COMMERCIAL ITEM DESCRIPTION

ROD ASSORTMENT, STIRRING, LABORATORY, Glass

The General Services Administration has authorized the use of this commercial item description in preference to Federal Specification NNN-R-560.

This commercial item description covers one type of glass stirring rod for laboratory use.

Salient characteristics:

Shall be an assortment of glass stirring rods suitable for laboratory use. Rods shall be fabricated of a suitable soda-lime or borosilicate glass material. Rod ends shall be fire polished for soda-lime glass or smoothly rounded for borosilicate glass. Each rod shall have a length of 300 millimeters (± 3 mm) and shall have nominal diameters of 4, 6, and 8 mm. The assortment of rods shall be packaged in a suitable container and each package shall weigh 1 pound nominal. Each package shall contain equal numbers of rods for each diameter. Lot quantities shall not be of mixed types of glass. Glass rods shall be in conformance with ASTM E-438, Laboratory Glass.

Rods shall have none of the following which will be considered defects.

Cracks or chips
Scratches
Dirt or grease
Cloudiness
Incorrect length or diameter
Unequal number of rods in each size

Workmanship. Individual glass rods shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing an assortment of rods (having an equal number of rods for each diameter) constitutes the unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2025-10-26 08:50Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspections. Inspections, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. Unless otherwise specified, sampling for examination shall be conducted at inspection level S-II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-II, AQL 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

The following National Stock Number is covered by this document:

6640-00-438-5950 ROD ASSORTMENT, STIRRING, LABORATORY, Glass

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

Project No. 6640-1353

Location S3-8-13

U.S. GOVERNMENT PRINTING OFFICE: 1984 - 603-034/45197

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.