A-A-51075B
26 February 1988
SUPERSEDING
A-A-51075A
15 March 1986

COMMERCIAL ITEM DESCRIPTION

COVER, X-RAY CASSETTE

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a sterile, disposable, large size, non-woven fabric, pillowcase-type cover for x-ray cassettes used during intra-operative radiographic procedures.

Salient characteristics:

The cover shall be sterilized, disposable, large size, pillowcase-type, of a non-woven fabric, for x-ray cassettes consisting of a removable inner sleeve and outer bag with cuff.

The cover shall have a removable blue inner sleeve composed of 1.5 mil thick polyethylene or polyester anti-static film, embossed. The sleeve shall protect the top outer portion and top inner portion of the cassette cover against possible contamination.

The outer bag shall consist of a white, 2.75 mil thick polyethylene anti-static spunbonded non-woven fabric, embossed.

The cover shall be fluid resistant to prevent contamination from blood, saline, or other fluids. The cover shall have a seamless bottom for strength.

The cover shall have a fold-over cuff, inner sleeve and outer bag, a minimum of 4 inches (10.16 cm) long, for aseptic handling and easy closure after cassette has been placed in the cover. The cover shall meet NFPA Standard No. 99 for anti-static materials.

AMSC N/A FSC 6525

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

The dimensions of the cover after the folding of the cuff shall be 24 inches (61 cm) ± 1/4 inch in height and 20 inches (51 cm) ± 3/4 inch in width. The cover shall easily accept a 5/8 by 14 by 17 inch x-ray cassette.

Each cover shall be folded to approximately 7 by 12 inches and individually wrapped in a peel-open package.

Workmanship. The covers shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing 20 individually packaged covers, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one (1) cover. Unless otherwise specified, sampling for examination shall be conducted at Inspection Level II, AQL 2.5, and sampling for tests shall be conducted at Inspection Level S-1, acceptance number is zero.

Sterility. Sterilization testing shall be in accordance with FDA's current Guideline on General Principles of Process Validation.

Drop test. Insert one hand under each white cuff and open bag mouth. Have assistant drop cassette into bag from height where bottom edge of the cassette is on the same plane as the edge of the blue cuff. Reject if heat-seals or white polyethylene non-woven material severs upon impact. Use 5/8" x 14" x 17" cassette of weight 4 to 5 pounds for testing.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6525-01-253-8295

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6525-1224

PGC 70033

Location: CID/COVER/H13-25

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.