A-A-51086
4 November 1985

COMMERCIAL ITEM DESCRIPTION

LIGHT, MICROSCOPE (Adjustable, 115 volt, 60 Hz, AC)

The General Services Agency has authorized the use of this commercial item description.

This commercial item description covers a microscope illuminator with a tungsten bulb, that can be adjusted to any angle above or below the stage of a microscope.

Salient characteristics:

Shall be an adjustable microscope illuminator for low power inspection or dissection of entomological specimens.

Shall be used with a variable voltage transformer that is operable from a 115 volt, 60 Hz, AC supply line.

Shall include the following:

1. A lamp housing with a replaceable, clear glass, 6.5 volt, 2.75 ampere tungsten bulb.
2. A variable voltage transformer in a base housing.
3. A three-link jack-knife upright support that attaches to the base housing with a thumbscrew.
4. A focusing mount with a three lens condensing system that attaches to the front of the lamp housing.
5. A filter holder that includes a replaceable ground glass and a replaceable blue filter, and attaches to the front of the focusing mount.
6. A 3-wire power cord (one ground) with connection to the variable voltage transformer and termination with a 3-prong plug for use with receptacles specified in American National Standards Institute (ANSI) C73.11.
7. An electrical cord interconnection between the variable voltage transformer and the lamp housing.

The three-link jack-knife shall be detachable from the base housing for direct attachment to a microscope having an appropriate threaded insert.

The focusing mount and filter holder shall be detachable from the lamp housing.

FSC 6650

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

The lamp housing shall include a socket which can be easily detached from the lamp housing for replacement of the lamp.

The lamp housing shall be well ventilated and shall be provided with cooling fins to dissipate the heat.

Workmanship. The microscope light shall be free from defects which detract from its appearance or impairs its serviceability.

Unit. Each (EA). One adjustable microscope light, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspections. Inspections, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Each microscope light shall be tested to demonstrate its capability to operate properly.

Sampling for examination. Sampling for examination shall be conducted in accordance with commercial practice.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

The following National Stock Number is covered by this document:

6650-00-428-7010

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL ENGINEERING COORDINATING ACTIVITY:

VA-OSS

Project No. 6650-0132

Location: S3-7-12

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.