COMMERCIAL ITEM DESCRIPTION

ENDOSCOPIC INSTRUMENT, FIBER OPTIC, FLEXIBLE

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers one size and type of single channel endoscopic instrument suitable for examining, photographing, and biopsing the colon from the sigmoid to the cecum.

Salient characteristics:

General. The fiberoptic endoscope shall have twin light illumination, an air-water-carbon dioxide channel, a suction-biopsy channel, and an adjustable ocular.

Dimensions. The fiberoptic endoscope shall have the dimensions shown in Table I.

TABLE I

Length (cm)	
deflection section	7
flexible shaft	160
entire instrument	187
Diameter (mm)	
distal tip	13.5
deflection section	12.5
flexible shaft	13.0

Deflection section. The deflection section shall be capable of 180° deflection in the right, left, up and down directions. The maximum deflection in any direction shall be 220°. The deflection controls shall be of the ratchet type, and shall lock the deflection section in the selected position. There shall be two separate controls for right/left and up/down deflection. The position of 0° deflection (neutral, deflection section straight) shall be clearly marked on each control.

Free-wheeling mode. The fiberoptic endoscopic instrument shall have a free-wheeling mode which disengages the ratchet mechanism and allows free movement of the deflection section.

Optics. The instrument shall provide a forward field of view of 75°, and shall have a focal depth of 5-100 mm. The instrument shall have an ocular adjustable to plus or minus 4 diopters.

Air-water-carbon dioxide channel. The air water - CO_2 channel shall have a nozzle on the distal end which directs the discharge upon the objective lens. The control for the channel shall provide for air (or CO_2) only, water only, or air-water mixture. The proximal head of the instrument shall have a water port capable of delivering

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.
Source: http://assist.daps.mil -- Downloaded: 2023-05-26 08:57Z
Check the source to verify that this is the current version before use.

water to the channel from a syringe. The instrument shall be capable of accepting air from an automatic air system, an air bag, or any alternate low pressure air source. The instrument shall be capable of accepting water from a water bottle for automatic water delivery to the channel.

Suction-biopsy channel. The suction channel shall be capable of passing 8 French biopsy forceps, cytology brushes, and other ancillary devices. The suction channel control shall be on the proximal head of the instrument.

Electrosurgical capability. The instrument shall have a terminal on the proximal end for connection of an electrical bypass cable. The cable, when connected to the terminal, shall bypass capacitatively coupled electrical energy from internal conductive components to the dispersive electrode or electrosurgical unit return terminal. The instrument shall meet the requirements of Underwriters Laboratories (UL) Standard 544.

Markings. The flexible shaft shall have markings in 10 cm increments beginning at the distal end.

Manufacturer's name or trademark. The manufacturer's name or trademark shall appear on the proximal end.

Auxillary optics. The fiberoptic endoscopic instrument shall be compatible for use with cinephotography, direct television hook-up and video tape equipment.

Accessories. The fiber optic endoscopic instrument shall be furnished with the following accessories:

Accessory	Quantity
Biopsy forceps, 7 French	1
Cleaning brush, manifold	1
Cleaning brush, forceps channel	1
Forceps valve seal	1
Syringe adapter	1
Protective plug	1
Valve assembly, biopsy, spare	1
Silicone fluid	1
Carrying case	1
Operating and instruction manual	1

Material. The fiber optic endoscopic instrument shall be manufactured from material suitable in all aspects for the purpose intended. The distal tip, flexible shaft, suction channel, and air-water-CO_2 channel shall be manufactured from a material capable of repeated chemical sterilization. Plastic and neoprene components making contact with the patient shall meet the requirements of the United States Pharmacopeia for Class II plastics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

Workmanship. The fiber optic endoscopic instruments shall be free from defects which detract from their appearance or impair their serviceability.

PRODUCT QUALITY LEVEL

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with the requirements specified.

Unit. Each (EA). One fiber optic endoscope, with accessories and case, as specified, constitutes one unit of issue.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one fiber optic endoscope.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, ?rahndling, etc., involving shipment from the supply source to the receiving ?ractivity, plus reshipment from receiving activity, and shall conform to applicable c?rarrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	Item Identification
6515-01-204-2662	Endoscopic Instrument, Fiber Optic, Flexible

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA - DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

Project No. 6515-3436

Location: S1-10-2

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.