A-A-51092A
15 June 1990
SUPERSEDING
A-A-51092
30 October 1985

COMMERCIAL ITEM DESCRIPTION

ILLUMINATOR, X-RAY FILM

(Four Section, Mobile)

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Type V of Federal Specification GG-I-446E which is cancelled.

This Commercial Item Description covers a mobile x-ray film illuminator with four 14 by 17 inch sections, side by side, suitable for viewing medical x-ray films of all sizes.

Salient characteristics:

The x-ray film illuminator shall be a mobile x-ray film illuminator with four 14 inch by 17 inch sections, side by side, suitable for viewing medical x-ray films of all sizes up to 14 inches by 17 inches simultaneously. The illuminator shall operate satisfactorily on 110-125 volts, 60 Hz, AC.

The illuminator shall consist of a metal frame, suitably finished, supporting an x-ray film illuminator having a white, translucent, plastic viewing panel approximately 16-7/8 inches high and 56 inches long, in a suitable housing with four switches controlling the illumination of the panel (4 areas), and four film holding devices of the continuous gravity or strip spring design. The frame shall be equipped with a minimum of four swivel casters, two of which shall have foot-operated brakes. The illuminator shall be rigidly constructed, stable, and free of excessive vibration during normal use.

The illuminator housing shall be constructed of carbon steel sheet at least 0.032 inch thick (No. 21 Standard Gage), molded fibrous glass-reinforced plastic, or aluminum. The carbon steel and aluminum housing shall be finished with lacquer, baked-on enamel, or an epoxy powder coating. The

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

aluminum housing may be clear anodized in lieu of a painted finish. The housing shall be the manufacturer's standard color.

The illuminator housing shall be mounted on a stable base equipped with four ball-bearing, rubber-tired, swivel casters, at least 3 inches in diameter. The casters shall roll smoothly and quietly. The base shall have two shelves spaced for convenient film storage. The shelves shall be sheet carbon steel, a minimum of 0.047 inch thick (No. 18 Standard Gage), or an equivalent strength aluminum. The mobile base may be supplied with a drawer located on the right hand underside of the upper shelf. The drawer shall be at least 20 inches wide, 15 inches deep, 4 inches high, and shall have an easily releasable catch.

The lower edge of the illuminator housing shall be at least 48 inches from the floor, and the top shall be not more than 76 inches from the floor. The illuminator housing shall be pivot-mounted to allow positioning from the vertical through the horizontal (face-up) and shall lock in any position without the use of tools. The housing, combined with the viewing panel, shall provide a rigid enclosure for the fluorescent lamps and the electrical accessories.

The illuminator panel shall be large enough to accommodate four 14 inch by 17 inch radiographic films, side by side. Light baffles, located behind the panel, shall divide the panel into four sections, and allow any combination of films to be viewed simultaneously. The baffles shall block light from adjoining sections, but they shall not be visible from the front of the viewing panel when they are between the illuminated sections. Each section shall have a brightness of not less than 400 foot-lamberts over a square, 6 inches by 6 inches, at the center of the viewing area with overall distribution of even brightness to provide satisfactory film viewing. Switches, conveniently located, shall allow illuminating the film sections either individually or simultaneously.

The panel shall be distortion-free and be 0.125 inch to 0.040 inch thick, with a light transmission factor between 50 and 60 percent. The panel shall have a smooth glossy surface on at least one side and shall be free from bubbles, variations in thickness, and flaws in the finish. When the viewing panel is on place on the housing, the smooth, glossy side shall be exposed.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

The roller-gravity or strip-spring type film holder shall be continuous along the top of each section of the illuminator. The film holding devices shall be an assembly permitting easy removal of the viewing panel or the film holding devices. The device shall permit easy insertion and removal of any medical x-ray film at any point across the full width of each panel, without scratching the film.

The illuminator shall be equipped with a three-conductor line cord assembled with a hospital grade, grounded, three-prong plug which conforms to ANSI C73.11. The exposed portion of the cord shall be a minimum of 10 feet long. The illuminator shall show no signs of dielectric breakdown when tested for dielectric withstand as specified in UL 153. The illuminator shall meet the requirements of UL 544, Medical and Dental Equipment, as applicable.

Surfaces to be welded shall be free from grease, oil, oxides, or other foreign matter. All weld joints shall blend into the adjacent metal in gradual, smooth curves. Beads shall be smooth and free of slag, excessive undercut, or excessive splatter. Sufficient metal shall be added to provide a suitable fillet of reinforcement. Welds shall not be oxidized. All welds shall be continuous with no visible voids, pin holes, or cracks.

All exposed edges of the illuminator shall be free of burrs or sharpness. The illuminator finish shall be free of chipped, cracked, scratched, or rough areas.

Workmanship. The x-ray film illuminator shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One illuminator, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 4.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one x-ray film illuminator as specified in this description.

Sampling for tests. Sampling for tests, including dimensional tests, shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of S-2. The unit of product for test purposes shall be one x-ray film illuminator.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6525-00-604-0125

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

A-A-51092A

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6525-1250

Location:
CID/ILLUMINATOR/FOUR/H-13-16

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.