A-A-51093A
23 November 1990
SUPERSEDING
A-A-51093
30 October 1985

COMMERCIAL ITEM DESCRIPTION

SAFELIGHT, DARKROOM, RADIOGRAPHIC

The General Services Administration has authorized the use of this commercial item description as a replacement for Federal Specification G-G-S-86C, which is cancelled

This Commercial Item Description covers a radiographic darkroom safelight

Salient Characteristics

The radiographic darkroom safelight shall be suitable for mounting in an x-ray darkroom. The safelight shall be one of the following styles:

Style 1 - Wall type

Style 2 - Wall or table type

Style 3 - Ceiling type

Style 1

The wall type safelight shall be housed in a metal box having the following overall dimensions

Width	7-1/4 inches plus or minus 1/2 inch
Height	10-1/2 inches plus or minus 1/2 inch
Depth	5-1/4 inches plus or minus 1/2 inch

The safelight shall consist of an angulated safelight and a film viewer, each of which is 5 inches plus or minus 1/2

AMSC N/A FSC 6525

DISTRIBUTION STATEMENT A Approved for public release, distribution is unlimited

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

inch in width by 7 inches plus or minus 1/2 inch in length. The safelight shall be furnished with two 10 watt incandescent lamps.

The style 1 safelight shall be provided with eyelets or keyhole slots, or other suitable device, to allow the safelight to be mounted on a wall

Style 2

The wall or table type safelight shall be housed in a metal box having the following overall dimensions:

Width: 11-1/4 inches plus or minus 1/2 inch

Height: 16-1/2 inches plus or minus 1/2 inch

Depth: 6-3/4 inches plus or minus 1/2 inch

The safelight shall have a film viewer 9 inches plus or minus 1 inch wide by 5-1/4 inches plus or minus 1/2 inch high located in the upper portion of the front of the metal box, and a safelight lens 10 inches plus or minus 1 inch wide by 8 inches plus or minus 1 inch high located below the film viewer in the front of the metal box The safelight shall be furnished with two 20 watt incandescent lamps

The style 2 safelight shall be provided with eyelets or keyhole slots, or other suitable device, to allow the safelight to be mounted on a wall As an alternative, the style 2 safelight shall be designed to stand firmly on a shelf, bench, tabletop, or other flat, horizontal surface without the need for special mounting

Style 3

The ceiling type safelight shall have a medium-screw base rigidly attached to the lamp housing to allow the safelight to be screwed into a standard medium-screw socket

The face of the lamp housing shall be 5-1/2 inches plus or minus 1/2 inch in diameter and shall be covered with an x-ray filter held in place by a seal

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

General.

All metal housings shall be made of carbon steel, not less than 0 025 inch thick, or of other material and thickness adequate for providing an equivalent strength and rigidity All seams shall be joined to provide rigidity and to prevent emission of white light when filtered light only is required. Electrical outlets, cords, switches, mounting devices, and other parts attached to the housing shall also be installed so as to prevent emission of white light when filtered light only is required Filter holders and retaining devices shall be so constructed that there is no leakage of "white" light when the filter holder assembly is in use. It shall be possible to remove and replace filters without the use of any tools other than a common flat-blade screwdriver.

Both the safelight and the illuminator on styles 1 and 2 shall have their own on-off switch, with the switch for the safelight being identifiable by touch as to prevent inadvertently turning on the white light when a safe light is desired. As an alternate, a three-function rocker switch can be used (i e top position is the white light, bottom position as the safe light, and middle positon is off) In no case shall so-called "sequential" switches be used

The safelight and safe-light illuminators shall be provided with medium screwbase sockets to accommodate the required lamps The sockets shall be rigidly, but removably, attached to the lamp house Each safelight shall be furnished with an appropriate means of notification of the maximum wattage of the lamps The recommended lamps shall be rated for 115 volt, 60 Hz AC power operation

All safelight electrical wiring shall conform to National Fire Protection Association (NFPA) Publication No 70

Styles 1 and 2 safelights shall be furnished with a flexible, 3-wire power cord with a hospital grade plug, the cord being at least six feet in length All power cords and plugs shall be Underwriter Laboratories (UL) approved

All welding shall be durable, sound in every detail, and in accordance with the latest practices recommended by the American Welding Society Safelight metal finishes shall be

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

hard, smooth, and highly resistant to abrasion, chipping, and cracking Manufacturer's standard color shall be furnished

Workmanship The radiographic dark room safelight shall be free from defects which detract from its appearance or impair its serviceability.

Unit Each (EA) One safelight of the style and size, specified, constitutes one unit

Quality Assurance Provisions

Responsibility for inspection Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item

Sampling for examination Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 4.0 (percent defective) and an inspection level of I The unit of product for examination purposes shall be one radiographic darkroom safelight

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests, shall be conducted in accordance with MIL-STD-105, with an AQL of 0 (percent defective) and an inspection level of S-3. The unit of product for test purposes shall be one radiographic darkroom safelight.

Sampling for dimensional tests, shall be conducted in accordance with MIL-STD-105, with an AQL of 6 5 (percent defective) and an inspection level of S-4. The unit of product for test purposes shall be one radiographic darkroom safelight.

Tests Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics Tests shall include, but shall not be limited to, the following·

Light tightness test A lamp of the required wattage shall be installed in the safelight An opaque blank, dimensioned to the filter size, shall be installed in the filter holder of the safelight The safelight shall be mounted in its normal operating position in a totally darkened room. Sheets of unexposed Eastman Kodak Company Ortho G film or equal, not less than 8 inches by 10 inches, shall be placed in positions where film would normally be handled. The safelight shall be turned on At the end of 30 minutes, the film shall be processed and examined for evidence of light fog. Any light fogging shall constitute a failure

Temperature rise test. The safelight shall be operated under conditions of normal use in an ambient temperature of 90°F, with a lamp of required wattage and at rated voltage The safelight shall have a green filter of suitable size installed in the filter holder The safelight shall be operated until thermal equilibrium is established but not less than 1 hour The temperature of the hottest part of the filter shall be determined by means of a thermocouple attached to the filter by tape or other suitable material The ambient temperature during the test and the thermocouple temperature reading shall be recorded A difference of more than 40°F between the two temperatures shall constitute a failure

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract

Regulatory requirements

Federal Food, Drug and Cosmetic Act If the product covered by this document has been determined by the U S Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances and regulations

Recovered materials The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23 4 to the maximum extent practical

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document.

6525-00-299-8495 (Style 1)
6525-00-612-7840 (Style 2)
6525-00-612-7825 (Style 3)

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number 6525-1264

Location SLITECID 13A

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4956 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-51093A	SAFELIGHT, DARKROOM, RADIOGRAPHIC

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 1426 82 MAR

PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.