A-A-51101B
13 May 1991
SUPERSEDES
A-A-51101A
7 August 1989

COMMERCIAL ITEM DESCRIPTION

BEAKER, LABORATORY (Borosilicate Glass)

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers Griffin low form borosilicate glass laboratory beakers.

Salient Characteristics:

Shall be graduated borosilicate glass beakers, with single or double pour out lips, for general laboratory use.

Shall be in accordance with all the requirements of ASTM Designation E960 for type I Griffin beakers and applicable sizes (see "Ordering Data") for the 10 to 2000 milliliter (mL) rated capacity range.

Shall have graduation markings that comply with the requirements in ASTM Designation E960.

Workmanship. The beakers shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing 12 beakers for capacity sizes 10 to 400 ml, as specified, constitutes one unit. One package containing 6 beakers for capacity sizes 600 to 1000 ml, as specified, constitutes one unit. One package containing 4 beakers for capacity sizes 1500 to 2000 ml, as specified, consitutes one unit.

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspections and tests. Sampling for inspections and tests shall be conducted in accordance with commercial practice.

Inspections and tests. Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Other publications. The following documents form a part of this specification to the extent that is specified herein. Unless otherwise indicated, the effect on date of invitation for bids or request for proposals shall apply.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

ASTM E960 Standard Specification for Laboratory Glass Beakers

(Application for copies should be addressed to American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Numbers are covered by this document:

NSN	CAPACITY	ASTM DESIGNATED	ITEM DESCRIPTION
6640-01-330-9374	10 ML	1	BEAKER, LABORATORY, Borosilicate Glass, 10 ML, 12's
6640-01-330-3800	30 ML	3	BEAKER, LABORATORY, Borosilicate Glass, 30 ML, 12's
6640-01-328-9684	50 ML	4	BEAKER, LABORATORY, Borosilicate Glass, 50 ML, 12's
6640-01-141-1222	100 ML	5	BEAKER, LABORATORY, Borosilicate Glass, 100 ML, 12's
6640-01-329-6487	250 ML	7	BEAKER, LABORATORY, Borosilicate Glass, 250 ML, 12's
6640-01-329-7539	400 ML	8	BEAKER, LABORATORY, Borosilicate Glass, 400 ML, 12's
6640-01-328-9683	600 ML	9	BEAKER, LABORATORY, Borosilicate Glass, 600 ML, 6's
6640-01-328-9682	1000 ML	11	BEAKER, LABORATORY, Borosilicate Glass, 1000 ML, 6's
6640-01-328-9681	2000 ML	13	BEAKER, LABORATORY, Borosilicate Glass, 2000 ML, 4's

Ordering data. (Capacity size(s), intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.)

MILITARY INTERESTS: PREPARING ACTIVITY:

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

Custodians:

Army - MD
Navy - MS
Air Force - 03

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project Number: 6640-1710

Location: ENABLE/BEAKERCD/S31

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I. RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-51101B	13 May 1991

3. DOCUMENT TITLE
BEAKER, LABORATORY (Borosilicate Glass)

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION

c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center	(215)737-2870	444-2870

c. ADDRESS (Include Zip Code)	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT:
ATTN: DPSC-MSTE 2800 S. 20th Street Philadelphia, PA 19101	Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete — 96/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.