A-A-51103
20 November 1985

COMMERCIAL ITEM DESCRIPTION

FLASK, DISTILLING, 125 ml, Borosilicate Glass

The General Services Administration has authorized the use of this commercial item description in preference to Type III, Class 1 of Federal Specification NNN-F-240.

This specification covers one type of distilling flask for laboratory use.

Salient characteristics:

The flask shall be fabricated from borosilicate glass conforming to Type I, Class A, of ASTM E-438. Shall have a rounded bottom and a long neck with a reinforced mouth to accommodate a proportionally sized rubber stopper. The neck shall have a side arm at a 75° lower angle to be 100 mm in length and have a nominal diameter of 7 mm. The side arm shall be attached 137 mm. (± 3 mm.) from the bottom of the flask. Shall be uniform in construction and suitable for laboratory use.

One side of the flask shall provide a spot for marking. Each flask shall have the manufacturer's name or registered trademark. The capacity of the flask shall be marked on the flask. All markings on the flask shall be permanent and legible. Flasks shall not have discolorations, cracks or deficiencies in construction. Shall be free of internal foreign matter or grease and dirt, inside and out which cannot be easily removed with normal washing.

Workmanship. The flask shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One flask, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:50Z
Check the source to verify that this is the currentversion before use.

The following National Stock Number is covered by this document:

6640-00-423-8500 FLASK, DISTILLING, 125 ml.

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Project No. 6640-1342

Location S3-8-8

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:50Z
Check the source to verify that this is the current version before use.