COMMERCIAL ITEM DESCRIPTION

FLASK, ERLENMEYER (Graduated, Plastic)

The General Services Administration has authorized the use of this commercial item description in preference to Military Specification MIL-F-36747.

This commercial item description covers one type of plastic Erlenmeyer flask.

Salient characteristics:

The flasks shall be the Erlenmeyer type, fabricated from clear, shatterproof, chemically resistant plastic capable of withstanding autoclaving. Shall be of the following sizes and graduations as listed in Table I.

TABLE I

Size	Capacity (ml)	Graduations (ml)	Range (ml)
1	50	10	20- 50
2	250	25	50- 250
3	1000	50	200-1000

The flasks shall have a narrow mouth with a reinforced lip and shall be shatterproof. Shall be suitable for laboratory procedures and shall be permanently and legibly marked. Each unit shall have the manufacturer's name or registered trademark.

Flasks shall have an accuracy to ±5% of any specific graduation marking.

Workmanship. The flasks shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). Each flask, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:50Z
Check the source to verify that this is the current version before use.

uperformance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspections. Inspections, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of pnysical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. Unless otherwise specified, sampling for examination shall be conducted at inspection level S-II, AQL 1.0, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:50Z
Check the source to verify that this is the currenversion before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

The following National Stock Numbers are covered by this document:

6640-00-010-8360	FLASK, ERLENMEYER, Graduated, 50 ml Size 1
6640-00-010-8361	FLASK, ERLENMEYER, Graduated, 250 ml Size 2
6640-00-010-8362	FLASK, ERLENMEYER, Graduated, 1000 ml Size 3

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

Project No. 6640-1337
Loc. CID, S3 7-17

Location S3-7-17

U.S. GOVERNMENT PRINTING OFFICE: 1984 - 605-036/45196

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:50Z
Check the source to verify that this is the current version before use.