A-A-51105
20 November 1985

COMMERCIAL ITEM DESCRIPTION

CLAMP, LABORATORY, SUPPORT ROD

The General Services Administration has authorized the use of this commercial item description in preference to Federal Specification NNN-C-440.

This commercial item description covers one type of clamp rod support for laboratory use.

Salient characteristics:

Shall be a swivel clamp suitable for securing laboratory apparatus to support stands.

Shall consist of two separate V-grooved sections which are freely movable through an angle of 360° and in parallel planes capable of a fixed position by means of a set screw. Shall have a nominal overall length of 6 inches. Shall be fabricated from an aluminum or zinc base alloy. The V-grooved areas shall be fitted with thumbscrews for attachment to support rods or anchoring clamp shafts. Thumbscrews shall be fabricated from a corrosion-resistant metal. Shall be capable of direct attachment to rods up to 1/2 inch in diameter without passing over the end of the rods and attachment to rods up to 11/16 inch (10 mm.) in diameter by passing over the end of the rod.

Clamps shall have none of the following which will be considered defects.

Wobble while attached
Rust or corrosion
Casting or machine flaws
Not perpendicular to rod in which it is attached
Sharp or rough edges

Workmanship. The clamp shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One clamp, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspections. Inspections, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one swivel clamp. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil - Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

COMMERCIAL ITEM DESCRIPTION A-A-51105

The following National Stock Number is covered by this document:

6640-00-428-2460 CLAMP, LABORATORY, SUPPORT ROD

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

Project No. 6640-1347
Loc. CID,3 7-29

U.S. GOVERNMENT PRINTING OFFICE: 1986 — 605-036/44560

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.