A-A-0051491 (DM)
10 July 1986

INTERIM COMMERCIAL ITEM DESCRIPTION

FORCEPS, BRONCHOSCOPIC (Jackson)

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA, 19101, based upon currently available technical information.

This interim commercial item description covers Jackson design bronchoscopic forceps for use in removing foreign objects from the human bronchial tract and esophagus.

Salient characteristics.

Shall be a bronchoscopic forceps, side curved, regular, with box lock type handle, large serrated jaws, Jackson design, having a working length of approximately 50 cm. Shall be suitable for use, through a tube 5 mm in diameter or larger, in removing foreign objects from the human bronchial tract and esophagus.

The forceps shall consist of a cannula containing the forceps jaws stylet mechanism with the tube and stylet attached to a scissors type actuation handle for use in positioning the forceps and in opening and closing the jaws.

Material. Instrument components shall be fabricated from the following materials:

a. Handle and jaws shall be martensitic corrosion-resisting steel.

b. Stylet and handle set screws shall be austenitic corrosion-resisting steel with not more than 1% manganese.

c. Cannula shall be nickel copper alloy.

d. Linkage parts (cannula to handle) and stylet nut shall be nickle silver or austenitic corrosion-resisting steel with not more than 2% manganese.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.

Austenitic and martensitic corrosion-resisting steels shall conform to Class 3 and Class 4 respectively, stainless steel specified in ASTM F899. The copper nickel and nickel silver alloys shall conform to Class 8 composition, specified in Federal Specification GG-I-526.

Tests. The forceps shall show no evidence of corrosion when subjected to the corrosion testing specified in Federal Specification GG-I-526.

Workmanship. The bronchoscopic forceps shall be free from defects which detract from its appearance or impair its serviceability. Shall be free from corrosion, pits, deformation, burs, sharp edges and foreign matter.

Unit. Each (EA). One forceps, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Testing shall be at Inspection Level of S-2 with dimensional testing at a 1.5% AQL and hardness and corrosion testing at a 2.5% AQL. The unit of product for test purposes shall be one bronchoscopic forceps.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-00-331-9100	FORCEPS, BRONCHOSCOPIC, Regular, Jackson, 50 cm

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:56Z
Check the source to verify that this is the current version before use.

A-A-0051491 (DM)

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS

PGC 41621

Project No. 6515-3761

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:56Z
Check the source to verify that this is the current version before use.