A-A-0051492 (DM)
17 June 1986

INTERIM COMMERCIAL ITEM DESCRIPTION

INCISION-DRAINAGE SET, STERILE, DISPOSABLE

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA, 19101, based upon currently available technical information.

This interim commercial item description covers a incision and drainage set suitable for minor surgical procedures.

Salient characteristics.

General. The incision and drainage set shall be sterile, disposable and suitable for use in minor surgical incision and drainage procedures.

The set shall consist of the following items as a minimum:

One plastic bag with twist tie sufficient in capacity when used in conjunction with the complete incision and drainage set.

One fenestrated drape and one fluid resistant towel both of a nonwoven fabric, and of a size suitable for standard incision and drainage procedures.

One alcohol prep swab 4 inches by 3 inches (10 cm x 8 cm) with dimensions within 1/2 inch.

One Povidone Iodine prep swab 4 inches by 3 inches (10 cm x 8 cm)(with dimensions within 1/2 inch.

Four cotton-filled sponges with minimum dimensions of 3 by 3 inches 8 cm x 8 cm).

One metal scalpel with an attached number 11 blade. The blade shall be smooth, and void of any burrs, chips, pits, nicks, ragged edges, corrosion or foreign matter.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.

One tissue forceps 5 inches (13 cm) with dimensions within 1/2 inch and one straight Kelly hemostat 5-1/2 inches (14 cm) with dimensions within 3/4 inch. These instruments shall be fabricated of nickel-plated steel and shall be without rough edges, burs and corrosion. The instruments shall comply with their respective performance tests.

Forceps performance test. The instrument shall have no looseness or free play of either shank which would result in any misalignment of the tips in closing. Upon closing, tips which do not meet shall be rejected. Upon completion of test, the instrument shall undergo repetition a second time with the same criteria.

Hemostat performance test. In addition to compliance with the forceps test, the hemostat pin or rivet shall be permanently secured, insuring no free play of the shanks. Upon closing the hemostat, tips that do not successively tighten from the extreme distal tip shall be rejected. In the fully closed position, the hemostat shall have no visible gap on the mating surface.

Sterility and pryogenicity. Components shall be sterile and free from pyrogens. Kits shall have a minimum sterility assurance level of 0.0001 and shall be in accordance with the USP.

Workmanship. The incision and drainage set shall be free from defects which detract from its appearance or impair its serviceability.

Expiration dating. Expiration dating period of the set shall not be less than 24 months. Not less then 5/6 of the expiration dating period shall remain at the time of delivery to the Government.

Unit. Package (PG). One package containing twenty sets, each set packaged sterile in a sealed peel-open tray with items positioned within the tray in sequence of use, as specified constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:56Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances of procedures) and testing (determination by technical and chemical properties) of the item.

Examination. Kits shall be examined to determine compliance with specification requirements. Sampling shall be in accordance with MIL-STD-105, sampling level II, AQL 2.5.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one kit. Sampling shall be at an inspection level of S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-153-4910	INCISION-DRAINAGE SET, DISPOSABLE, STERILE, 20s

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS

PGC 59780

Project No. 6515-3763

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:56Z
Check the source to verify that this is the currentversion before use.