A-A-51496A
5 October 1987
SUPERSEDING
A-A-0051496(DM)
28 October 1986

COMMERCIAL ITEM DESCRIPTION

RECOVERY PACK, BLOOD CELL COLLECTION

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a sterile, pyrogen free, quadruple red blood cell recovery pack for use with a Blood Cell Washer and a Wash Bowl and Waste Bag Set.

Salient characteristics.

General. Shall be a sterile, pyrogen free, quadruple red blood cell recovery pack for use with a Blood Cell Washer and a Wash Bowl and Waste Bag Set.

Material. The recovery pack shall consist of four PVC plastic 600 milliliter bags, or as an alternate, two 600 milliliter and two 350-400 milliliter PVC plastic bags, connected by various length of PVC tubing, a standard plastic spike and polypropylene control clamps.

Characteristics. The recovery pack system shall consist of the spike connected by six inches of tubing to a plastic "Y" fitting connecting to two 34-inch lengths of tubing, each one attached to a 600 ml bag (primary).

Each primary bag shall be connected to a blood bag (transfer) by eight inches of tubing. Each leg, except the spike to "Y" fitting, shall contain a control clamp. The entire system shall be closed and tamperproof. All tubing shall have an outer diameter of .158 $\pm$ .003 inches with a wall thickness of .020 $\pm$ .001 inches.

The two primary bags shall be used to store two previously frozen and washed units of red blood cells suspended in approximately 250 ml of sodium chloride-glucose solution. The two transfer bags shall be used to receive 250 ml or more of the sodium chloride-glucose solution to be discarded prior to transfusion.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:56Z
Check the source to verify that this is the current version before use.

Bags shall be shaped to permit centrifugation in a standard 1000 ml trunnion cup. The two primary bags shall each be provided with two blood outlet ports, of a standard length and diameter, containing an easily puncturable diaphragm that forms a leakproof, air-proof, seal that shall require no more than normal manual force to remove. As an alternate, transfer bags may also possess outlet ports.

All ports shall be fully compatible with a Blood Administration Set. The primary bags shall have a horizontal slot to suspend the bag for administration.

Each primary bag shall be supplied with a permanently applied "commonality label" describing product to be "Red Blood Cells, Deglycerolized" product number 06400. This label shall be printed according to the latest edition of FDA's "Guidelines for Uniform Labeling" to include manufacturer and lot number.

Additions to this label by the processing facility shall consist of, but not limited to, name of processing activity, unit number and unit blood type.

The tubing connecting the "Y" junction and each primary bag shall bear a different set of at least ten indelibly imprinted numbers along the tubing for the purpose of dividing the donor tubing into segments for pilot samples if required. The center distance between numbered segments shall be between 2-1/2 and 3-1/2 inches.

The spike shall be suitably protected with a plastic cover to maintain sterility, and form a leak-proof, air-tight seal with a Wash Bowl and Waste Bag Set.

If a roller shut-off clamp is provided, the security of the roller in the clamp body shall be tested as follows:

The roller clamp shall be removed from the set together with a piece of tubing approximately one inch longer than the clamp body. The tubing shall be present in the roller clamp in its normal position with a 1/2 inch length extending from each end of the roller clamp. The clamp, see Figure 1, shall be held vertically, with the large end down, over a supporting gap so that the roller, if forced out of the body, is free to fall. A polished steel rod is applied to the top of the roller, so that the roller is able to rotate in the course of the test and the vertical downward force required to separate the roller from the clamp body is measured.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.

As an alternate, any position of the clamp is acceptable as long as the clamp body is not distorted and the load measurement is accurate. Roller shall sustain a load of 3.65 kg without separation from the clamp body.

Pyrogenicity and toxicity. The fluid path shall be pyrogen-free and non-toxic meeting the requirements of the USP for Class III plastics.

Sterility. The fluid path shall meet the sterility requirements when tested in accordance to the USP.

Workmanship. The recovery pack shall be free from defects which detract from its appearance or impair its serviceability. The recovery pack shall be free from cuts, tears, flash or weak spots.

Unit. Package (PG). One package containing ten individually packaged recovery packs, as specified, constitutes one unit. The individual package shall be a sealed container capable of maintaining sterility of contents unless package is damaged or opened.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:56Z
Check the source to verify that this is the currentversion before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. At inspection Level II, the AQL shall be 1.5%. The unit of product for test purposes shall be one recovery pack.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.

NOTE 1: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-235-6088	RECOVERY PACK, Blood Cell Collection, 10s

NOTE 2: Blood Cell Washer indicted herein is covered by NSN 6640-01-235-6131. Wash Bowl and Waste Bag Set indictated herein is covered by NSN 6640-01-234-9582. The 1000 ml standard tunnion cup indicated herein is part of centrifuge NSN 6640-01-139-7457. (Equal to Dupont Model RC3B centrifuge) Blood Administration Set indicated herein conforms to NSN 6515-01-128-1407.

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PBS
FDA-MPQAS

PGC S0074

Project No. 6515-4065

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:56Z
Check the source to verify that this is the currentversion before use.




FIGURE 1
METHOD OF APPLYING LOAD TO ROLLER

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-51496A	RECOVERY PACK, BLOOD CELL COLLECTION

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)*
b. ADDRESS *(Street, City, State, ZIP Code)*	

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

(FOLD ON THIS LINE)

(FOLD ON THIS LINE, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.