A-A-51507
23 June 1986

COMMERCIAL ITEM DESCRIPTION

TUBES, TRACHEAL, MAGILL

The General Services Administration has authorized the use of this commercial item description in preference to Military Specification MIL-T-36145 (Bu Med), 16 October 1963.

This commercial item description covers sterile, plastic Magill type tracheal tubes.

Salient characteristics:

General. Tracheal tubes shall meet all requirements of American National Standards Institute (ANSI) Standard Z79.14 for uncuffed Magill-type tracheal tubes with the exceptions specified herein. Tracheal tubes shall be capable of chemical disinfectant or steam sterilization, or sterilization with boiling water. The recommended method of sterilization shall be marked on the unit package in accordance with ANSI Standard Z29.14.

Sterility. Tracheal tubes shall be sterile. Sterility shall be tested in accordance with the current edition of the United States Pharmacopiea (USP).

Biocompatibility (inertness). Plastic used in the manufacture of the tubing shall be tested for biocompatibility in accordance with the requirements for Class IV plastics of the current USP.

Workmanship. Tracheal tubes shall be uniform in quality and free from any defects which affect the appearance or utility of the tube. Tracheal tubes shall be free from foreign matter, embedded particles or air, burrs, flashing, rough or sharp areas, cracks, and other defects.

PRODUCT QUALITY LEVEL

Visual and dimensional examination. Tracheal tubes shall be examined to determine compliance with specification requirements.

Unit. Package (PG). One package containing 10 tracheal tubes constitutes one unit. The tubes shall be individually packaged sterile in a peel-open type container capable of maintaining the sterility of the contents unless the package is damaged or opened.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by of for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Certificates of biocompatibility. If the contractor has performed biocompatibility testing, in accordance with the requirements of this specification, within two years of the date of solicitation, and if the contractor certifies that there have been no changes in material formulation or processing since actual testing then a certificate of biocompatibility may be submitted in lieu of said testing by the contractor. The certificate shall include the date of actual testing, the test method and results, the name and title of the person evaluating results, the current contract number and NSN, a statement to the effect that material processing and formulation have not changed since actual testing was performed, and the signature and title of the person making the certification.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packing, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers (NSNs) are covered by this document:

NSN	Item Identification
6515-01-127-2197	Tube, Tracheal, Magill, 3.5 mm ID, 10s
6515-01-127-2198	Tube, Tracheal, Magill, 4.0 mm ID, 10s
6515-01-127-2199	Tube, Tracheal, Magill, 4.5 mm ID, 10s
6515-01-127-2200	Tube, Tracheal, Magill, 3.0 mm ID, 10s
6515-01-127-2201	Tube, Tracheal, Magill, 2.5 mm ID, 10s
6515-01-128-4838	Tube, Tracheal, Magill, 5.0 mm ID, 10s

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Ordering date (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS

PGC 06009

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Project No. 6515-3434

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. *All users of military standardization documents are invited to provide suggestions.* This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 WASHINGTON D. C.
POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-ATTS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A A-51507	Tubes Tracheal, Magill

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*: _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 82 MAR 1426

PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.