A-A-0051508 (DM)
18 June 1986

INTERIM COMMERCIAL ITEM DESCRIPTION

SYRINGE, ASPIRATING, EAR AND ULCER

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA, 19101, based upon currently available technical information.

This interim commercial item description covers a sterile, disposable, plastic syringe for use in the aspiration of mucous from the ear, nose and mouth of premature or newborn infants.

Salient characteristics.

General. Shall be a sterile, disposable syringe for the aspirating of mucous from the ear, nose and mouth of premature and newborn infants.

Material. The syringe shall be manufactured entirely from flexible polyvinyl chloride. The syringe shall meet the requirements for Class II plastics as stated in the United States Pharmacopeial (USP).

Characteristics. The size 1 syringe shall have a nominal capacity of 2 ounces. The size 2 syringe shall have a nominal capacity of 3 ounces. The syringe shall have a tapered tip. The syringe shall have molded rings designed to provide a secure grip. The syringe tip shall have a nominal length of 2-1/2 inches. The tapered tip shall be free from any restrictions or occlusions. The syringe shall create a vacuum of 110 mm Hg (minimum).

Physical properties. The polyvinyl chloride used in the m manufacture of the syringe shall have the properties shown in Table I. A sufficient amount of the raw material, which is representative of the lot or batch, shall be used as a sample for testing.

AMSC N/A FSC 6515
DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

TABLE I

Property	Value	Test Method
Tensile strength, break	1100-1400 psi	ASTM D638
Elongation %, break	450-500	ASTM D638
Tensile modulus 100%	350-400 psi	ASTM D638
Hardness (shore durometer)	55-63A	ASTM D2240

Hardness. Using ASTM D2240, the gage reading shall be made 10 seconds after application of the instrument.

Vacuum test. Attach a manometer or pressure gage to the tip of the syringe. Manually evacuate the air from the syringe. Release the syringe. The syringe shall create a vacuum of 110 mm Hg (minimum) and hold the vacuum for a minimum of one minute.

Accelerated aging test. Preheat an air oven to 140°F + 4°F prior to placing the sample in the oven. The sample shall be placed in the oven for seven days. The appearance of the sample before and after the aging test shall be noted. The physical properties listed in Table I shall not decrease by more than 25 percent due to the aging test. The vacuum suction pressure shall not be less than 110 mm Hg as a result of the aging test. The physical properties of the representative polyvinyl chloride sample shall be established prior to performing the accelerated aging test.

Sterilization, safety and freedom from pyrogens. The syringe shall be sterile, free from pyrogens and safe. The syringe shall be tested for sterility, freedom from pyrogens and safety in accordance with the USP.

Certificates of biocompatibility. A certificate of compliance, stipulating that offered items meet the biocompatibility test, shall be acceptable, provided that the contractor has supplied acceptable items to the Government within the past three years and that there have been no changes to material formulation and method of processing since actual testing was performed. The certificate shall include the date the testing was performed, the test method and results, the name and title of the person evaluating the results, the current contract number and NSN, a statement certifying that material processing and formulation have not changed since the time that biocompatibility was performed and the signature and title of the person making the certification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Workmanship. The aspirating syringe set shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing fifty syringes of one size only constitutes one unit. Each syringe shall be packaged sterile in a sealed peel-open type container capable of maintaining sterility of contents.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one aspirating syringe. Sampling shall be S-2, AQL 2.5.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:57Z
Check the source to verify that this is the currentversion before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-045-0029	SYRINGE, ASPIRATING, Ear and Ulcer, Size 2
6515-01-198-0181	SYRINGE, ASPIRATING, Ear and Ulcer, Size 1

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:57Z
Check the source to verify that this is the currentversion before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS

Project No. 6515-3756

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.