A-A-51508A
27 April 1987
SUPERSEDING
A-A-0051508 (DM)
18 June 1986

COMMERCIAL ITEM DESCRIPTION

SYRINGE, IRRIGATING (Ear and Ulcer)

The General Services Administration has authorized the use of this Commercial Item Description.

This commercial item description covers a sterile, disposable, plastic syringe used to irrigate the eye, ear and ulcers.

Salient characteristics.

General. Shall be a sterile, disposable syringe used in irrigation of the eyes, ears and ulcers.

Material. The syringe shall be manufactured entirely from flexible polyvinyl chloride. The syringe shall meet the requirements for Class II plastics as stated in the United States Pharmacopeia (USP).

Characteristics. The syringe shall have a nominal capacity of 3 ounces. The syringe shall have a tapered tip. The syringe shall have molded rings designed to provide a secure grip. The syringe tip shall have a length of 2-1/2 +1/2 - 1/8 inches. The tapered tip shall be free from any restrictions or occlusions. The syringe shall create a vacuum of 76 mm Hg (minimum).

Hardness. The syringe shall have a Shore durometer hardness of 55-63A when tested in accordance with ASTM D2240. Using ASTM 2240, the gage reading shall be made 10 seconds after application of the instrument.

Vacuum test. Manually evacuate the air from the syringe by compressing the top of the bulb until it touches the opposite wall. Attach the syringe to a mercury filled "U" tube with one foot of 1/4 inch I.D. flexible tubing. The syringe shall create a vacuum of 76 mm Hg (minimum)

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:57Z
Check the source to verify that this is the current version before use.

Sterilization, safety and freedom from pyrogens. The syringe shall be sterile, free from pyrogens and safe. The syringe shall be tested for sterility, freedom from pyrogens and safety in accordance with the USP.

Certificates of biocompatibility. A certificate of compliance, stipulating that offered items meet the biocompatibility test, shall be acceptable, provided that the contractor has supplied acceptable items to the Government within the past three years and that there have been no changes to material formulation and method of processing since actual testing was performed. The certificate shall include the date the testing was performed, the test method and results, the name and title of the person evaluating the results, the current contract number and NSN, a statement certifying that material processing and formulation have not changed since the time that biocompatibility was performed and the signature and title of the person making the certification.

Workmanship. The aspirating syringe set shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing fifty syringes of one size only constitutes one unit. Each syringe shall be packaged sterile in a sealed peel-open type container capable of maintaining sterility of contents.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:57Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one syringe. Sampling shall be S-2, AQL 2.5.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-045-0029	SYRINGE, IRRIGATING, Ear and Ulcer

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 46922

Project No. 6515-3867

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.