A-A-51509A
11 April 1990
SUPERSEDING
A-A-51509
23 June 1986

COMMERCIAL ITEM DESCRIPTION

TUBE, TRACHEAL, Radiopaque (Murphy)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers Murphy-type tracheal tubes commonly used in human patients in the practice of anesthesiology and in respiratory care.

Salient characteristics:

General. Except as specified below, shall be in accordance with American National Standard for anesthetic equipment-tracheal tubes, ANSI Z79.14. The tracheal tubes shall be the Murphy-type and shall be provided with a tracheal cuff. The tracheal tubes shall be supplied sterile and they shall be disposable.

Material. The tracheal tube, tracheal tube cuff, the inflation valve, and the pilot balloon shall be fabricated of a polyvinyl chloride plastic material which shall conform to the United States Pharmacopeia (USP) requirements for Class IV plastics. The poppet shall be fabricated from neoprene. If a valve spring is used, it shall be fabricated of AISI 300 Series corrosion-resistant steel. The Luer fitting on the inflation valve shall be in accordance with ANSI/HIMA MD70.1, Standard for Medical Materiel Luer Taper Fittings. Each tracheal tube shall be provided with a 15 mm tracheal tube straight connector.

The connector shall be fabricated of polypropylene and shall comply with the requirements of Z79.2, American National Standard for tracheal tube connectors and adapters. The tracheal tubes shall have a radiopaque stripe extending the entire length of the tube.

Leakage. The cuff and inflation system shall show no evidence of leakage when it has been inflated to an air pressure of 22 mm Hg for not less than 30 minutes.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Lead content. The tracheal tube and cuff shall contain not more than 5 parts per million (ppm) of lead.

Sterility. The tracheal tube and connector shall be sterile in accordance with the USP or AAMI guidelines. Sterility Assurance level shall be .000001.

Workmanship. Tracheal tubes shall be uniform in quality and free from any defects which affect the appearance or utility of the tube. Tracheal tubes shall be free from foreign matter, embedded particles or air, burrs, flashing, rough or sharp areas, cracks and other defects.

Unit. Package (PG). One package containing 10 tracheal tubes of one size only constitutes one unit. The tubes shall be individually packaged sterile in a peel-open type container capable of maintaining the sterility of the contents unless the package is damaged or opened.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for tests purposes shall be one tracheal tube. Sampling shall be Inspection level S-3 with an AQL of 2.5 (percent defective).

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards and quality assurance practices. The Government reserves the right to require the proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:57Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities, labels and markings shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers, their respective sizes and item identification are applicable to this document:

NSN	Size	Item Identification
6515-00-105-0653	3.0	Tube, Tracheal, Radiopaque, Murphy, Disposable, 3.0 mm I.D., 10s.
6515-00-105-0664	4.0	Tube, Tracheal, Radiopaque, Murphy, Disposable, 4.0 mm I.D., 10s.
6515-00-105-0707	5.0	Tube, Tracheal, Radiopaque, Murphy, Disposable, 5.0 mm I.D., 10s.
6515-00-105-0720	6.0	Tube, Tracheal, Radiopaque, Murphy, Disposable, 6.0 mm I.D., 10s.
6515-00-105-0744	7.0	Tube, Tracheal, Radiopaque, Murphy, Disposable, 7.0 mm I.D., 10s.

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:57Z
Check the source to verify that this is the current version before use.

6515-00-105-0759	8.0	Tube, Tracheal, Radiopaque, Murphy, Disposable, 8.0 mm I.D., 10s.
6515-00-105-0808	9.0	Tube, Tracheal, Radiopaque, Murphy, Disposable, 9.0 mm I.D., 10s.
6515-00-105-0815	10.0	Tube, Tracheal, Radiopaque, Murphy, Disposable, 10.0 mm I.D., 10s.
6515-01-036-9034	7.5	Tube, Tracheal, Radiopaque, Murphy, Disposable, 7.5 mm I.D., 10s.
6515-01-036-9035	8.5	Tube, Tracheal, Radiopaque, Murphy, Disposable, 8.5 mm I.D., 10s.

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4772

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.