A-A-51511B
24 March 1989
SUPERSEDING
A-A-51511A
30 January 1987

COMMERCIAL ITEM DESCRIPTION

CABINET, STORAGE, X-RAY FILM

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a metal cabinet with light limit switch for storage of x-ray film.

Salient characteristics:

The x-ray film storage cabinet shall be constructed of reinforced welded sheet metal material. The cabinet shall accommodate a minimum of six cartons of x-ray film (75 or 100 sheet cartons) size 14 inches by 17 inches, or smaller. The cabinet shall sit evenly on a flat surface and shall have the following dimensions:

- Overall width: 18-1/4 inches plus or minus 1 inch.
- Overall depth: 15-1/2 inches plus or minus 1 inch.
- Overall height: 32-9/16 inches plus or minus 1 inch.

The cabinet shall have a light limit switch with necessary actuator, contacts, or adjustments to protect the film from white room light exposure when drawer is even slightly opened. The limit switch shall be rated for use on 125 volts and approximately 15 amps. Installation data shall be supplied if light limit switch assembly, or adjustment, is necessary.

The cabinet shall be suitably protected against corrosion by means of a baked enamel paint, epoxy coating, or plated finish. All cabinet parts which are not inherently corrosion resistant shall be finished. All parts shall be properly cleaned or primed prior to finishing. The cabinet finish shall be uniform and free of coarse grind marks and scratched areas. The cabinet shall be the manufacturer's standard color.

AMSC N/A FSC 6525

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

All surfaces to be welded shall be free from grease, oil, dirt, or other foreign material. All weld joints shall blend into the adjacent metal in gradual, smooth curves. Beads shall be smooth and free of slag or excessive spatter. Sufficient metal shall be added to provide a suitable fillet or reinforcement. All welds shall be continuous with no visible voids, pin holes, pitting, cracks, or discoloration.

Cabinet fabrication shall be first class throughout. All exposed corners and edges shall be broken, rounded, or otherwise treated to eliminate sharp edges. The cabinet shall be free from oil, grease, dirt, or other foreign material.

Workmanship. The x-ray film storage cabinet shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One cabinet, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 4.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one x-ray film storage cabinet, as specified in this description.

Defects.

Cabinet is not constructed of welded sheet metal.

Cabinet does not accommodate minimum of six cartons of 14 inches by 17 inches x-ray films.

Cabinet does not have a baked enamel paint, epoxy coating, or plated finish.

Cabinet does not sit evenly on a flat surface.

Welds are not continuous, not clean, or not free of pitting, splatter, or discoloration.

Exposed or accessible edges, corners, or surfaces are not broken, rounded, or otherwise treated to eliminate sharp edges.

Finish is not uniform and free of course grind marks and scratched areas.

Cabinet is not clean and free of oil, grease, dirt or other foreign material.

Sampling for tests. Sampling for tests (dimensional) shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-3. The unit of product for test purposes shall be one x-ray film storage cabinet.

a. First article. Where a first article sample is required, all tests shall be performed on that unit.

b. Acceptance testing. When a first article sample has not been required, one unit shall be subjected to the performance test specified herein. If a test article sample has successfully passed all tests, performance test shall not be required at acceptance.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Tests.

Limit switch performance. The x-ray film storage cabinet shall be placed on a level surface with the light limit switch and switch contact correctly adjusted for use. The limit switch shall be connected to a suitable lamp which shall be "ON" while the cabinet drawer is completely closed.

The drawer shall then be cracked open slightly. Failure of the lamp to go "OFF" shall constitute a defect. The drawer shall then be closed completely. Failure of the lamp to go "ON" shall constitute a defect.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:57Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6525-01-115-9184

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6525-1253

Location:
CID/CABINET/STORAGE/H-13-16

Source: https://assist.dla.mil -- Downloaded: 2026-04-26T08:57Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-51511B	CABINET, STORAGE, X-RAY FILM

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)*
b. ADDRESS *(Street, City, State, ZIP Code)*	

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge (DO NOT STAPLE), mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4956 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC- RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.