A-A-51520B
27 January 1992
SUPERSEDING
A-A-51520A
11 February 1988

COMMERCIAL ITEM DESCRIPTION

SCREW, BONE, CANCELLOUS

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers cancellous bone screws suitable for scaphoid and routine large cancellous fragment procedures.

Salient characteristics:

General. Each bone screw shall have a major thread diameter of 6.5 mm. Shaft diameter shall be 4.5 mm. Core diameter shall be 3.0 mm. Screw shall be in accordance with ISO 5835/1.

Each bone screw shall have the following lengths:

Type	Length
1	50 mm
2	55 mm
3	60 mm

Each bone screw shall have an effective thread length of 32 mm. Thread profile shall be in accordance with ISO 5835/1 for deep, code HB 6.5 thread.

Tolerances shall be in accordance with ISO 5835/1.

Material. Bone screws shall be fabricated from cold drawn wire conforming to the composition specified in ASTM Designation F138, Grade 2.

Length. Type 1 bone screws shall have a length of 50 mm. Type 2 shall have a length of 55 mm. Type 3 shall have a length of 60 mm.

Surface preparation. Description of surface characteristics and methods of surface preparation shall be in accordance with ASTM Designation F86.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Finish. The completed bone screws shall have a mirror finish, properly passivated.

Corrosion properties. Bone screws shall show no evidence of corrosion when subjected to the following tests:

Intergranular corrosion. The intergranular corrosion susceptibility test specified in ASTM Designation F138.

Boil. Bone screws shall be scrubbed with soap and water, rinsed throughly in distilled water, and dried. After cleaning, the instruments shall be boiled in distilled water in a glass beaker for 30 minutes, allowed to remain submerged for an additional 24 hours.

Workmanship. The bone screws shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Each (EA). One bone screw of the diameter specified, consitutes one unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one bone screw. Sampling shall be inspection level II with an acceptable quality level (AQL) of .065.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one bone screw. Sampling shall be inspection level S-2 with an AQL of .065.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers'own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE 1: Copies of International Standards (ISO) may be obtained from the American National Standards Institute (ANSI), 11 West 42nd Street, New York, NY, 10036.

NOTE 2: The following National Stock Numbers are covered by this document:

NSN	TYPE	ITEM IDENTIFICATION
6515-01-168-0109	1	SCREW, BONE, Cancellous, Spherical Head, Hexagonal Socket, 6.5 mm Thread Diameter by 50 mm Length
6515-01-168-0110	2	SCREW, BONE, Cancellous, Spherical Head, Hexagonal Socket, 6.5 mm Thread Diameter by 55 mm Length
6515-01-168-3791	3	SCREW, BONE, Cancellous, Spherical Head, Hexagonal Socket, 6.5 mm Thread Diameter by 60 mm Length

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

PGC 06808

Project No. 6515-5057
Location: c:\word\robert\screw.doc

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-51520B	27 January ~~1992~~

3. DOCUMENT TITLE
SCREW, BONE, CANCELLOUS

4. NATURE OF CHANGE (*Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.*)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)
	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center ATTN: DPSC-MSE	(215) 737-2870	444-2870
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

Source: https://assist.dla.mil - Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.