A-A-51521
7 July 1986

COMMERCIAL ITEM DESCRIPTION

SEPARATING FLUID, DENTAL

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a liquid separating medium for use as an interface for easy separation of plaster, stone, investment and other dental gypsums.

Salient characteristics:

Shall be a liquid separating medium capable of producing, upon a single application, clean and easy separation of plaster, stone, investment and other dental gypsums at a desired interface. Shall leave no discernable film. Shall spread evenly and be easily applied by brush or applicator. Shall dry within 30 seconds so that plaster, stone or investment can be poured against it. The dry separating medium layer shall be a thin, hard, non-peeling coating. The dry separating medium layer shall be insoluble to water and create an effective moisture barrier at the interface. The separating medium shall be a minimum of eight (8) fluid ounces (236.6 ml) and shall be supplied in a resealable container.

Material. Shall be composed of the following ingredients:

Petrolatum Amber
Stearic Acid
Isopropyl Alcohol
Naptha
D&C Red #17

Shall have a viscosity of 3.5-6.5 centipoises. Shall be orange in color.

Workmanship. The liquid separating medium shall be free of particulates or sediment or any defect which detracts from its appearance or impairs its serviceability.

Unit. Bottle (BT). One bottle containing 8 fluid ounces (236.6 ml), minimum, of liquid separating medium with appropriate instructions for use, as specified, constitutes one unit.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:57Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one eight-ounce bottle of liquid separating medium. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local stautes, ordinances, and regulations.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6520-01-153-6067 SEPARATING FLUID, DENTAL, 8 fl oz (236.6 ml).

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

Civil Agency Coordinating Activity:
VA-OSS

52052

Project No. 6520-1936

Location: NCR

☆U.S. GOVERNMENT PRINTING OFFICE: 1986 — 605-036/44829

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:57Z
Check the source to verify that this is the current version before use.