A-A-51525A
18 June 1990
SUPERSEDING
A-A-51525
30 June 1986

COMMERCIAL ITEM DESCRIPTION

CENTRIFUGE, LABORATORY (Small)

The General Services Administration has authorized the use of this commercial item description in preference to Military Specification MIL-C-37040 which was cancelled.

This commercial item description covers a table model centrifuge with a six or eight place fixed angle or horizontal head for 15 milliliter tubes, suitable for use in military medical laboratories having limited space and where the character of work does not require centrifuging large volumes.

Salient characteristics.

Shall be a table model centrifuge suitable for small volume centrifuging on 115 volt, 50/60 Hz, AC supply lines in military facilities having limited space.

Shall consist of a stainless steel or aluminum base, guard bowl and hinged cover, universal type motor, speed control device, and line cord.

Shall be supplied with a detachable six (6) or eight (8) place fixed angle or horizontal head for 15 milliliter (ml) tubes and a 15 ml plastic shield or corrosion-resisting steel shield with rubber cushion for each trunnion ring or fixed angle position.

Shall be capable of producing a relative centrifugal force (RCF) of not less than 1470 gravities and not more than 2000 gravities. The RCF shall be tested with 15 ml tubes fully loaded to capacity and measured at the tips of the tubes with the speed control setting at maximum.

The base shall support the guard bowl, and shall house the motor and speed control device. The bottom of the base shall be provided with flexible rubber feet for cushioning and stability.

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

The guard bowl shall provide a chamber for the detachable head, and tubes that are inserted for horizontal centrifuging. The top view of the guard bowl shall be circular, and the outer diameter shall not exceed 15-1/2 inches. The interior of the bowl shall be finished with a durable baked-on enamel. The cover shall be hinged to the bowl. The center of the cover shall have a cutout (hole) which allows for safe use of a speed counter device (which is not required to be supplied).

The motor shall be a universal speed type that operates on 115 volt, 50/60 Hz, AC supply lines. The head shall be capable of being secured to the motor shaft for a motor shaft adapter.

The speed control device shall be an autotransformer or rheostat which provides for six or seven steps of speed variation. The speed selection or variation shall be through use of an accessible selector knob or handle control that includes an "off" position.

The overall height of the centrifuge with closed cover shall not exceed 13 inches.

The line cord shall be a three-wire electrical cord that terminates with a non-detachable 3-blade male plug that is suitable for use with receptacles specified in American National Standards Institute (ANSI) C73-11 (Dimensions of Plugs and Receptacles - General Purpose, 125 Volts, 15 Amperes, 2-Pole, 3-Wire, Grounding Type). The exposed length of the line cord shall be not less than five feet.

Workmanship. The centrifuge shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Each (EA). One centrifuge, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for tests. Sampling for tests shall be conducted in accordance with commercial practice. The unit of product for tests purposes shall be one centrifuge.

Tests. Tests shall be conducted to determine compliance with specification requirements.

Operation test. An operational check shall be performed on all centrifuges to demonstrate that the centrifuges and their controls are capable of proper operations.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/ contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6640-00-930-9034

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6640-1613

Location: CID/CENTRIFUGE/S31/12

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-51525A	CENTRIFUGE, LABORATORY, (SMALL)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil - Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC- RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.