A-A-0051530 (DM)
10 July 1986

INTERIM COMMERCIAL ITEM DESCRIPTION

ANESTHESIA SET, SPINAL

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA, 19101, based upon currently available technical information.

This interim commercial item description covers a spinal anesthesia set suitable for anesthetizing patients for surgical procedures.

Salient characteristics.

General. The spinal anesthesia set shall be sterile, disposable and suitable for use in anesthetizing patients for surgical procedures.

The set shall consist of the following items as a minimum:

One 25 gauge by 3-1/2 inch (9 cm) needle with introducer.

One 3 ml ampul 10% Dextrose Injection, USP.

One 2 ml ampul 1% Procaine Injection, USP, with directions for use.

One 1 ml ampul Ephedrine Sulfate Injection, USP, 50 mg/ml.

One 1 ml ampul 0.1% Epinephrine Injection, USP.

One 2 ml ampul 1% Tetracaine, USP, Sterile.

One 2-1/2 ml Syringe (standard Luer tip without lock) with 25 gauge by 5/8 inch (1.6 cm) needle.

One 5 or 6 ml Syringe (standard Luer tip without lock) with 20 or 21 gauge by 1-1/2 inch (3.8 cm) needle.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

One fenestrated waterproof drape, and prep sponges, gauze sponges and towels, sufficient in capacity to prepare the patient for anesthesia.

Material. The glass ampuls shall be Type I or Type III, prescored, meeting the requirements of the Pharmacopeia of the United States (USP). The ampuls shall be made of molded glass or drawn glass tubing, with walls of uniform thickness with drawn, rolled-in, or stuck-on necks. Ampuls shall stand upright, at right angles to the supporting surface, without wobbling. Ampuls shall be supplied ready for opening and shall break cleanly.

The plastics shall meet the requirements of the USP for Class II plastics.

The needles shall be austenitic and in accordance with ASTM F899. The needles shall be furnished with protectors.

The set shall be packaged in a container capable of maintaining the sterility of the tray and contents until manually opened. The tray shall be ready for use and have all new components in sequence lined packaging.

The anesthesia set shall be furnished with no less than 12 months to the expiration date at the time of delivery to the Government.

Sterility and pyrogenicity. Components shall be sterile and free from pyrogens. The set shall have a minimum sterility assurance level of 0.0001.

Workmanship. The anesthesia set shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One anesthesia set, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. The anesthesia set shall be examined to determine compliance with specification requirements. Sampling shall be in accordance with MIL-STD-105, sampling level II, AQL 2.5.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one anesthesia set. Sampling shall be at an inspection level of S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-00-491-4909	ANESTHESIA SET, SPINAL, DISPOSABLE

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS

PGC 59494

Project No. 6515-3779

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:57Z
Check the source to verify that this is the current version before use.