A-A-51530B
21 December 1989
SUPERSEDING
A-A-0051530A(DM)
15 July 1987

COMMERCIAL ITEM DESCRIPTION

ANESTHESIA SET, SPINAL

The General Services Administration has authorized the use of this interim commercial item description.

This commercial item description covers two spinal anesthesia sets suitable for anesthetizing patients for surgical procedures.

Salient characteristics:

The spinal anesthesia sets shall be sterile, disposable and suitable for use in anesthetizing patients for surgical procedures.

Type I shall contain one 2 ml ampul 1% Tetracaine Hydrochloride, USP, sterile.

Type II shall contain lyophilized Tetracaine Hydrochloride, USP, to yield 2 ml at 1%.

The set shall consist of the following items as a minimum:

One 25 gauge by 3-1/2 inch (9 cm) spinal needle with introducer.

One 5 ml ampul 10% Dextrose Injection, USP.

One 2 ml ampul 1% Procaine Injection, USP, with directions for use.

One 1 ml ampul 5% Ephedrine Sulfate Injection, USP.

One 1 ml ampul 0.1% Epinephrine Injection, USP.

One 2-1/2 ml Syringe (standard Luer tip without lock) with 25 gauge by 5/8 inch (1.6 cm) needle.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:57Z
Check the source to verify that this is the current version before use.

One 5 or 6 ml Syringe (standard Luer tip without lock) with 20 or 21 gauge by 1-1/2 inch (3.8 cm) needle.

One 5 micron procedural filter.

One fenestrated waterproof drape, and prep sponges, gauze sponges and towels, sufficient in capacity to prepare the patient for anesthesia.

Material. The glass ampuls shall be Type I or Type III, prescored, meeting the requirements of the Pharmacopeia of the United States (USP). The ampuls shall be made of molded glass or drawn glass tubing, with walls of uniform thickness with drawn, rolled-in, or stuck-on necks. Ampuls shall stand upright, at right angles to the supporting surface, without wobbling. Ampuls shall be supplied ready for opening and shall break cleanly.

The plastics shall meet the requirements of the USP for Class II plastics.

The needles shall be austenitic and in accordance with ASTM Designation F899. The needles shall be furnished with protectors.

The components of the set shall be within opaque packaging capable of maintaining the sterility of the tray and contents until manually open.

The anesthesia sets shall be furnished with no less than 18 months to the expiration date at the time of delivery to the Government.

Sterility. Components shall be sterile and have a minimum sterility assurance level of $1X10^{-6}$.

Workmanship. The anesthesia sets shall be free from defects which detract from its appearance or impair its serviceability. Needles shall be without burrs or signs of corrosion.

Unit. Set (SE). One anesthesia set, each set packaged sterile, as specified, constitutes one unit. The package container shall be capable of maintaining sterility of contents unless the package is damaged or opened.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. The anesthesia sets shall be examined to determine compliance with specification requirements. Sampling shall be in accordance with MIL-STD-105, sampling level II, AQL 2.5.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one anesthesia set. Sampling shall be at an inspection level of S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 as amended, as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION	TYPE
6515-00-491-4909	ANESTHESIA SET, SPINAL, DISPOSABLE, with 1% Tetracaine HCl Solution, EA.	I
6515-01-156-2494	ANESTHESIA SET, SPINAL, DISPOSABLE, with dry Tetracaine HCl, EA.	II

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
MPQAS

Project No. 6515-4469

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-51530B	ANESTHESIA SET, SPINAL

3a. NAME OF SUBMITTING ORGANIZATION

b. ADDRESS *(Street, City, State, ZIP Code)*

4. TYPE OF ORGANIZATION *(Mark one)*

- ☐ VENDOR
- ☐ USER
- ☐ MANUFACTURER
- ☐ OTHER *(Specify)*: _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional

b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional

c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional

8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 6, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 Alexandria, VA.

POSTAGE WILL BE PAID BY

Commander
Defense Personnel Support Center
ATTN: DPSC-RST
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.