
FED. SUP CLASS
6640

USER ACTIVITIES: ARMY - ME

REVIEWER ACTIVITIES: ARMY - ME
NAVY - YD



PART NUMBER	CAPACITY *µL (λ) (±%)
MS35986-5	1 ± 1.0
MS35986-6	5 ± 0.5
MS35986-1	10 ± 0.5
MS35986-2	25 ± 0.5
MS35986-7	50 ± 0.3
MS35986-3	100 ± 0.3
MS35986-4	250 ± 0.2

*Microliter

MATERIAL: Borosilicate glass in accordance with Federal Specification DD-G-541, Type 1, Class A or B.

DESIGN: Micropipets shall be of the Kirk design with a tapered top (2.5 ± 0.5 mm od) and long tapered delivery tip with a safety bulb near the top end of the pipet.

Maximum outside diameter at the safety bulb shall be no more than 1 mm more than the maximum outside of the tubing.

Outside edge of each end shall be slightly ground and fire polished.

Micropipets shall be calibrated with mercury "to contain" at 20°C.

Marking of capacity in lambdas and calibrations shall be permanently etched or fused on the pipet.

NOTES:
1. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.
2. For design feature purposes, this standard takes precedence over procurement document referenced herein.
3. Referenced documents shall be of the issue in effect on date of invitations for bid.

(B) REDRAWN AND REVISED
(A) ENTIRE STANDARD REVISED

APPROVED 23 Dec 1959 REVISED (A) 14 July 1969 (B) 21 OCT 1983

P.A. EA	INTERNATIONAL INTEREST	TITLE PIPET, VOLUMETRIC, MICRO Kirk design	MILITARY STANDARD MS35986
Other Cust MS 82			
Procurement Specifications (B) ASTM E193	SUPERSEDES:		SHEET 1 OF 1

PROJECT NUMBER 6640-1304

DD FORM 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:32Z
Check the source to verify that this is the current version before use.