

FED. SUP CLASS
6640

REVIEW ACTIVITIES: ARMY-CE
AIR FORCE - 82
NAVY-AS





DIMENSIONAL DATA					
MORTAR DATA				PESTLE DATA	
PART NUMBER	TOP OD mm (± 5%)	HEIGHT mm (± 5%)	CAPACITY ml (± 10%)	PART NUMBER	LENGTH mm (± 5%)
MS36103-1	90	58	135	MS36103-5	135
MS36103-2	130	80	400	MS36103-6	190
MS36103-3	CANCELED INACTIVE			MS36103-7	CANCELED INACTIVE
MS36103-4	CANCELED INACTIVE			MS36103-8	CANCELED INACTIVE

MATERIAL: Mortar and pestle shall be of porcelain.

DESIGN: Mortar shall be bowl-shaped with a flat base.

Base diameter shall be large enough to provide a stable resting surface.

Top edge shall be flat with a pour-out.

Pestle shall be designed to afford hand comfort in grinding. The grinding end shall be rounded to facilitate grinding.

Mortar and pestle shall be in accordance with Federal Specification NNN-M-560, Type I, Class A.

FINISH: Mortar shall be glazed outside with exception of the bottom.

Pestle shall be glazed with the exception of the grinding surface.

NOTES:
1. For design feature purposes, this standard takes precedence over procurement documents referenced herein.
2. Referenced documents shall be of the issue in effect on date of invitations for bid.

(A) ENTIRE STANDARD REVISED

APPROVED 22 SEPT 1960 REVISED (A) 31 MAY 1987

This military standard is approved by the Department of Defense and is mandatory on all activities. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

P.A. MU Other Cust MS 82	TITLE MORTAR AND PESTLE, LABORATORY porcelain	MILITARY STANDARD MS36103	
Procurement Specification NNN-M-560	SUPERSEDES:	SHEET 1 OF 1	

DD FORM 1 SEP 63 672 (Coordinated)

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:34Z
Check the source to verify that this is the current version before use.