MILITARY SPECIFICATION

ANESTHESIA SET, EPIDURAL, DISPOSABLE

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a sterile, disposable continuous epidural anesthesia set suitable for use in obstetrical and general surgical procedures. Potency is 36 months, which is determined by the shortest dated item (Epinephrine Injection) in the set.

2. APPLICABLE DOCUMENTS

2.1 Specifications and standards. The following documents of the issue in effect on date of invitation for bids or request for proposals form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

GG-I-526	Instruments, Dental and Surgical; General Specifications.
GG-N-196	Needles, Hypodermic, for Luer Syringes.
PPP-B-585	Box, Wood, Wirebound.
PPP-B-601	Box, Wood, Cleated-Plywood.
PPP-B-621	Box, Wood, Nailed and Lock-corner.
PPP-B-636	Box, Fiberboard.

Military

MIL-L-10547	Liners, Case and Sheet Overwrap; Water-Vaporproof, or Waterproof, Flexible.
MIL-N-36124	Needle, Hypodermic, Disposable.
MIL-S-36157	Syringes, Hypodermic, Disposable.
MIL-N-36214	Needle, Hypodermic, Spinal Anesthesia.
MIL-S-36586	Sterilization, Test Strip Set, Bacterial Spore.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:59Z
Check the source to verify that this is the current version before use.

STANDARDS

Federal

Fed. Test Method STD-191 Textile Test Methods.

Military

MIL-STD-105 Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 Marking for Shipment and Storage.

(Copies of specifications and standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals, shall apply.

THE U. S. PHARMACOPEIAL CONVENTION, INC.

The Pharmacopeia of the United States of America (USP)

(Application for copies should be addressed to the Mack Publishing Company, Easton, Pa. 18042).

U. S. DEPARTMENT OF HEALTH, EDUCATION, AND WELFARE
Bureau of Biologics. Food and Drug Administration
Code of Federal Regulation Part 273.
Minimum Requirements (as applicable).

(Application for copies should be addressed to the Food and Drug Administration, 5600 Fishers Lane, Rockville, Maryland 20852.)

AMERICAN NATIONAL STANDARDS INSTITUTE, INC.

A.N.S.I. Z70.1 American Standard Dimensions of Glass and Metal Luer Tapers of Medical Applications.

(Application for copies should be addressed to the American National Standards Institute Inc., 1430 Broadway, New York, N.Y. 10018).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:59Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 Material.

3.1.1 Plastic components. All plastic components shall be of the best quality and suitable in all respects for the intended purpose. All plastic components shall be sterilizable, free from pyrogens, compatible, and safe (nontoxic) when tested as specified in Section 4. Those plastic components coming in contact with solutions being administered, and end protectors, shall contain not more than 5 parts per million (p.p.m.) of lead when tested as specified in 4.4.1. The plastics shall meet the requirements of the Pharmacopeia of the United States (USP) for Class II plastics and for transfusion and infusion assemblies, when tested as specified in 4.4.2.

3.1.1.1 Needle adapter. The needle adapter shall be fabricated from a suitable rigid plastic.

3.1.1.2 End protectors. End protectors shall be fabricated from plastic tubes or sheaths.

3.1.2 Needles.

3.1.2.1 Cannulae of needles. Cannulae of needles shall be fabricated from 304 AISI stainless steel, and shall meet the corrosion resistance, elasticity, and resistance-to-breakage tests as specified in GG-N-196. The cannulae shall be fabricated from either seamless drawn, or welded and drawn tubing.

3.1.2.2 Hubs. Hubs shall be class 3, type 302, 303, or 304, or class 8, alloy number 260, 270, or 360 of GG-I-526; or hubs shall be plastic. Class 8 (brass) hubs shall be nickel plated.

3.1.3 Other material. All other materials as specified herein shall be of the best quality and suitable in all respects for the purpose intended.

3.2 Style, design and construction. Shall be a continuous epidural anesthesia set of sterile items consisting of two compartmented plastic trays, one containing drugs and the other preparation supplies, as follows:

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:59Z
Check the source to verify that this is the current version before use.

Drug Tray

Component	Quantity required
Needle, 25 gage by 5/8 inch with Luer hub	1
Needle, 22 gage by 1-1/2 inch with Luer hub	1
Needle, 18 gage by 1-1/2 inch with Luer hub	1
Needle, 18 gage T.W. by 3-1/2 inch modified Husted with Luer hub	1
Syringe, plastic, 1cc	1
Syringe, plastic, 2-1/2 cc	1
Syringe, glass, with metal Luer-Lock, 10cc	1
Cup, medicine, with cover	1
Adapter, catheter, 23 gage by 5/8 inch	1
Catheter, epidural, plastic, radiopaque	1
Catheter pad, adhesive	1
Epinephrine Injection, USP, 1:1000, 1 cc ampul	1
Sodium Chloride Injection, USP, 5 ml	

Preparation Tray

Component	Quantity required
Sponge, 3 by 3 inch	4
Towel	2
Drape, moisture resistant, fenestrated	1
Cup, antiseptic solution	1
Swab with handle	3

3.2.1 Needle, 25 gage. Shall be in strict accordance with MIL-N-36124 for size 7, except the packaging requirements of that specification shall not apply. As an alternate, the stylus knob shall be fabricated from a suitable plastic.

3.2.2 Needle, 22 gage. Shall be in strict accordance with GG-N-196 Type I, Style A, Size 1-1/2 inches long except the packaging requirements of that specification shall not apply.

3.2.3 Needle, 18 gage. Shall be in strict accordance with MIL-N-36124 for size 1, except the packaging requirements of that specification shall not apply. As an alternate, the stylus knob shall be fabricated from a suitable plastic.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:59Z
Check the source to verify that this is the current version before use.

3.2.4 Needle, 18 gage T.W. Shall be a 18 gage Thinwall by 3-1/2 inch modified Husted needle with Luer hub. The hub shall meet the requirement and test for hubs as specified in MIL-N-36214 except material for hub shall be plastic.

3.2.5 Syringe, plastic, 1 cc. Shall be a 1 cc disposable plastic syringe. The material, style, design, dimensions and construction shall be in accordance with MIL-S-36157 for Type II, size 1, except that packaging requirements of that specification shall not apply.

3.2.6 Syringe, plastic, 2-1/2 cc. Shall be a 2-1/2 cc disposable plastic syringe. The material, style, design, dimensions and construction shall be in accordance with MIL-S-36157 for Type II, size 2, except that packaging requirements of that specification shall not apply.

3.2.7 Syringe, glass, with metal Luer Lock, 10 cc. Shall be a 10 cc glass syringe. The style, design and dimensions shall be in accordance with MIL-S-36157 for Type II, size 4, except that packaging requirements of that specification shall not apply.

3.2.8 Cup, medicine. Shall be a medicine cup with cover used for mixing drugs.

3.2.9 Adapter, catheter. A plastic 23 gage by 5/8 inch catheter adapter shall be provided. Shall meet the leakage test specified in 4.4.7.

3.2.10 Catheter, epidural, plastic, radiopaque. Shall be a radiopaque polytetrafluorethylene epidural catheter. Shall be marked at 1 cm intervals for a distance of 5 cm at the needle hub end, with a metal stylet.

3.2.11 Catheter pad. A catheter pad with adhesive back shall have a slit for holding catheter.

3.2.12 Epinephrine injection. Shall be Epinephrine Injection, USP, 1:1000, 1 cc. The epinephrine injection shall be in accordance with the requirements of the current specification (Defense Medical Purchase Description) for NSN 6505-00-299-8760, except as follows:

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:59Z
Check the source to verify that this is the current version before use.

The designation "25s" shall not apply.

Page 2 of Purchase Description:

The name and address of the manufacturer need only appear on the ampul, providing the contractor's name and address appear on the anesthesia set.

Page 3 of Purchase Description:

The labeling of the unit packages shall not apply, except for the statement in subparagraph (h) which shall appear on the anesthesia set.

The circular, brochure, or other printed matter shall be furnished in each anesthesia set.

Page 4 of Purchase Description:

Packaging, Packing, and Marking shall not apply.

3.2.13 Sodium chloride injection. Shall be sodium chloride injection, USP, 5 cc. Shall be in accordance with the requirements of the current specification (Defense Medical Purchase Description) for NSN 6505-00-559-8456, except as follows:

The designation "25s" shall not apply.

Page 2 of Purchase Description.

The name and address of the manufacturer need only appear on the ampul, providing the contractor's name and address appear on the anesthesia set.

Page 3 of Purchase Description:

The labeling of the unit packages shall not apply.

Page 4 of Purchase Description:

Packaging, Packing, and Marking shall not apply.

3.2.14 Sponge. The gauze sponge shall be 12 ply, approximately 3 inches by 3 inches when folded.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:59Z
Check the source to verify that this is the current version before use.

3.2.15 Towel. Towel shall be a white cellulose wipe, minimum of 3 ply, approximately 13 inches by 19 inches.

3.2.16 Drape. The drape shall be moisture resistant, and fenestrated, approximately 22-1/4 inches by 36 inches. The drape shall meet the requirement for waterproof when tested in accordance with 4.4.3.

3.2.17 Cup, antiseptic solution. A cup shall be provided for use of antiseptic solution.

3.2.18 Swab with handle. The swab shall be foam sponge, and shall be approximately 1" x 1-1/2" x 1/2". Shall be securely mounted on a durable handle.

3.2.19 Sterilization indicator. A sterilization indicator showing a color change when sterilized shall be provided.

3.3 Sterility. All components of the set shall be sterile. When tested with a biological indicator as specified in 4.4.4, the needles, syringes, medicine cup and cover, adapter, catheter, catheter pad, sponges, towels, drape, solution cup and swabs shall be sterile.

3.3.1 Certificates of sterility. A certificate of sterility, stating that samples from each sterilizer lot have been tested and found to be sterile, shall be submitted to the procuring activity. The certificate of sterility shall include:

a. Item identification.
b. National stock number.
c. Contract or purchase order number.
d. Sterility lot or control number.
e. Date of sterilization.
f. Quantity of product in sterilizer lot.
g. Type of sterilization (gas, radiation, heat, steam, etc.)
h. Form of biological indicator (i.e., inoculated simulated product).
i. Test organism(s) and population density used.
j. Test medium, incubation temperature, incubation time, and date of tests.
k. Number of test units.
l. Statement that samples comply with sterility tests.
m. Signature of head of laboratory or designated qualified alternate.

When inspection is made at destination, the certificates of sterility shall be submitted by the contractor to the Directorate of Medical Materiel, ATTN: DPSC-ATQ, Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pa., 19101, and a copy forwarded with each shipment to the consignee. When inspection is made at source, certificate of sterility shall be furnished to the cognizant Government Assurance Representative for submission to the procuring activity.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:59Z
Check the source to verify that this is the current version before use.

3.3.2 Sterility lot. A sterility lot is that single quantity of product subjected to the same manufacturing operation and simultaneously sterilized in the same sterilizer chamber. Each unit package shall be clearly marked with a sterility lot (control) number and date of sterilization as specified in 5.3.1. Each sterility lot shall be traceable to the pertinent manufacturing lot(s).

3.4 Nontoxicity. When tested as specified in 4.4.5, needles, syringes, medicine cup and cover, adapter, catheter, catheter pad, sponges, towels, drape, solution cup and swabs shall comply with the requirements for safety (nontoxicity) of the USP.

3.5 Freedom from pyrogens. When tested as specified in 4.4.6, the fluid path, end connectors, and needles shall be pyrogen free.

3.6 Certificates of sterility, freedom from pyrogens, safety, and biological tests. A certificate listing all lot or control numbers shall state that representative samples from each lot or control have been tested and found to be sterile, free from pyrogens, nontoxic (safe), and have passed the biological tests. This certificate shall include the specific tests performed and the tests results. For destination inspection, certificates shall be submitted by the contractor directly to Defense Personnel Support Center, Directorate of Medical Materiel, Technical & Quality Assurance Division, 2800 South 20th Street, Philadelphia, Pa. 19101, Attention: Quality Assurance Branch. One copy of each shall be forwarded with each shipment to the consignee. For source inspection, certificates shall be furnished to the procuring activity via the government inspector.

3.7 Manufacturer's records. The manufacturer shall place a lot control number on the paper label, or on the printed carton, which can be verified against the manufacturer's records of all tests made available for inspection by any duly qualified medical officer of the Armed Forces, National Institutes of Health, any official of the Federal Food and Drug Administration, or any duly authorized Government representative. No shipment shall be made until all the tests have been completed, verified, and recorded in the permanent records of the manufacturer.

3.8 Compatibility. The materials of which the sets are manufactured shall not react with, or be acted upon by the liquids with which they will be in contact during normal use and operation, nor shall there be any materials incorporated in the set which will cause any deleterious effect when the set is used for the purpose intended.

3.9 Instructions. Each set shall be furnished with a list of the contents of the pack and procedural instructions for use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:59Z
Check the source to verify that this is the current version before use.

3.10 Workmanship. Workmanship shall be first class throughout. All materials, treatments, and procedures used in the fabrication and assembly of the sets shall be such that the finished sets shall not show any evidence of decomposition, deterioration, or other conditions which may adversely affect the appearance or serviceability of the sets at time of delivery or during storage.

3.11 Delivery. Not less than 33 months of the expiration dating (potency period) shall remain at time of delivery to the Government.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the plastic, metal, and glass, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:59Z
Check the source to verify that this is the current version before use.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels shall be as indicated in Table I. The unit of product for examination shall be one anesthesia set.

Table I. Sampling for Examination

Type of defect	Inspection level	AQL (percent defective)
Major	II	1.0
Minor	II	2.5

Inspection lot shall be the same as sterilization lot unless otherwise specified.

4.2.2 For tests.

4.2.2.1 For sterility test. Sampling for the sterility test shall be in accordance with United States Pharmacopeia (USP) for parenteral devices.

4.2.2.2 For pyrogen, and safety tests. Sampling procedures for pyrogen, and safety tests shall be conducted as indicated in the USP, for transfusion and infusion assemblies.

4.2.2.3 For lead test. Samples shall be taken from each lot of plastic as indicated in 4.4.

4.2.3 For other tests. Sampling for other tests shall be conducted in accordance with MIL-STD-105, and as indicated in Table II.

Table II. Sampling for Other Tests

Component (unit of product)	Characteristic	Req.	Test procedure	Insp. level	AQL (percent defective)
Plastic components	Material - Class II	3.1	4.4.1	**	*
	Biological test	3.1	4.4.2	**	*
Drape	Waterproof drape test	3.2.2	4.4.3	S-2	2.5
Set	Sterility	3.3	4.4.4	****	*
Set	Safety	3.4	4.4.5	****	*
Set	Freedom from pyrogen	3.5	4.4.6	***	*

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:59Z
Check the source to verify that this is the current version before use.

* Acceptance number shall be zero for all sample sizes.
** Each lot of product need not be biologically tested provided evidence is furnished that samples of the formulation and identical method of processing have previously been tested and that they are in accordance with USP requirements for class II plastics and that they are the same as the material being furnished.
*** Sampling for tests shall be on component lots.
**** Sterility, Safety, and Freedom from pyrogen shall be tested by sampling sterility lots.

4.2.4 Sample unit usage. Where possible, the sample unit shall be used for determination of two or more test characteristics.

4.3 Examination. The sets shall be examined to determine compliance with all requirements contained in this specification. Examination shall be conducted in accordance with MIL-STD-105 and the classification of defects specified in Table III. Examination however, shall not be restricted to the classified possible defects listed in Table III.

Table III. Classification of Defects

Categories	Defects
Critical	None defined.
Major	
101	Set incomplete, one or more components missing.
102	Plastic or metal components not smoothly finished; not free of fits, blisters, scratches, tears, cracks, nicks, bubbles, roughness, tackiness, bloom or crazing, as applicable.
103	Components not free of foreign matter.
104	Towels not free from defects such as cuts and tears.
105	Labeling of drugs not as specified.
106	Swab not securely mounted on handle.
Minor	
201	Packaging, packing and marking not as specified.
202	Instructions missing, incomplete, or illegible.
203	Poor workmanship (not resulting in a major defect).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21 09:59Z
Check the source to verify that this is the current version before use.

4.4 Tests.

4.4.1 Test for lead. A 2 gm composite sample shall be obtained from plastic components. The material shall be cut into small slivers, so that no thickness exceeds 3 mm., and eluted for 72 hours at 50° C., 24 hours at 70° C., or 60 minutes at 121° C. in two menstrua: 0.9% sodium chloride and cottonseed oil. Each gram of the material shall be covered with 5 milliliters (ml.) of the menstruum. The menstrua are collected, agitated after elution, and then submitted to spectographic analysis for lead. The lead content shall be in parts per million. Thus 1.0 ppm in elute is equivalent to extracting 1.0 mg. from 200 g test material dispersed in 1,000 ml of elution menstruum. Alternates to the above manner specified in determining lead content may be considered acceptable, provided the alternate method is formally submitted to the contracting officer for approval, prior to shipment of contracted items.

4.4.2 Biological tests. The plastics shall be tested in accordance with the USP for Class II plastics.

4.4.3 Waterproof drape test. The drape shall be tested for waterproof requirement in accordance with Test Method 5526 of Fed. Test Method Std. 191.

4.4.4 Sterility test. The anesthesia set shall be tested for sterility in accordance with the method specified in the USP for parenteral devices using inoculated product, inoculated simulated product, or paper spore strip (carrier). The testing for sterility shall be performed on at least needles, catheter, syringes, towels, sponges and swabs as representative of the sterility of all the set components. Product containing a biological indicator shall be marked or labeled so that it clearly differentiates from product intended for distribution. The incubation medium and minimum time in each case shall be soybean casein digest for 7 days for inoculated product, inoculated simulated product, or paper spore strip (carrier). The incubation temperature shall be as indicated in paragraph 4.4.4.1. The minimum number of test units, regardless of lot size, shall be 10, except for inoculated carrier which shall be 20 for Type I sterilization and 30 for Type II Sterilization including inoculated carrier and product test units. The inspection level shall be as indicated in Table II, Sampling Criteria for Tests.

4.4.4.1 Inoculated product or simulated product. For each type of sterilization, the organism used shall be as listed below:

Type of Sterilization	Organism	Incubation Temperature
Steam	B. stearothermophilus	55° - 65° C.
Dry Heat	B. stearothermophilus or	55° - 65° C.
	B. subtilis var. niger (or globigii)	30° - 35° C.
Gas	B. subtilis var. niger (or globigii)	30° - 35° C.
Radiation	B. pumilus	30° - 35° C.

The number of viable cells shall be sufficient to provide assurance that the product has been sterilized.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:59Z
Check the source to verify that this is the current version before use.

4.4.4.2 Inoculated carrier. Inoculated carriers (paper spore strips) if used shall be in accordance with MIL-S-36586. Other paper carriers shall be inoculated to be of equivalent resistance. The method of incubation and number of test units shall be as indicated in 4.4.4 Sterility test and 4.4.4.1 Inoculated product or simulated product. The method of sterilization will control the organism in the spore strip to be tested for as shown in 4.4.4.1 NOTE: When an inoculated carrier is chosen, product samples must also be tested as required by the USP.

4.4.4.3 Alternate method. Suppliers conducting sterility tests or designating lots in a manner other than that indicated herein shall submit proof that the alternate method provides equal assurance concerning the sterility of the product. However, all sterility tests shall be conducted using biological indicators. Requests for approval of any alternate method of testing for sterility shall be made to the Contracting Officer, and approval obtained before employment of the alternate method on a government contract.

4.4.5 Safety (nontoxicity) test. Samples of the finished sets shall be tested for safety, in accordance with the USP, for transfusion and infusion assemblies.

4.4.6 Freedom from pyrogens. Samples of the finished sets shall be tested for freedom from pyrogens, in accordance with the USP, for transfusion and infusion assemblies.

4.4.7 Fluid tightness test. With the spinal needle in a vertical position and a water pressure of 5 pounds per square inch applied for 10 seconds to the point of the cannula, the Luer adapter plug fit with the spinal needle shall show leakage not exceeding one drop.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be Level A, or C as specified (see 6.2).

5.1.1 Unit of issue. One set as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in Table IV shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table IV. Packaging Quantities

Unit package	Intermediate package	Exterior container
1	Not required	10

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:59Z
Check the source to verify that this is the current version before use.

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package. Components comprising each set shall be placed into two trays. Trays shall be recessed to receive and retain components in position (after overwrapping). Needles shall be furnished with suitable protectors. Both trays shall be contained in a durable overwrap and packed in a semirigid seamless plastic tub which shall maintain sterility until manually opened.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be Level A, B or C as specified (see 6.2).

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant, grade V3c. Closure and strapping shall be as specified in the appendix of the applicable box specification. Special requirements of applicable table shall apply for PPP-B-636 box.

5.2.1.2 Case liner. Each Level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Fiberboard box shall be waterproof in accordance with paragraph 30.4 of PPP-B-636.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification. Fiberboard box shall be closed by method II of PPP-B-636. Special requirements of applicable table shall apply for PPP-B-636 box.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:59Z
Check the source to verify that this is the current version before use.

5.3 Marking.

5.3.1 Unit package. Each unit package shall be marked in accordance with MIL-STD-129. Marking shall include type I shelf-life markings as specified in MIL-STD-129, lot (control) number, the word "STERILE" and the legend "Sterility guaranteed unless package is damaged or open" or similar statement. In addition, marking shall include the method of sterilization (G-Gas, R-Radiation, S-Steam, D-Dry Heat Sterilization). Example: A-4/74-G). Final sterilization shall be on or after indicated packaging date. Marking shall also include the legends: "DO NOT PERMIT TO FREEZE" and "STORE AT CONTROLLED ROOM TEMPERATURE (59° - 86°F.)."

5.3.2 Exterior container. Exterior container shall be marked as required for the unit package except the word "STERILE" and sterility legend shall not be shown.

6. NOTES

6.1 Intended use. The epidural anesthesia set is for use in obstetrical and general surgical procedures).

6.2 Ordering data. The procurement documents should specify the following:

(a) Title, number and date of this specificatcn.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 National Stock Number (NSN) coverage. This specification is intended to cover the following item appearing in the Federal Supply Catalog:

NSN	Item
6515-00-148-7106	ANESTHESIA SET, EPIDURAL, Disposable.

It does not include all types as indicated by the title nor all types which are commercially available.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1890

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:59Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-A-37108

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426

REPLACES EDITION OF 1 JAN 68 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:59Z
Check the source to verify that this is the current version before use.