MILITARY SPECIFICATION

CUFF, SPHYGMOMANOMETER, CONDUCTIVE

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers three sizes of conductive sphygmomanometer cuffs and contained compression (arm) bags suitable for use with aneroid and mercurial sphygmomanometers.

1.2 Classification. Sphygmomanometer cuffs shall be furnished in the following sizes:

(a) Size I, Cuff, Sphygmomanometer, Child, Conductive.
(b) Size II, Cuff, Sphygmomanometer, Adult, Conductive.
(c) Size III, Cuff, Sphygmomanometer, Adult, Conductive, Giant Size.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposals form a part of the specification to the extent specified herein.

SPECIFICATIONS

Federal

V-T-295 - Thread, Nylon.
PPP-B-566 - Boxes, Folding, Paperboard.
PPP-B-636 - Boxes, Shipping, Fiberboard.
PPP-B-676 - Boxes, Set-up.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:59Z
Check the source to verify that this is the current version before use.

Military

MIL-C-577	-	Cloth, Wind Resistant, Sateen, Cotton.
MIL-F-21840	-	Fastener Tapes, Hook and Pile, Synethetic.
MIL-T-43548	-	Thread, Polyester, Cotton Cornered.

STANDARDS

Federal

FED-STD-123	-	Marking for Domestic Shipment (Civilian Agencies).
FED-STD-151	-	Metal, Test Methods.
FED-STD-191	-	Textile Test Methods.
FED-STD-601	-	Rubber, Sampling and Testing.
FED-STD-751	-	Stitch, Seam and Stitching.

Military

MIL-STD-105	-	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	-	Marking for Shipment and Storage.

(Copies of specifications and standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer).

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

National Fire Protection Association (NFPA)

No. 56A Standard for the use of Inhalation Anesthetics

(Application for copies should be addressed to the National Fire Protection Association, 60 Batterymarch Street, Boston, MA 02110.)

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:59Z
Check the source to verify that this is the currentversion before use.

AMERICAN NATIONAL STANDARDS INSTITUTE (ANSI)

Z70.1 Glass and Metal Luer Tapers for Medical Applications, Dimensions of.

(Application for copies should be addressed to the American National Standards Institute, 10 East 40th Street, New York, New York 10018).

3. REQUIREMENTS

3.1 Fire and casualty hazards.

3.1.1 The rubber in the subject item shall comply with the fire and casualty hazards provisions of the applicable standards for operation in hazardous locations, Class 1, Group C and as prescribed by the National Fire Protection Association, (NFPA) No. 56, for conductive accessories, and be approved for the purpose intended, as determined by a nationally recognized testing agency, adequately equipped and competent to perform such services, as shall be evidenced by the issuance of a certificate. The certificate of the Underwriters' Laboratories, Inc., or other testing laboratory approved by the contracting agency shall be considered acceptable evidence of such approval; however, in addition to the aforementioned, the contracting agency reserves the right to require the contractor to submit a laboratory test report which shall include test results and methods used to determine conformance.

3.1.2 Compliance with the above requirements regarding fire and casualty hazards, does not absolve the contractor from complete compliance with the requirements of this specification in order to secure the acceptance of this commodity.

3.2 Material.

3.2.1 Metal parts. Metal parts shall be fabricated from stainless steel, aluminum, nickel-plated brass, or chromium over nickel plated brass.

3.2.2 Rubber.

3.2.2.1 Compression bag. Compression bag shall be fabricated of natural rubber containing not less than 80 percent of pure virgin latex suitably compounded, processed, or treated to render it conductive. Shall be of seamless construction.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:59Z
Check the source to verify that this is the current version before use.

3.2.2.2 Tensile strength. Tensile strength of the rubber in the compression bag shall be 3000 pounds per square inch minimum when tested according to 4.4.1. Tensile strength requirement of tubing integral to the bag (dipped latex tubing) shall not apply. All additional tubing shall have tensile strength of 1900 psi minimum. After accelerated aging (4.4.2), tensile strength shall not decrease by more than 25 percent. After aging, bag and tubing shall show no stiffening, softening or cracking.

3.2.3 Coated cloth.

3.2.3.1 Base material. Shall be type I cloth as specified in MIL-C-577 except the color requirements shall not apply.

3.2.3.1.1 Nylon. As an alternate, approximately 3.5 ounce/square yard nylon oxford cloth, 200 denier, 70/54 weave may be used. A 3/4 ounce/square yard urethane coating may be used. The cloth shall have a minimum tensile strength of 200 by 160 pounds per inch. Color shall be blue, gray or tan.

3.2.3.1.2 Polyester/cotton. As another alternate, polyester/cotton approximately 8.5 ounce/square yard, 104/38 weave may be used. The cloth shall have a minimum tensile strength of 150 by 90 pounds per inch. Color shall be blue, gray or tan.

3.2.3.2 Coatings.

3.2.3.2.1 Inner coating. Base material for the cuff proper shall have a neoprene inner coating of 0.002 $\pm$ 0.001 inch.

3.2.3.2.2 Outer conductive coating. Base material for the cuff proper shall have a conductive neoprene outer coating in the range of 0.0005 to 0.0024 inch.

3.2.3.3 Finished fabric. The finished fabric shall be black or slate color, and except as stated in 3.2.3.1.1 and 3.2.3.1.2, shall be in accordance with the requirements listed in Table I when tested as specified in 4.4.4.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:59Z
Check the source to verify that this is the current version before use.

TABLE I. Physical requirements of conductive finished fabric

Breaking strength, warp (pounds per inch min.)	90
Breaking strength, fill (pounds per inch min.)	90
Elmendorf tear strength, warp (gms. min.)	700
Elmendorf tear strength, fill (gms. min.)	525
Mullen bursting strength, (p.s.i. min.)	150
Stiffness (cantilever method for MIL-C-577 type cloth):	
Cuff proper, warp and fill (cms. max.)	9.5

3.2.4 Sewing thread. Thread for sewing shall be nylon thread in accordance with V-T-295, Size B. Thread shall be of the same color as material sewn. An alternate thread shall be cotton covered polyester thread in accordance with MIL-T-43548.

3.2.5 Nylon tape fasteners. Shall be 2-inch nylon tape fasteners in accordance with MIL-F-21840 for hook and pile. There shall be no visible electrostatic discharge at the hood and pile interface when tested as described in 4.4.9.

3.3 Style, design and dimensions. Shall consist of a cuff proper and the contained compression bag, ready for use. The components shall be as specified in the following paragraphs:

3.3.1 Cuff proper. The cuff shall be fabricated from material as specified in 3.2.3. The cuff shall be rectangular in shape and shall be provided with a pocket with an opening to allow insertion of the rubber arm bag. The cuff shall be of the dimensions as indicated in Table II. A flap shall be present which measures 4 x 2 inches for Size I cuffs and 4 x 2-1/2 inches for the Size II and Size III cuffs. The opening in the pocket of the cuff shall be reinforced by double stitching. Stitching may be present at the center of the opening. Design shall permit the compression bag tubes to protrude. A manometer loop consisting of a 1-1/8 ± 1/8 inch by 1-5/8 ± 1/8 inch nylon webbing shall be located on the outside upper center portion and 13/16 ± 3/16 inch from the top of the cuff.

Source: https://assist.dla.mil -- Downloaded: 20260-21T09:59Z
Check the source to verify that this is the current version before use.

TABLE II. Cuff dimensions (in inches)

Cuff	Length	Width	Pocket	Pocket opening	Area (minimum values in square inches) Hook	Pile
Size I	16-1/4 +5/8 -1/4 or 13-15/16 +1 -0	4-3/16 +1/4 -1/8	7-1/2 +1/4 -1/8	2-3/4 + 5/8, -1/4	11-1/2	20-1/2
II	20-3/4 +5/8 -0	5-1/2 +7/16 -1/8	9-3/4 +1 -0 or 10-13/16 ±5/16	2-3/4 + 5/8, -1/4	17-1/2	22-1/2
III	40 ±3/4	7-3/4 +1/4 -1/8	15 ±1/4	2-3/4 + 5/8, -1/4	71	56

3.3.1.1 Stitching and seams. All stitching shall be type 301 of Federal Standard No. 751 with 8 to 14 stitches per inch. Seam types shall be in accordance with the requirements of Federal Standard No. 751; type SSa-1 and type SSae-2 for hemming.

3.3.1.2 Closure. Closure of the cuff shall be accomplished through the use of nylon tape fasteners as specified in 3.2.5. A length of pile tape, not less than 8 inches, shall be sewn to one side and end (opposite to end containing compression bag) of the cuff, and two lengths of hook tape, not less than 3-3/4 inches, shall be sewn, side by side, on opposite side and end of the cuff directly over the compression bag pocket. As an alternate, a length of hook tape not less than 10 inches shall be sewn to one side and end (the same end containing the compression bag).

3.3.2 Compression bag. The compression bag shall be fabricated from rubber as specified in 3.2.2.1. The dimensions of the compression bags shall be as indicated in Table III. Two lengths of 5/32 inch inside diameter conductive tubing at least 8 inches in length shall be securely attached to a long edge of the arm bag to provide air entry to and exit from the bag. One tube shall be fitted with a female Luer connector.

Source: https://assist.dla.mil -- Downloaded: 202640-21T09:59Z
Check the source to verify that this is the currentversion before use.

TABLE III. Compression bag dimensions (in inches)

Size	Length	Width	Tube Lengths (minimum values)
I	6-5/8 to 7-3/8	3-1/4 to 3-13/16	8 + 8 plus extension tube of 25 inches
II	8-5/8 to 9-1/4 or 9-15/16 to 10-7/16	4-5/8 to 5-1/4	8 + 8 plus extension tube of 25 inches
III	13 to 14-1/2	6-3/4 to 7-1/2	8 + 8 plus extension tubes of 38 and 25 inches

3.3.3 Extension tubes. The extension tube shall be fabricated from 5/32 inch inside diameter conductive rubber tubing. Lengths shall be as shown in Table III. One end of the extension tubes shall be fitted with a male Luer connector in order to be fitted to the female Luer fitting present on the compression bag tubes.

3.3.4 Luer connectors. The metal female Luer connector of the compression bag shall be in accordance with American National Standards Institute (ANSI) Z70.1, and shall make a leakproof connection with the metal male Luer connector of the extension tube. When tested as specified in 4.4.7 the connector shall fit the feeler gage within the GO - NO GO limits. In addition, when the plug gage is fitted onto the connector as specified, there shall be no movement (wobble) of the connector on the gage when the connector is pressed lightly from side to side in all directions. One end of the extension tubing shall be fitted with a metal male Luer connector in accordance with ANSI Z70.1 When tested as specified in 4.4.8, the connector shall fit the ring gage within the GO - NO GO limits. In addition, when the ring gage is fitted onto the connector as specified, there shall be no movement (wobble) of the gage on the connector when the gage is pressed lightly from side to side in all directions.

3.4 Autoclaving. When tested as specified in 4.4.5, cuff (cuff proper, bag, tubing and metal parts) shall not lose shape or decompose.

3.5 Leakage.

3.5.1 Leakage of compression bag. The compression bag with extension tube attached shall be capable of withstanding an internal pressure of 12 pounds per square inch for one minute without leakage when tested as specified in 4.4.3.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:59Z
Check the source to verify that this is the current version before use.

3.5.2 Leakage of assembly. After autoclaving, the bag, tubing, connectors and extension tubing shall be assembled. The leakage of the assembly shall not exceed 10 mm. per minute when tested as specified in 4.4.6.

3.6 Identification marking. The cuff and compression bag shall be legibly and permanently marked with contractor's name or registered trademark. In addition, cuff and compression bag shall be marked with date of manufacture (month and year) and a suitable legend indicating that the item is conductive. Tubing shall be marked conductive.

3.7 Workmanship. Workmanship shall be first class throughout. Cuffs and compression bags shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the materials, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the National Stock Number, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:59Z
Check the source to verify that this is the current version before use.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified in Table IV. The unit of product for examination shall be one cuff with compression bag, extension tube and fittings.

TABLE IV. Sampling criteria for examination

Type of defect	Inspection level 1/	AQL (percent defective)
Major	II	1.0
Minor	II	1.5

1/ If acceptance number is not exceeded, lot shall be accepted. If rejection number is equaled or exceeded, inspection at level II may be performed to determine final acceptance or rejection.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as indicated in Table V.

TABLE V. Sampling criteria for tests

Component	Characteristic	Reqmt.	Test procedure	Unit of product	Insp. level	AQL
Compression bag	Tensile strength	3.2.2.2	4.4.1	1-bag	S-1	2/
Compression bag	Accelerated aging	3.2.2.2	4.4.2	1-bag	S-1	2/
Tubing	Tensile strength	3.2.2.2	4.4.1	1 or 2 extension tubes 3/	S-1	2/
Tubing	Accelerated aging	3.2.2.2	4.4.2	1 or 2 extension tubes 3/	S-1	2/
Cuff	All requirements of Table I	3.2.3.3	4.4.4	1/	1/	1/
Female connector	Diameter and depth of cavity	3.3.4	4.4.7	1-connector	II	1.0
Male connector	Taper	3.3.4	4.4.8	1-connector	II	1.0
Cuff complete	Autoclaving	3.4	4.4.5	1-cuff	S-2	2/
Compression bag, tubing, connectors	Leakage	3.5.1	4.4.3	1-bag	S-1	2/
Compression bag, tubing, connectors	Leakage of assembly, after autoclaving	3.5.2	4.4.6	1-bag	S-2	2/
Hook and pile fastener	Electrostatic charge	3.2.5	4.4.9	1-cuff	S-1	2/

1/ Sufficient samples shall be taken at random from each lot to perform all tests specified in 3.2.3.3. Failure to comply with any tests shall be cause for rejection of the lot represented.

2/ Acceptance number shall be zero for all sample sizes.

3/ One extension tube for sizes I & II. Two extension tubes for size III compression bag.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:59Z
Check the source to verify that this is the current version before use.

4.2.3 Where possible, the same sample unit shall be used for the determination of two or more test characteristics.

4.3 Examination.

4.3.1 Examination of cuffs. The cuffs shall be examined to determine compliance with all the requirements contained in this specification. Examination shall be conducted as specified in Table IV and Table VI. Examination shall not be restricted to defects listed in Table VI. Any cuff in the sample containing one or more defects will be rejected and if the number of defects in any sample exceeds the acceptance number for the sample the lot represented by the sample shall be rejected.

TABLE VI. Classification of defects

Categories	Defects
Critical	None defined.
Major	
101	Component parts damaged.
102	Rubber parts stiff or brittle.
103	Webbing missing.
104	Pile fastener loose or missing.
105	Luer connectors missing.
Minor	
201	Label missing.
202	Color not as specified.
203	Sewing of cuff pocket not as specified.
204	Identification marking not as specified.

4.3.2 Examination for preparation for delivery. An examination shall be made to determine whether the packaging, packing, and marking comply with the requirements of Section 5. Defects shall be scored as specified in Table VII. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 and the acceptable quality level shall be 4.0 defects per hundred units.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:59Z
Check the source to verify that this is the current version before use.

TABLE VII. Examination for preparation for delivery

Examination	Defect
Containers	Not as specified.
Quantity	Quantity in interior packages and shipping containers not as specified.
Packaging and packing materials	Any component missing or damaged.
Workmanship	Inadequate application of components, such as incomplete closure of container flaps, tape improperly applied, or containers distorted.
Markings	Omitted, incorrect, illegible, improper location, sequence, or method of application.

4.3.2.1 Examination of closure and reinforcement for containers. When shipping containers are required to comply with PPP-B-636, examination for defects in closure and reinforcement shall be in accordance with the appendix to that specification.

4.4 Tests. Tests shall be conducted, as necessary, to determine compliance with specification requirements.

4.4.1 Tensile strength test. Tensile strength shall be determined by the applicable method specified in Federal Test Method Standard No. 601, Method 4111.

4.4.2 Accelerated aging test. Rubber tubing and seamless arm bag shall be subject to the accelerated aging test specified in Test Method 7111 of Federal Test Standard No. 601.

4.4.3 Leakage (compression bag) test. The compression bag with extension tube attached shall be subjected to the required pressure (12 p.s.i.g.). The bag shall be secured from ballooning by use of screening of sufficient strength to withstand the pressure.

4.4.4 Finished fabric test. The finished fabric used in the cuffs shall be examined for compliance with specification requirements and tested in accordance with the test methods specified in Fed. Test Method Std. 191 for the following characteristics:

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:59Z
Check the source to verify that this is the currentversion before use.

Characteristic	Test Method
Breaking strength (warp and fill)	5100
Elmendorf tear strength (warp and fill)	5132
Mullen bursting strength	5122
Stiffness (MIL-C-577 type cloth)	5204 Cantilever Method

4.4.5 Autoclaving. Cuffs, compression bags, tubing and connectors shall be steam autoclaved at 15 p.s.i.g and 250° F (for 15 minutes and then allowed to cool at room temperature (68° - 72° F.) for one hour. This test shall be performed three times.

4.4.6 Leakage of assembly test. After autoclaving in accordance with 4.4.5, the compression bag and extension tube with Luer connectors shall be assembled and subjected to a pressure of 300 millimeters of mercury for one minute. The bag shall be secured from ballooning by the use of screening of sufficient strength to withstand the pressure.

4.4.7 Diameter and depth of cavity of female connector. The cavity shall be tested on the male Luer taper gage in accordance with Figures 5 and 6 of ANSI Z70.1. This gage shall be as shown in ANSI Z70.1. The connector shall be slipped onto this gage with very light pressure. The fit of the connector on this gage shall be tested by means of the feeler gage in accordance with Figures 5 and 6 of ANSI Z70.1.

4.4.8 Taper of male connector. The tapered end of the connector shall be tested for proper dimensions and taper by means of the ring gage shown in Figure 7 of ANSI Z70.1. This gage shall be slipped onto the tip with very light pressure.

4.4.9 Electrostatic discharge of hood and pile fastener. The completed cuff shall be conditioned at 10 to 20% relative humidity and 70 to 75 F for 16 hours. The hook and pile fastener shall then be tested in a dark room of the same relative humidity by attaching and rapidly detaching the fastener material. There shall be no electrostatic discharge at the hook-pile interface. The test shall be repeated five times with the maximum area of engagement of the hook and pile fastener material.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A, or C, as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Unit package. Each cuff of the size specified shall be packaged in a 3 mil polyethylene bag with the cuff and tubing arranged in such a manner that it will not kink or be permanetnly deformed. Excess air shall be evacuated from the polyethylene bag and closure effected by heat sealing. The item shall be further packaged in a snug fitting box constructed in accordance with PPP-B-566 or PPP-B-676 except that commercial colors are acceptable. Closure of the box shall be secured to prevent opening under normal handling.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:59Z
Check the source to verify that this is the current version before use.

5.1.1.2 Intermediate package. Twelve unit packages containing cuffs of one size only shall be packaged in a snug fitting box constructed in accordance with PPP-B-636, type CF, class domestic. Closure shall conform to method IV as specified in PPP-B-636.

5.1.2 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A. Four intermediate packages containing Size III cuffs or six intermediate packages containing Size I or Size II cuffs shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-636 class weather-resistant. Closure, strapping and waterproofing shall be as specified in the appendix of the box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2.2 Level B. Four intermediate packages containing Size III cuffs or six intermediate packages containing Size I or Size II cuffs shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-636, type CF, class domestic. Closure of the box shall conform to method I or II as specified in PPP-B-636.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size and kind commonly used for the purpose, constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.2.4 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. Contract or purchase order number and name of contractor shall be shown.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:59Z
Check the source to verify that this is the current version before use.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. The sphygmomanometer cuff and compression (arm) bag shall be used in the determination of blood pressure. The cuff and compression bags are conductive for possible operating room usage.

6.2 Ordering data. Procurement documents should specify the following:

a. Number, title and date of this specification.
b. Applicable levels of packaging and packing.
c. National stock number of item ordered (NSN).
d. Size.

6.3 This specification is intended to cover the following items appearing in the Federal Supply Catalog:

NSN	Item Identification
6515-00-116-0533	CUFF, SPHYGMOMANOMETER, Adult, Conductive, Giant Size
6515-00-817-2530	CUFF, SPHYGMOMANOMETER, Child, Conductive,
6515-00-904-5630	CUFF, SPHYGMOMANOMETER, Adult, Conductive,

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1888
CS

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:59Z
Check the source to verify that this is the currentversion before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-C-37107 CUFF, SPHYGMOMANOMETER, CONDUCTIVE

NAME OF ORGANIZATION AND ADDRESS	CONTRACT NUMBER
	MATERIAL PROCURED UNDER A ☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*	TELEPHONE NO.
	DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

☆ GPO 1974—603-660/6126 2-1

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:59Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

U.S.MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:59Z
Check the source to verify that this is the current version before use.