MILITARY SPECIFICATION

CLIP, HEMOSTATIC, RAINEY, 5/8 INCH

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a hemostatic clip, of the Rainey design, suitable for hemostasis during cranial surgical procedures.

2. APPLICABLE DOCUMENTS

2.1 Specifications, standards and drawings. The following documents of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein.

SPECIFICATIONS

Federal

PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Box, Wood, Wirebound.
PPP-B-601	Box, Wood, Cleated-plywood.
PPP-B-621	Box, Wood, Nailed and Lock-corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Box, Set-up.

Military

MIL-P-116	Preservation, Methods of.
MIL-L-10547	Liners, Case and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:01Z
Check the source to verify that this is the current version before use.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center (DPSC)

Drawing No. 1995 Clip, Hemostatic, Rainey, 5/8 inch.

(Copies of specifications, standards and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring acitvity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material. The clip shall be fabricated from tempered, spring grade, SAE 1095 carbon steel. The finished clip shall be nickel plated in accordance with QQ-N-290, Class 1, Type III (KS).

3.2 Style, design and dimensions. Shall be a hemostatic clip of the Rainey design, suitable for hemostasis of the scalp edge during cranial surgical procedures. The clip shall be suitable for use with National Stock Number 6515-00-332-6200 Forceps, Hemostatic, Clip, Applying, Rainey, and shall be in accordance with DPSC Drawing No. 1995.

3.3 Sterilization. The clip shall be capable of steam sterilization. When tested as specified in 4.5.1, there shall be no evidence of corrosion, distortion, cracking, discoloration, peeling or blistering.

3.4 Finish. The clip shall possess a smooth, uniform finish throughout. The clip shall be free from burrs, sharp edges (except for clamping teeth). cracks, pits, scale, corrosion and non-uniform plating.

3.5 Permanent set. The clip shall exhibit a permanent set of 5/32 inch maximum when tested as specified in 4.5.2.

3.6 Workmanship. Workmanship shall be first class throughout. Clips shall be free from defects which detract from their appearance or impair their serviceability.

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:01Z
Check the source to verify that this is the currentversion before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificate of quality, supplied by the manufacturer of the tempered, spring grade steel, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, thename of the manufacturer or supplier, the National Stock Number (NSN), the item identification, the name of the material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and as indicated in Table I. Unit of Product for purposes of examination shall be one clip.

Table I. Sampling Criteria for Examination

Type defect	Inspection level	AQL (percent defective)
Major	II	1.0
Minor	None defined	

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:01Z
Check the source to verify that this is the current version before use.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and as indicated in Table II. Unit of product for test purposes shall be one clip. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics.

Table II. Sampling Criteria for Tests

Component	Characteristic	Requirement	Test procedure	Inspection level	AQL*
Clip	Dimensions	3.2	---	S-1	2.5
	Sterilization test	3.3	4.5.1	S-1	1.5
	Permanent set	3.5	4.5.2	S-1	1.5

*Percent defective

4.3 Examination.

4.3.1 Clip. The clips shall be examined to determine compliance with all requirements contained in this specification. Examination shall be conducted as specified in Tables I and III. Examination, however, shall not be restricted to defects listed in Table III. Any clips in the sample containing one or more defects shall be rejected and if the number of defective clips in any sample exceeds the acceptance number for the sample, the lot represented by that sample shall be rejected.

Table III. Classification of Defects

Categories	Defects
Critical	None defined.
Major	
101	Clip not plated as specified.
102	Plating not uniform and not free from flaking.
103	Clip not free from cracks, burrs, pits, scale and corrosion.

4.3.2 Preparation for delivery. An examination shall be made to determine whether the packaging, packing and marking comply with the requirements of Section 5. Defects shall be scored as specified, but not limited to those defects, in Table IV. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 with an AQL of 4.0 expressed in terms of defects per hundred units.

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:01Z
Check the source to verify that this is the currentversion before use.

Table IV. Examination of Preparation for Delivery

Examination	Defect
Containers	Not as specified.
Contents	Not as specified.
Markings	Omitted; incorrect; illegible; improper size, location, sequence, or method of application.
Packaging & Packing Materials	Component missing or damaged.
Workmanship	Bulging or distortion of containers. Cushioning inadequate, improper or missing.

4.4.1 Examination of closure and reinforcement for containers. When shipping containers are required to comply with PPP-B-636, examination for defects in closure and reinforcement shall be in accordance with the appendix to that specification.

4.5 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.5.1 Sterilization test. The clips shall be placed into a suitable sterilizer at a temperature of 250° Fahrenheit for a period of at least 10 minutes. The clips shall then be removed from the sterilizer and allowed to cool to room temperature. This test shall be performed a total of 3 times on the sample unit.

4.5.2 Permanent set. Measure the existing opening of the sample clip. The clip shall then be held in the hand, and with a suitable instrument placed at the center of the clip, spread open to a distance of 3/8 ± 1/32 inch and then released. The clip shall return to within 5/32 inch maximum of the starting dimension. The test shall be repeated 3 times.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One clip, as specified, constitutes one unit of issue.

5.1.2 Level A.

5.1.2.1 Unit package. Each clip shall be thoroughly clean, wrapped in neutral tissue and packaged in accordance with MIL-P-116, method IC-1 except that a preservative shall not be utilized.

5.1.2.2 Intermediate package. Twelve unit packages shall be packaged in a snug fitting box constructed in accordance with PPP-B-566 or PPP-B-676. Commercial colors are acceptable on intermediate boxes and the closure shall be secured to prevent spilling of contents under normal handling.

Source: https://assist.dla.mil -- Downloaded: 202640-21T10:01Z
Check the source to verify that this is the current version before use.

5.1.3 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A. One hundred and twenty intermediate packages shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-636, class weather-resistant. The box shall also conform to the special requirements as specified in PPP-B-636. Closure, strapping and water-proofing shall be as specified in the appendix of the box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level B. One hundred and twenty intermediate packages shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-636, class domestic. The box shall conform to the special requirements as specified in PPP-B-636. Closure of the box shall conform to method I or II as specified in the appendix of the box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.2.4 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:01Z
Check the source to verify that this is the current version before use.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. Clip, Hemostatic, Rainey is used for hemostasis during cranial surgical procedures.

6.2 Ordering data. Procurement documents should specify the following:

(a) Title, number and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 This specification is intended to cover the following items listed in the Federal Supply Catalog:

NSN	Item Identification
6515-00-375-4600	CLIP, HEMOSTATIC, RAINEY, 5/8 INCH

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1902
cs

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:01Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:01Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-C-37124 CLIP, HEMOSTATIC, RAINEY, 5/8; INCH

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 202640-21T10:01Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, Pa. 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304



OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:01Z
Check the source to verify that this is the current version before use.