MILITARY SPECIFICATION

CURETTE, EYE, KNAPP, 5 INCH

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a curette, of the Knapp design, suitable in surgical procedures of the eye.

2. APPLICABLE DOCUMENTS

2.1 Specifications, standards and drawing. The following documents of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein.

SPECIFICATIONS

Federal

| | |
|---|---|
| GG-I-526 | Instruments, Dental and Surgical; General Specifications. |
| PPP-B-585 | Boxes, Wood, Wirebound. |
| PPP-B-601 | Boxes, Wood, Cleated-Plywood. |
| PPP-B-621 | Boxes, Wood, Nailed and Lock-Corner. |
| PPP-B-636 | Box, Fiberboard. |
| PPP-B-676 | Boxes, Set-up. |

Military

| | |
|---|---|
| MIL-P-116 | Preservation, Methods of |
| MIL-L-10547 | Liners, Case and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible. |

STANDARDS

Military

| | |
|---|---|
| MIL-STD-105 | Sampling Procedures and Tables for Inspection by Attributes. |
| MIL-STD-129 | Marking for Shipment and Storage. |



Source: https://assist.dla.mil -- Downloaded: 202610-21T10:04Z
Check the source to verify that this is the current version before use.

DRAWING

Defense Personnel Support Center (DPSC)

No. 20319 Curette, Eye, Knapp, 5 inch

(Copies of specifications, standards, and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 General requirements. The eye curette shall be in accordance with the applicable requirements of GG-I-526, and as specified herein.

3.2 Material. The curette shall be fabricated from class 4, type 420 material.

3.3 Hardness. After proper stress relieving, Rockwell hardness shall be C40 to C45.

3.4 Style, design and dimensions. Shall be a curette, Knapp design, suitable for use involving surgical procedures of the eye. The distal end shall have a shallow scoop, egg-shaped configuration, possessing a well rounded, polished, non-cutting edge. The curette shall be in strict accordance with DPSC Drawing No. 20319.

3.5 Finish. Finish shall be mirror, or mirror and satin.

3.5.1 Passivation. The curette shall be properly passivated.

3.6 Identification marking. In addition to the requirements specified in GG-I-526, curettes of foreign manufacture shall be marked "stainless".

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:04Z
Check the source to verify that this is the current version before use.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the corrosion-resistant steel, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and as indicated in Table I. The unit of product for examination purposes shall be one curette.

Table I. Sampling Criteria for Examination

| Type defect | Inspection level | AQL (percent defective) |
|---|---|---|
| Major | II | 1.0 |
| Minor | II | 2.5 |

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and as indicated in Table II. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics. The unit of product for test purposes shall be one curette.



Source: https://assist.dla.mil -- Downloaded: 202640-21T10:04Z
Check the source to verify that this is the currentversion before use.

Table II. Sampling Criteria for Tests

| Component | Characteristic | Requirement | Test procedure | Inspection level | AQL |
|---|---|---|---|---|---|
| Forceps | Hardness test | 3.3 & GG-I-526 | GG-I-526 | S-2 | 1.5 |
| | Dimensions | 3.4 | -- | S-2 | 2.5 |
| | Copper sulfate test | GG-I-526 | GG-I-526 | S-2 | 1.0 |
| | Boil test | GG-I-526 | GG-I-526 | S-2 | 1.0 |

4.2.3 Raw Materials. A sufficient quantity of each lot of metal used in the construction of the forceps shall be selected for chemical analysis.

4.3 Examination.

4.3.1 Examination of curettes. The curette shall be examined to determine compliance with all the requirements contained in this specification. Examination shall be conducted as specified in Tables I and III. Examination, however, shall not be restricted to defects listed in Table III. Any curettes in the sample containing one or more defects shall be rejected and if the number of defective curettes in any sample exceeds the acceptance number for the sample, the lot represented by that sample shall be rejected.

Table III. Classification of Defects

| Categories | Defects |
|---|---|
| Critical | None defined. |
| Major | |
| 101 | Curette not of the design specified. |
| 102 | Working edge sharp. |
| 103 | Working edge not well rounded. |
| 104 | Curette not free from cracks, pits, burrs, sharp edges, scale, rough areas, and corrosion. |
| Minor | |
| 201 | Curette not identified as specified in 3.6. |



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:04Z
Check the source to verify that this is the current version before use.

4.3.2 Examination of Preparation for Delivery. An examination shall be made to determine whether the packaging, packing, and marking comply with the requirements of Section 5. Defects shall be scored as specified in Table IV. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 and the acceptable quality level shall be 4.0 defects per hundred units.

Table IV. Examination of Preparation for Delivery

| Examination | Defect |
|---|---|
| Containers | Not as specified. |
| Quantity | Quantity in interior packages and shipping containers not as specified. |
| Materials | Any component missing or damage. |
| Workmanship | Inadequate application of components such as incomplete closure of container flaps, tape improperly applied, or containers distorted. |
| Markings | Omitted, incorrect, illegible, improper size, location, sequence, or method of application. |

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

5. PREPARATION FOR DELIVERY

5.1 Preservation and Packaging. Preservation and packaging shall be level A or C as specified (see 6.1).

5.1.1 Level A.

5.1.1.1 Cleaning and drying. Each curette shall be thoroughly cleaned and dried in accordance with the applicable method as outlined in MIL-P-116.

5.1.1.2 Preservative. Preservative shall not be applied.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:04Z
Check the source to verify that this is the current version before use.

5.1.1.3 Unit package. Each curette shall be wrapped in neutral tissue or other suitable anticorrosive materials. Curette shall be further packaged in accordance with MIL-P-116, method 1C-1.

5.1.1.4 Intermediate package. Six (6) curettes shall be packaged in a box of appropriate size and design constructed in accordance with PPP-B-566, PPP-B-636, Type CF, class domestic or PPP-B-676. Closure shall be adequate to prevent accidental opening.

5.1.2 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.1).

5.2.1 Level A. Three hundred sixty (360) curettes shall be packed in a box designed for a type 2 load, and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2 or PPP-B-636, type CF, class weather-resistant. Closure and strapping shall be in accordance with the appendix of the applicable box specification. Special requirements of applicable table shall apply for PPP-B-636 box.

5.2.1.2 Case liner. Each Level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Fiberboard boxes shall be waterproofed in accordance with PPP-B-636.

5.2.2 Level B. Three hundred sixty (360) curettes shall be packed in a box designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:04Z
Check the source to verify that this is the current version before use.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Title, number, and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 This specification is intended to cover National Stock Number 6515-00-366-1300 CURETTE, EYE, Knapp, 5 inch, appearing in the Federal Supply Catalog.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1911

is



Source: https://assist.dla.mil -- Downloaded: 202610-21T10:04Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:04Z
Check the source to verify that this is the current version before use.

# STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-C-37150 CURETTE, EYE, KNAPP, 5 INCH

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426

REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:04Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, Pa. 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

U.S.MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:04Z
Check the source to verify that this is the current version before use.