MILITARY SPECIFICATION

HOLDER, DENTAL MIXING PAD

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 This specification covers a holder suitable for holding a paper pad during dental operative mixing procedures.

2. APPLICABLE DOCUMENTS

2.1 The following documents, of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

QQ-W-470	Wire, Steel, Carbon, Spring, Music.
PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Boxes, Set-up.

Military

MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:58Z
Check the source to verify that this is the current version before use.

DRAWING

Defense Personnel Support Center (DPSC)

24222 Holder, Dental Mixing Pad.

(Copies of specifications, standards and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material. All materials shall be suitable for the intended purpose.

3.1.1 Frame. The frame shall be fabricated from aluminum alloy or corrosion-resisting steel sheet.

3.1.2 Clip, clip retention pin, and rivets. The clip, clip retention pin, and rivets shall be fabricated from carbon steel. As an alternate, the rivets shall be fabricated from aluminum alloy.

3.1.3 Clip spring. The clip spring(s) shall be fabricated from music-spring, carbon steel wire 0.055 or 0.067 inch in diameter, as specified in QQ-W-470, except that the tensile strength shall be a minimum of 250,000 pounds per square inch (p.s.i.).

3.1.4 Backing. The backing shall be fabricated from open-celled, resilient, synthetic rubber sheet. The exposed face backing shall have a screen grained finish.

3.1.5 Adhesive. The adhesive shall be pressure-sensitive, and shall be suitable for firmly attaching one side of the backing to the underside of the frame.

3.2 Style, design, and dimensions. The holder shall consist essentially of an aluminum frame, with clipboard features, and a rubber mat cemented to the bottom of the frame. The holder shall be suitable for the reception, retention, and stabilization of a six-by-eight-inch, dental, parchment-paper mixing pad (NSN 6520-00-542-2550) during the spatulation and mixing of viscous dental impression materials. The style, design, and dimensions of the holder shall be in accordance with DPSC Drawing 24222.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:58Z
Check the source to verify that this is the current version before use.

3.2.1 Drawings. The drawings forming a part of this specification are specification control drawings. The supplier shall be responsible for preparing his own shop drawings. Where tolerances prescribed could cumulatively result in improper design, the supplier shall provide a product with dimensions within the tolerances prescribed on the drawing to insure proper design, fit, function, assembly and operation of the item. No deviation from the prescribed dimensions or tolerances is permissible without prior approval of the contracting officer.

3.3 Construction. The spring clip shall be firmly riveted to the top side of the frame, and the nonskid backing shall be firmly attached to the underside of the frame. The spring action of the clip shall be provided by either two spring wires or, as an alternate, one spring wire. The clip shall permit ready adjustability for insertion and removal of the pad specified in 3.2. Any bend (which is optional) near the top of the lower clip member shall not permit any portion of the clip to touch or mar a flat, level working surface, when the pad is in use. The top side of the backing shall contain sufficient pressure-sensitive adhesive to enable the backing to adhere firmly to the underside of the metal frame. A bend at the lower end of the frame shall provide a suitable vertical support for one of the six-inch edges of the mixing pad. The corrugated rubber sheet backing shall prevent movement of the holder during use.

3.4 Plating. The entire top side of both upper and lower members of the clip shall be nickel plated or chormium-over-nickel plated. The rivets, if manufactured from carbon steel, shall be rendered suitably corrosion resistant.

3.5 Finish. The frame of the holder shall have a smooth satin or polished finish. The top of the upper and lower clip members shall have a bright, polished finish. The clip retention pin shall have a smooth, polished finish, except that the head of the pin may be knurled. The spring shall have a satin or bright finish.

3.6 Workmanship. Workmanship shall be first class throughout. The holders shall be free from defects which detract from their appearance of may impair their serviceability.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:58Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in this specification, where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the frame, spring clip, and backing, may be furnished in lieu of actual performance of such testing by the contractor, provided that lot identity has been maintained and can be demonstrated to the Government. The Certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the name of the component, the lot number, the lot size, the sample size, the NSN, the Item Identification, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Quality conformance inspection. The examinations and tests required to assure conformance of those items or lots of items to be offered for acceptance are classified as specified in Table I (see 4.2.1) and Table II (see 4.3).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:58Z
Check the source to verify that this is the current version before use.

4.2.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels (AQL's) shall be as indicated in Table I. The unit of product for examination purposes shall be one holder.

Table I. Examination criteria

Type defect	Inspection level	AQL (percent defective)
Major A	II	1.0
Major B	S-2	1.5
Minor A	II	2.5
Minor B	S-2	4.0

4.2.2 Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, inspection level S-2, except that the acceptance number shall be zero for all sample sizes. The unit of product for test purposes shall be one holder.

4.3 Examination. The holders shall be examined for defects including, but not limited to, those listed in Table II.

Table II. Classification of defects

Category	Defect
Major A:	
101	Style and design not as specified.
102	Rubber backing not nonskid, not resilient.
103	Adhesive not pressure-sensitive, does not firmly attach backing to frame.
104	Clip not in proper operating order.
105	Any component missing.
106	Any portion of bottom of lower clip member touches working surface while in use.
Major B:	
151	Distance from inner edge of lower member to beginning of 90° bend at bottom of frame not as specified.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:58Z
Check the source to verify that this is the current version before use.

Table II. Classification of defects (cont'd)

Minor A:	
201	Rubber backing does not completely cover underside of frame.
202	Type of finish not as specified.
203	Finish not free from coarse grind marks or scratched surface areas.
204	Packaging and packing not as specified.
205	Marking on packaging and packing not as specified.
Minor B:	
251	Any dimension (other than 151) not as specified on drawing.

4.4 Tests. Tests shall be performed to determine compliance with specification requirements.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be Level A or C as specified (see 6.1).

5.1.1 Unit of issue. One holder, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in Table III shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table III. Packaging quantities

Unit package	Intermediate package	Exterior container
1 unit	Not required	24 units

5.1.2.1 Variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:58Z
Check the source to verify that this is the current version before use.

5.1.3 Unit package. Each unit shall be packaged in a box of appropriate size and design constructed in accordance with PPP-B-566 or PPP-B-676, except that commercial colors are acceptable. Closure shall be adequate to prevent spilling of contents under normal handling. Box shall be designed with suitable interior fittings or shall contain sufficient cushioning material to insure against shifting of, or damage to, the contents or package.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be Level A, B or C as specified (see 6.1).

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant, grade V3c. Closure and strapping shall be as specified in the appendix of the applicable box specification. Special requirements of applicable table shall apply for PPP-B-636 box.

5.2.1.2 Case liner. Each Level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Fiberboard box shall be waterproof in accordance with paragraph 30.4 of PPP-B-636.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification. Fiberboard box shall be closed by method II of PPP-B-636. Special requirements of applicable table shall apply for PPP-B-636 box.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:58Z
Check the source to verify that this is the current version before use.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Title, number, and date of this specification.

(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 National Stock Number (NSN) coverage. This specification covers the following item appearing in the Federal Supply Catalog:

NSN	Item Identification
6520-00-180-5974	HOLDER, DENTAL MIXING PAD, Metal, Dental Parchment Paper Pad.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6520-1382
is

Source: https://assist.dla.mil -- Downloaded: 202640-21T09:58Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-H-37098

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426

REPLACES EDITION OF 1 JAN 68 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:58Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304



OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:58Z
Check the source to verify that this is the current version before use.