MILITARY SPECIFICATION

KNIFE, EYE, NEEDLE TYPE, ZIEGLER, 6 mm

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers the requirements for an eye knife, needle type of the Ziegler design, suitable for surgical procedures of the eye.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposals form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

GG-I-526	Instruments, Dental and Surgical, General Specifications.
PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Boxes, Set-Up.

Military

MIL-P-116	Preservation, Methods of.
MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:00Z
Check the source to verify that this is the current version before use.

STANDARDS

Military

MIL-STD-105 Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center (DPSC)

1859 Knives, Eye.

(Copies of specifications, standards, and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 General requirements. The eye knife shall be in accordance with the applicable requirements of GG-I-526 and the following option exercises under that specification:

3.2 Materials and finishes. The eye knife shall be fabricated and finished as specified in Table I and in accordance with GG-I-526. When more than one class and type is indicated, either one shall be acceptable.

Table I. Material and Finish Requirements

Item	Material		Finish
	Class	Type	
Blade	1	1095	Crocus or satin
Handle	1	1095	chrome over nickel
	4	410	mirror or mirror and satin
		420	
	8	260	chrome over nickel
		270	
		360	

3.3 Hardness. The blade and blade shank shall have a Rockwell hardness of C-50 to C-60 after proper stress relieving.

3.4 Design. The eye knife shall be of the Ziegler design, consisting of a one-piece blade and blade shank, with a handle firmly and permanently attached to the blade shank. The dimensions shall be in accordance with DPSC Drawing No. 1859, Knives, Eye.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:00Z
Check the source to verify that this is the current version before use.

3.5 Workmanship. Workmanship shall be first class throughout. The eye knife shall be free from burrs, sharp edges, cracks, holes, pits, scale, corrosion, flaking or incomplete plating, foreign particles, and any other defects which detract from its appearance or impair its serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection shall be as defined in GG-I-526.

4.1.3 Certificates of Quality. Certificates of quality, supplied by the manufacturer of the material, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificates shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the National Stock Number (NSN), the item identification, the name of the material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection level and acceptable quality levels shall be as indicated in Table II. The unit of product for examination purposes shall be one eye knife.

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:00Z
Check the source to verify that this is the currentversion before use.

Table II. Sampling for examination

Type of defect	Inspection level	AQL (percent defective)
Major	II	1.0
Minor	II	4.0

4.2.2 Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as indicated in Table III. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Table III. Sampling for Tests

Unit of product	Characteristic	Requirement	Test procedure	Inspection level	AQL (percent defective)
Knife	Composition analysis	3.2 and GG-I-526	GG-I-526	**	*
	Hardness	3.3 and GG-I-526	GG-I-526	S-2	1.5
	Dimensions	3.4	-	S-1	3.5
	Salt Spray test (for Class B)	GG-I-526	GG-I-526	S-2	1.5
	Boil test (handle only)	GG-I-526	GG-I-526	S-2	1.5

* Acceptance number shall be zero for all sample sizes.
** Sufficient amount from each lot of each metal used in the construction of the eye knife shall be selected for composition analysis.

4.3 Examination.

4.3.1 Eye knife. The eye knife shall be examined to determine compliance with specification requirements. Examination shall be conducted in accordance with MIL-STD-105 and the following classification of defects. Examination shall include, but shall not be limited to, the defects listed in Table IV.

Table IV. Classification of Defects

Categories	Defects
Major	
101	Style and design not as specified.
102	Instrument not constructed as specified.
103	Instrument not free from sharp edges, rough areas, burrs, cracks, grind marks or overlaps.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:00Z
Check the source to verify that this is the current version before use.

Table IV. Classification of Defects (Cont'd)

Categories	Defects
Minor	
201	Finish not as specified.
202	Instrument not permanently and legibly marked with either the name or trademark of the contractor.

4.3.2 Preparation for delivery. An examination shall be made to determine whether the packaging, packing, and marking comply with the requirements of Section 5. Defects shall be scored as specified in Table V. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 and the acceptable quality level shall be 4.0 (defects per hundred units).

Table V. Examination of preparation for delivery

Examination	Defect
Containers	Not as specified.
Quantity	Quantity in interior packages and shipping containers not as specified.
Packaging and Packing Materials	Any component missing or damaged.
Workmanship	Inadequate application of components such as incomplete closure of container flaps, tape improperly applied, or containers distorted.
Markings	Omitted, incorrect, illegible, improper size, location, sequence, or method of application.

4.3.2.1 Examination of closure and reinforcement for containers. When shipping containers are required to comply with PPP-B-636, examination for defects in closure and reinforcement shall be in accordance with the appendix to that specification.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:00Z
Check the source to verify that this is the current version before use.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C as specified (see 6.2).

5.1.1 Unit of issue. Each knife constitutes one unit of issue.

5.1.2 Level A.

5.1.2.1 Unit package. Each knife shall be thoroughly cleaned in accordance with the applicable cleaning method in MIL-P-116. The blade of the knife shall be protected by inserting into a block of styrafoam or by other suitable means. The knife shall be wrapped in neutral tissue or other suitable barrier material and packaged in accordance with MIL-P-116, method IC-2 except that a preservative shall not be utilized.

5.1.2.2 Intermediate package. Six unit packages shall be packaged in a snug fitting box constructed in accordance with PPP-B-566 or PPP-B-676. Closure of the box shall be adequately secured to prevent spilling of contents under normal handling.

5.1.3 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A. Forty eight intermediate packages (288 units) shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, overseas type; PPP-B-621, class 2 or PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed as specified in the appendix of PPP-B-636.

5.2.2 Level B. Forty eight intermediate packages (288 units) shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification except the closure of fiberboard boxes shall conform to method I or II as specified in PPP-b-636.

Source: https://assist.dla.mil -- Downloaded: 202640-21T10:00Z
Check the source to verify that this is the currentversion before use.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. The needle type eye knife of the Ziegler design is intended for use in surgical procedures of the eye.

6.2 Ordering data. Procurement documents should specify the following:

(a) Title, number and date of this specification.
(b) Selection of applicable levels of packaging and packing. (see 5.1 and 5.2).

6.3 This specification is intended to cover the following items listed in the Federal Supply Catalog:

NSN	Item Identification
6515-00-344-2900	Knife, Eye, Needle Type, Zeigler, 6 mm.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1893
cs

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:00Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:00Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE: MIL-K-37115

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:00Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, Pa. 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:00Z
Check the source to verify that this is the current version before use.