MILITARY SPECIFICATION

MASK, OXYGEN, PLASTIC, DISPOSABLE

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 This specification covers a disposable under-the-chin type mask of flexible plastic complete with tube and inlet connector for the purpose of administering oxygen therapy.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-636	Box, Shipping, Fiberboard

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

(Copies of specifications and standards, required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:00Z
Check the source to verify that this is the current version before use.

2.2 Other publication. The following document forms a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

D-412 Tension Testing of Vulcanized Rubber

(Application for copies should be addressed to The American Society for Testing and Materials, 116 Race Street, Philadelphia, Pennsylvania 19103.)

3. REQUIREMENTS

3.1 Material. The oxygen mask shall be made of pliable transparent or translucent plastic. There shall be a rotatable tubing connector made of rigid polyethylene plastic attached to the center of the mask and to which there is permanently attached a green-tinted, pliable, transparent length of medical grade polyvinyl chloride tubing. The tubing at its distal end shall have a green molded polyvinyl chloride, or other suitable material, female connector permanently attached to it. The elastic head band shall be made of rubber and nylon, rayon, or cotton.

3.2 Style, design, and dimensions. The mask shall be designed to be usable on a wide variety of facial shapes and sizes. A plastic tubing connector shall be present on the mask at a location corresponding to the zone beneath the nose and near the mouth of the wearer. The tubing connector shall be rotatable 360° in either direction, and may have a baffle diffuser for directing oxygen flow across the interior surface of the mask. An exhaust opening or openings shall be present in the mask. The tubing shall be 60 to 102 inches long, resilient and non-tacky. The distal end shall have a tapered, female connector for attachment to the standard nipple of a regulator from an oxygen supply. An elastic head band shall be present and attached to the face mask. The oxygen mask shall be the Monaghan No. 1105-S Oxygen Mask, Med-Econ Plastics Co. No. 1421-8, or Oxequip No. 122-AB, or equal.

3.2.1 Tensile strength and elongation. There shall be a minimum tensile strength of 1800 pounds per square inch (psi) and ultimate elongation of 200 percent when the tubing is tested as specified in 4.4.1.

3.3 Safety (skin contact) test. The plastic mask material shall not cause skin irritation when tested as specified in 4.4.2 nor shall there be a delayed skin reaction 24 hours after removal of the patch.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:00Z
Check the source to verify that this is the current version before use.

3.4 Certificate of safety (skin contact) test. A certificate listing all lot or control numbers shall state that representative samples from each lot or control have been tested as specified in 3.3 and found to be non-dermatitic. This certificate shall be signed by a doctor of medicine and shall include the specific tests performed and the test results. For destination inspection, certificates shall be submitted by the contractor directly to Defense Personnel Support Center, Directorate of Medical Materiel, Technical Services Branch, Technical & Quality Assurance Division, 2800 South 20th Street, Philadelphia, PA 19101, ATTN: Quality Assurance Branch, and one copy of each shall be forwarded with each shipment to the consignee. For source inspection, certificates shall be furnished to the procuring activity via the government representative.

3.5 Security of mask-tube assembly. There shall be no evidence of separation at the junction of the tubing and the mask when subjected to stress as required in 4.4.3.

3.6 Workmanship. Workmanship shall be first class throughout. The oxygen mask, the tubing and other components shall be free from defects which detract from their appearance or impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:00Z
Check the source to verify that this is the current version before use.

4.1.3 Certificates of quality. Certificates of quality supplied by the manufacturer of the plastics may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and as specified in Table I. The unit of product for examination purposes shall be one oxygen mask assembly.

Table I. Sampling for Examination

Type of defect	Inspection level	AQL (percent defective)
Major	II	1.0
Minor	II	4.0

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as indicated in Table II. Where feasible, the same sample shall be used for the determination of two or more test characterisitcs.

Table II. Sampling for Tests

Component	Characteristic	Req't	Test Procedure	Insp. Level	AQL (percent defective)
Tubing	Tensile Strength	3.2.1	4.4.1	S-2	*
Mask	Safety	3.3	4.4.2	**	*
Assembly	Security	3.5	4.4.3	S-2	*

* Acceptance number shall be zero for all sample sizes.
** A minimum of five masks shall be tested from each lot.

Source: https://assist.dla.mil -- Downloaded: 202640-21T10:00Z
Check the source to verify that this is the currentversion before use.

4.3 Examination. The oxygen mask assembly shall be examined to determine compliance with all the requirements contained in this specification. Examination shall be conducted as specified in Tables I and III. Examination, however, shall not be restricted to defects listed in Table III. Any oxygen mask assembly in the sample containing one or more defects shall be rejected, and if the number of defective items in any sample exceeds the acceptance number for test sample, the lot represented by the sample shall be rejected.

Table III. Classification of defects

Category	Defect
Critical	None defined
Major	
101	Mask and tube not assembled and ready to use
102	Component part missing or damaged
103	Plastic brittle or aged
104	Tubing connector not rotatable at mask
105	Plastic tubing and distal connector not green tinted to identify oxygen use
106	Mask surface at face contact is not smooth
107	Foreign material molded in plastic
108	Foreign matter on parts
Minor	
201	Mask or tubing kinked or crushed out of shape

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Tensile strength and elongation. The tensile strength and ultimate elongation testing shall be performed in accordance with ASTM D-412.

4.4.2 Safety (skin contact) test. Apply a 0.5 inch by 0.5 inch piece of the face mask to the skin on the body side of the upper arm of five volunteers. The patch of plastic shall be held in place by means of a strip of adhesive tape of a width not exceeding the width of the plastic sample. The plastic sample shall be allowed to remain in contact with the skin for 24 hours. The patch shall be removed from the contact area and no reaction shall be visible. Within an additional 24 hours there shall be no delayed reaction on any of the subjects. If the patch unduly irritates the skin, the patch shall be removed immediately.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:00Z
Check the source to verify that this is the current version before use.

4.4.3 Security of mask-tube assembly. A two-pound weight shall be attached to the tubing while the mask is held in a clamp for 10 minutes.

4.5 Examination of preparation for delivery. An examination shall be made to determine whether the packaging, packing and marking comply with the requirements of Section 5. Defects shall be scored as specified in Table IV. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 with an AQL of 4.0 expressed in terms of defects per hundred units.

Table IV. Examination of Preparation for Delivery

Examine	Defect
Containers	Not as specified.
Contents	Not as specified.
Markings	Omitted; incorrect, illegible, improper size, location, sequence, or method of application.
Packaging and Packing Materials	Missing or damaged.
Workmanship	Bulging or distortion of containers. Cushioning inadequate, improper or missing.

4.5.1 Examination of closure and reinforcement for containers. When shipping containers are required to comply with PPP-B-636, examination for defects in closure and reinforcement shall be in accordance with the appendix to that specification.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One package containing fifty masks, as specified, constitutes one unit of issue.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:00Z
Check the source to verify that this is the current version before use.

5.1.2 Level A.

5.1.2.1 Individual package. Each mask shall be thoroughly clean and packaged along with instructions for use in a snug fitting, clear, 3 mil, heat sealed polyethylene bag. The mask and tubing shall be packaged to prevent the tubing and mask from being kinked or deformed.

5.1.2.2 Unit package. Fifty individual packages shall be packaged in a snug fitting box constructed in accordance with PPP-B-636, type CF, class domestic. Closure of the box shall conform to method I of PPP-B-636.

5.1.3 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A. Two unit packages shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-636, class weather-resistant. Closure, waterproofing and strapping shall be as specified in the appendix of the box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level B. Two unit packages shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-636, type CF, class domestic. Closure of the box shall conform to method I or II as specified in the appendix of the box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.2.4 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:00Z
Check the source to verify that this is the current version before use.

5.3 Marking.

5.3.1 Individual package. Each package shall be marked to identify contents and quantity. Marking may be printed on the bag or on an insert placed within the bag in such a manner so as to permit ready identification.

5.3.2 Unit packages. Each unit package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.3 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. The oxygen mask and attached tubing shall be used to administer oxygen therapy to a patient.

6.2 Ordering data. Procurement documents should specify the following:

(a) Number, title and date of this specification.
(b) Applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 This specification is intended to cover the following item appearing in the Federal Supply Catalog:

NSN	Item Identification
6515-00-392-0725	MASK, OXYGEN, Plastic Disposable, Under-the-Chin Type, 50s

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing Activity:
Defense Supply Agency - DM

Project No. 6515-1892
cs

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:00Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/ vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-M-37114 MASK, OXYGEN, PLASTIC, DISPOSABLE

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:00Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, Pa. 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

U.S.MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:00Z
Check the source to verify that this is the current version before use.