3.1.2 Barrel. Syringe barrels shall be fabricated from rigid, transparent, non-toxic, high-grade polypropylene plastic. The barrel shall be considered to be transparent if the scale can be seen in daylight when viewed through the empty barrel from the opposite side of the wall.

3.1.3 Plunger. The plunger rod shall be fabricated from rigid polypropylene plastic. The plunger tip shall be fabricated from fresh stock, pharmaceutical grade, non-blooming, non-leaching, non-toxic, age-resistant, inert, natural rubber or neoprene.

3.1.3.1 Inertness. The plunger tip shall show no evidence of a reaction with a high pH buffer solution when tested in accordance with 4.4.5.

3.1.4 Tube adapter. The tube adapter shall be fabricated from a suitable rigid plastic.

3.1.5 Tubing. The tubing shall be fabricated from clear medical grade polyvinyl chloride plastic and shall comply with the USP requirements for Class IV plastics.

3.1.6 Clamp. The slide clamp shall be fabricated from polypropylene.

3.1.7 Lubricant. The lubricant shall be an inert, non-toxic silicone compound suitable for lubricating the movement of the plunger in the barrel. The silicone compound shall be an FDA approved type.

3.2 Style, design and construction. The syringe shall be disposable with a specially formed and sized male nozzle to accommodate a mating female tubing adapter with a firm, non-leaking friction fit which shall be capable of withstanding the pressure and vacuum test hereafter described, without separation or leakage. The syringe shall be in accordance with DPSC Drawing No. 24354.

3.2.1 Barrel of syringe. The syringe barrel shall be closed at one end except for a specially tapered male nozzle, and open at the other end with a radial outside flange, of suitable size and shape to provide a finger grip and anti-roll features. The inside of the barrel shall have a very fine uniform surface, free from any imperfection that would affect the fluid tightness and smoothness of its fit with the rubber plunger tip. The inside surface of the barrel shall be lightly coated with the silicone lubricant specified in 3.1.7. Dimensions of the barrel shall be as specified in the DPSC drawing and shall permit a graduated capacity of 120 milliliters (ml) as indicated therein. The inside diameter shall be such as to provide an accurate fit with plunger tip, resulting in a fluid-tight fit, yet permitting the plunger to move freely throughout its excursion in the barrel. The ungraduated portion of the barrel, at the proximal end, shall be of sufficient length to permit the plunger tip to be aligned with the maximum graduation mark and still enclose the entire rubber tip. The ungraduated portion shall not exceed one-third of the overall length of the syringe barrel.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:59Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

POSTAGE AND FEES PAID

U.S. MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

DEFENSE PERSONNEL SUPPORT CENTER
DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:59Z
Check the source to verify that this is the current version before use.