MIL-T-37133A

29 September 1977
SUPERSEDING
MIL-T-37133
14 April 1975

MILITARY SPECIFICATIONS

TUBES, ENDOTRACHEAL, MURPHY, PLASTIC, DISPOSABLE

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers sterile, plastic, disposable endotracheal tubes of the Murphy design with inflatable cuff, pilot balloon and mechanical inflation-deflation valve.

1.2 Classification. Tubes shall be furnished in the following sizes, as indicated below (see 6.2).

Sizes		Inside Diameter (ID)
1	-	3 mm
2	-	4 mm
3	-	5 mm
4	-	6 mm
5	-	7 mm
6	-	7.5 mm
7	-	8 mm
8	-	8.5 mm
9	-	9 mm
10	-	10 mm

2 APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bids or request for proposals form a part of this specification to the extent specified herein.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Hq, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, DPSC-ATT, 2800 S. 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:02Z
Check the source to verify that this is the current version before use.

SPECIFICATIONS

FEDERAL

MMM-A-250	-Adhesive, Water-Resistant, (For Sealing Fiberboard Boxes).
PPP-B-566	-Boxes, Folding, Paperboard.
PPP-B-636	-Boxes, Shipping Fiberboard.
PPP-B-676	-Boxes, Set-up.
PPP-T-76	-Tape, Pressure-Sensitive Adhesive Paper, Water-Resistant, (For Carton Sealing).

MILITARY

MIL-S-36586	-Sterilization Test Strip Set, Bacterial Spore.

STANDARDS

MILITARY

MIL-STD-105	-Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	-Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center (DPSC)
Drawing No. 24339 ENDOTRACHEAL TUBES, MURPHY, PLASTIC, DISPOSABLE.

(Copies of specifications, standards, drawings, and publications required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply:

THE UNITED STATES PHARMACOPEIAL CONVENTION, INC.

The Pharmacopeia of the United States of America (USP).

(Application for copies shall be addressed to the Mack Publishing Company, Easton, PA 18042).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:02Z
Check the source to verify that this is the current version before use.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM) STANDARDS

ASTM D 412 - Tension Testing of Vulcanized Rubber.

ASTM D 2240 - Test for Indentation Hardness of Rubber and Plastics by Means of a Durometer.

(Application for copies should be addressed to American Society for Testing and Material, 1916 Race Street, Philadelphia, PA 19103.)

AMERICAN NATIONAL STANDARDS INSTITUTE (ANSI)

ANSI Z70.1 - American National Standard for Luer Tapers for Medical Applications.

ANSI Z79.5 - American National Standard for Murphy Tracheal Tubes.

(Application for copies should be addressed to the American National Standards Institute (ANSI, 1430 Broadway, New York, NY 10018.)

(Technical society and technical association specifications and standards are generally available for reference from libraries. They are also distributed among technical groups and using Federal agencies.)

3. REQUIREMENTS

3.1 Material. The material indicated herein shall be of the highest quality and suitable in all respects for the purpose intended.

3.1.1 Endotracheal tube. Shall be fabricated of polyvinyl chloride plastic and conform to the USP requirements for class IV plastics as described in 4.4.7. All sizes shall have a radiopaque stripe.

3.1.2 Cuff. Shall be made of polyvinyl chloride plastic conforming to the USP requirements for Class IV plastics as described in 4.4.7.

3.1.3 Injection tube and pilot balloon. Shall be made of polyvinyl chloride plastic.

3.1.4 Valve. The valve spring shall be made of stainless steel. The poppet in the valve shall be fabricated from neoprene. The balance of the valve shall be made of polyvinyl chloride plastic.

3.1.5 Connector. The connector shall be made of polypropylene or other suitable plastic.

3.2 Style, design and dimensions. Shall be a sterile endotracheal tube supplied with a 15 mm connector and with a permanently attached inflatable cuff, inflation tube, pilot balloon and mechanical inflation-deflation valve. Passage to the balloon shall be through a lumen in the wall of the endotracheal tube and proximally by means of a narrow inflation tube cemented into the lumen.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:02Z
Check the source to verify that this is the current version before use.

of the wall. At the proximal end of the inflation tube there shall be a pilot balloon and to this the mechanical inflation-deflation valve shall be attached. The endotracheal tube, cuff, pilot balloon and valve shall be in accordance with DPSC Drawing No. 24339. The radiopaque stripe shown in DPSC Drawing No. 24339 may be extended for the entire length of the endotracheal tube. Measurements shall be metric in accordance with ANSI Std. Z79.5. The Luer fitting on the inflation-deflation valve shall be in accordance with ANSI Std. Z70.1.

3.3 Pressure test. The cuff, inflation tube and inflation deflation valve assembly shall withstand an air pressure of 22 mm of mercury (Hg) for 30 minutes without leakage, when tested in accordance with 4.4.1. The 3 and 4 mm ID endotracheal tubes shall not be required to pass this requirement.

* 3.4 Tensile strength, elongation and hardness. The endotracheal tubes when tested as specified in section 4 shall have the following values:

Size	Tensile Strength (psi min.)	Elongation (min.)	Shore Durometer A (± 5)
1	2300	120%	85
2	2250	130%	85
3	2000	130%	85
4	2000	150%	83
5	2000	150%	83
6	2000	150%	83
7	2000	150%	83
8	1800	150%	83
9	1750	150%	83
10	1730	180%	83

3.5 Lead content. When tested as specified in 4.4.4 and 4.4.5, the endotracheal tube and cuff shall contain not more than 5 parts per million (ppm) of lead.

3.6 Sterility. The endotracheal tube shall be sterile when tested with a biological indicator as specified in 4.4.6.

3.6.1 Certificate of sterility. A certificate of sterility, stating that samples from each sterilizer lot have been tested and found to be sterile, shall be submitted to the procuring activity. The certificate of sterility shall include:

a. Item Identification.
b. National Stock Number.
c. Contract or purchase order number.
d. Sterility lot or control number.
e. Date of sterilization.
f. Quantity of product in sterilizer lot.
g. Type of sterilization (gas, radiation, heat, steam, etc.).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:02Z
Check the source to verify that this is the current version before use.

h. Form of biological indicator (i. e., inoculated paper spore strip or carrier, inoculated product, or inoculated simulated product).
i. Test organism(s) and population density used.
j. Test medium, incubation temperature, incubation time, and dates of tests.
k. Number of test units.
l. Statement that samples comply with sterility tests.
m. Signature of head of laboratory or a designated qualified alternate.

When inspection is made at designation, the certificate of sterility shall be submitted by the contractor to the Defense Personnel Support Center, Directorate of Medical Materiel, ATTN: DPSC-ATQ, 2800 South 20th Street; Philadelphia, PA 19101 and a copy forwarded with each shipment to the consignee. When inspection is made at source, certificate of sterility shall be furnished to the cognizant Government Quality Assurance Representative for submission to the procuring activity.

3.6.2 Sterility lot. A sterility lot is that single quantity of product subjected to the same manufacturing operation and sterilized simultaneously in the same sterilizer chamber. Each unit package shall be clearly marked with a sterility lot (control) number and date of sterilization as specified in 5.3.1. Each sterility lot shall be traceable to the pertinent manufacturing lot(s).

3.7 Certificates. When the contract provides for destination inspection only, certificates of sterility and biological safety, as specified below, shall be submitted by the contractor to Defense Personnel Support Center, Directorate of Medical Materiel, ATTN: DPSC-ATQ, 2800 South 20th Street, Philadelphia, PA 19101 and one copy of each shall be forwarded with each shipment to the consignee.

3.7.1 Certificate of sterility. A certificate of sterility, listing all sterility lot or control numbers, shall state that samples from each sterility lot or control have been tested for sterility, as rquired in 3.6 and 4.4.6, and subdivisions, and have been found to be sterile. The certificate shall reference the method used in determining sterility.

3.7.2 Certificates of biological safety. A certificate of biological safety, listing all lot or control numbers, shall state that samples from each lot or control have been tested for biological safety as specified in 3.1.1, 3.1.2 and 4.4.7, and were found to comply with such requirements.

3.8 Instructions. Each unit package shall be provided with instructions describing the proper techniques for dispensing the sterile endotracheal tubes from the containers. In addition, information as to the correct method of using the inflation-deflation valve shall be described.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:02Z
Check the source to verify that this is the current version before use.

3.9 Workmanship. Workmanship shall be first class throughout. The endotracheal tubes (including components) shall be free from defects which detract from their appearance or impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificate of quality. Certificate of quality, supplied by the manufacturer of the plastic, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the National Stock Number (NSN), the item identification, the name of component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as indicated in table I. The unit of product for examination purposes shall be one endotracheal tube.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:02Z
Check the source to verify that this is the current version before use.

TABLE I. Sampling criteria for examination.

Type defect	Inspection level	AQL (percent defective)
Major	S-3	2.5
Minor	S-2	4.0

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as indicated in table II. The unit of product for test purposes shall be one endotracheal tube. Where feasible, the same sample unit shall be used for the determination of two or more characteristics.

TABLE II. Sampling criteria for tests.

Component	Characteristic	Require-ment	Test procedure	Inspection level	AQL (percent defective)
Endotracheal tube	Dimensions	3.2	--	S-2	4.0
Inflation tube	Pressure test	3.3	4.4.1	S-2	1.5
Endotracheal tube	Tensile strength	3.4	4.4.2	S-1	1.5
Endotracheal tube	Elongation	3.4	4.4.2	S-1	1.5
Endotracheal tube	Hardness	3.4	4.4.3	S-1	1.5
Endotracheal tube	Lead	3.5	4.4.4 and 4.4.5	S-1	2.5
Cuff	Lead	3.5	4.4.4 and 4.4.5	S-1	2.5
Endotracheal tube complete	Sterility	3.6	4.4.6	S-2	1.0

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:02Z
Check the source to verify that this is the current version before use.

TABLE II. Sampling criteria for tests. Continued

Component	Characteristic	Require-ment	Test procedure	Inspection level	AQL (percent defective)
Endotracheal tube	Biological safety	3.1.1	4.4.7	S-1	1.5
Balloon	Biological safety	3.1.1	4.4.7	S-1	1.5

4.3 Examination. The endotracheal tubes shall be examined to determine compliance with all requirements contained in this specification. Examination shall be conducted as specified in tables I and III. Examination, however, shall not be restricted to defects listed in table III. Any endotracheal tube in the sample containing one or more defects shall be rejected and, if the number of defective tubes in any sample exceeds the acceptance number for that sample, the lot represented by the sample shall be rejected.

TABLE III. Classification of defects.

Categories	Defects
Major	
101	Endotracheal tubes not free from foreign matter.
102	Some components not present.
103	Depth markings incomplete, missing, illegible or incorrect (of drawing).
104	Endotracheal tubes not free from burrs, flashings or other defect affecting usage, or not free from roughness or defects which could be injurious to the patient.
105	Bevel edge at distal end of endotracheal tube harder than rest of tube.
106	Pilot balloon excessively thick or stiff and not readily indicative of inflation.
107	Cuff not reasonably symmetrical when inflated as required in 4.4.1.
Minor	
201	Instructions illegible, inadequate, or missing.
202	Marking incorrect, incomplete, illegible, or omitted.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:02Z
Check the source to verify that this is the current version before use.

4.3.3 Packaging inspection. The inspection of the unit package, intermediate package, packing and marking for shipment and storage shall be in accordance with quality assurance provision of the applicable container specification and the marking requirements of MIL-STD-129.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Pressure test. The cuff, inflation tube and inflation-deflation valve assembly shall be inflated to an air pressure of 22 mm Hg and this pressure maintained for 30 minutes. There shall be no evidence of leakage.

4.4.2 Tensile strength, elongation. Tensile strength and percentage elongation shall be determined in accordance with ASTM Method D 412, Tension Testing of Vulcanized Rubber.

4.4.3 Hardness. Hardness shall be determined in accordance with ASTM Method D 2240, Test for Indentation Hardness of Rubber and Plastics by Means of a Durometer.

4.4.4 Test for lead. Ten (10) Gms from the distal end of the endotracheal tube or from the cuff, whichever is being tested, shall be taken, ignited to char, and then further ignited in a muffle furnace at 700° C. to a white ash. The ash shall be boiled in 15 milliliters (ml) of 10 percent nitric acid, filtered into a volumetric flask and made up to 100 ml. An aliquot portion, equivalent to one gram of sample, shall be taken and placed in a separator. The dithizone test for lead, as given in the Pharmacopeia of the United States, shall be applied using a control of 5 ppm or 12 ppm of lead solution as given in the USP test. If the sample solution is darker than the limit, small aliquots shall be used until the color of the control is matched by the sample solution and the concentration of lead in ppm is determined.

4.4.5 Alternate test for lead. Accurately weigh a 5-gram sample into a 1-Oz. Vycor crucible. Reduce to ash in a muffle furnace at 650 C for eight hours. Let cool to room temperature and add 15 ml of 1:1 hydrochloric acid. Bring to a boil on a hot plate, allow to boil for 2 to 3 minutes, remove, let cool, and filter through Whatman #2 filter paper into a 50-ml volumetric flask. Rinse the residue on the filter with deionized water and dilute to 50 ml. Prepare standard lead solutions of 1, 2 and 5 ppm lead. Determine the absorbances of these solutions at the 283 millimicron wave length using a Hollow Cathode Pb lamp and 10x scale expansion on a spectrophotometer, such as the Perkin-Elmer Model 303 Atomic Absorption Spectrophotometr. Determine the absorbance of the sample solutions using the same conditions as with the standards.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:02Z
Check the source to verify that this is the current version before use.

Calculation:

$$\frac{A \times F \times \text{dilution}}{W} = \text{ppm lead wherein,}$$

A = Absorbance
F = Factor = $\frac{\text{(Concentration standard in ppm Pb)}}{\text{absorbance}}$ 1/

W = Sample weight in grams

1/ The three standards are averaged to give ppm/unit A.

4.4.6 Sterility test. The endotracheal tube shall be tested for sterility in accordance with the method specified in the USP for parenteral devices using inoculated product, inoculated simulated product, or paper spore strip (carrier). 1/ Product containing a biological indicator shall be marked or labeled so that it clearly differentiates from product intended for distribution. The incubation medium and minimum time in each case shall be soybean casein digest for 7 days for inoculated product, simulated product, or paper spore strip (carrier). The incubation temperature shall be as indicated in paragraph 4.4.6.1. The minimum number of test units, regardless of lot size, shall be 10.

1/ The preinoculum shall be placed at those surfaces which in the intact article are least accessible to the sterilization medium. Sterilization shall be accomplished in the assembled state. Sterility testing shall be done in the disassembled state, except for inoculated carrier which shall be 20 for Type I Sterilization and 30 for Type II Sterilization including inoculated carrier and product test units. The inspection level shall be as indicated in table II, Sampling criteria for tests.

4.4.6.1 Inoculated product or simulated product. For each type of sterilization, the organism used shall be as listed below:

Type of Sterilization	Organism	Incubation Temperature
Steam	B. stearothermophilus	55° - 65° C
Dry Heat	B. stearothermophilus or B. subtilis var. niger (or globigli)	55° - 65° C 30° - 35° C
Gas	B. subtilis var. niger (or globigii)	30° - 35° C
Radiation	B. pumilus	30° - 35° C

The number of viable cells shall be sufficient to provide assurance that the product has been sterilized.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:02Z
Check the source to verify that this is the current version before use.

4.4.6.2 Inoculated carrier. Inoculated carriers (paper spore strips) if used shall be in accordance with MIL-S-36586. Other paper carriers shall be inoculated to be of equivalent resistance. The method of incubation and number of test units shall be as indicated in 4.4.6 Sterility test and 4.4.6.1 Inoculated product or simulated product. The method of sterilization will control the one organism in the spore strip to be tested for as shown in 4.4.6.1. NOTE: When an inoculated carrier is chosen, product samples must also be tested as required by the USP.

4.4.6.3 Alternate method. Suppliers conducting sterility tests or designating lots in a manner other than that indicated herein shall submit proof that the alternate method provides equal assurance concerning the sterility of the product. However, all sterility tests shall be conducted using biological indicators. Requests for approval of any alternate method of testing for sterility shall be made to the Contracting Officer, and approval obtained before employment of the alternate method on a government contract.

4.4.6.4 Product testing. Testing for sterility of the product after receipt by the government shall be without a biological indicator and in accordance with the USP.

4.4.7 Biological safety. Samples of the endotracheal tube and of the cuff shall be tested in accordance with the USP requirements for Class IV plastics except that in the implantation test the following shall apply:

After "72 hours" add "and not to exceed 144 hours"; and delete: "Examine macroscopically the area of the tissue surrounding the center portion of each implant strip. Use a magnifying lens if necessary. Substitute: "The muscle tissue surrounding the implants shall be examined, both grossly and microscopically, by a pathologist certified by the American College of Veterinary Pathologists."

5. PACKAGING.

5.1 Packaging. Packaging shall be level A or C as specified (see 6.2).

5.1.1 Unit of issue. One package containing ten tubes of the size specified constitutes one unit of issue.

5.1.2 Level A.

5.1.2.1 Individual package. Each tube shall be packaged in a sealed peel-open type bag capable of maintaining sterility of contents unless damaged or opened. The bag shall be designed with tabs to facilitate opening.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:02Z
Check the source to verify that this is the current version before use.

* 5.1.2.2 Unit package. Ten individual packages, containing tubes of one size only, shall be packaged in a box conforming to PPP-B-566, or PPP-B-676. Closure of the bag shall be as specified in the appendix of the applicable box specification.

5.1.2.3 Intermediate package. Ten unit packages containing tubes of one size only shall be packed in a box conforming to PPP-B-636, type CF, class weather-resistant. Boxes shall be closed by taping the top and bottom flaps with 2-inch taper conforming to PPP-T-76. Closing of bottom flap with adhesive conforming to MMM-A-250 or metal stitches is permitted.

* 5.1.3 Level C. Ten tubes of one size only individually packaged as specified in 5.1.2.1 shall be further packaged in a manner that will afford adequate protection against deterioration and physical damage during shipment from the supply source to the first receiving activity.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A. Two intermediate packages, containing tubes of one size only, shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the appendix of the box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

* 5.2.2 Level B. Two intermediate packages, containing tubes of one size only shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-636, class domestic. Closure shall be as specified in the append of the box specification.

5.2.3 Level C. The packaged tubes shall be packed in shipping containers in a manner that will afford adequate protection against physical damage during direct shipment from the supply source to the first receiving activity. These packs shall conform to the applicable carrier's rules and regulations.

5.3 Marking.

5.3.1 Individual package. In addition to any commercial markings, each package shall be marked in accordance with the requirements of the Federal Food, Drug Cosmetic Act. In addition, the bag shall be marked with the name, and size of the commodity contained, the sterility lot (control) number, and the legends: "STERILE" and "STERILITY GUARANTEED UNLESS PACKAGE IS DAMAGED OR OPENED" or similar statement.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:02Z
Check the source to verify that this is the current version before use.

5.3.2 Unit package and intermediate package. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. Sterility lot (control) number, contract or purchase order number and name of contractor shall be shown. In addition, the method sterilization (G-Gas, R-Radiation C-Cold, S-Steam, D-Dry Heat Sterilization) shall be shown following the level of packaging and date of packaging. Example: "A-3/77-G". Final sterilization shall be on or after indicated packaging date.

5.3.3 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129. Sterility lot (control) number and the method of sterilization as specified in 5.3.2 shall be shown.

6. NOTES

6.1 Intended use. The tubes are intended for use in oral and nasal intubation.

6.2 Ordering data. Purchasers should exercise any desired options offered herein, and the procurement documents should specify the following:

(a) Title, number, and date of this specification.
(b) Size (see 1.2).
(c) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
(d) Applicable NSN.

6.3 National stock number coverage. This specification covers the following items appearing in the Federal Supply Catalog:

National Stock No.	Item Identification	Specification designation Size
6515-00-105-0653	TUBE, ENDOTRACHEAL, Murphy, with Cuff, 3 mm ID, 10s	1
6515-00-105-0664	TUBE, ENDOTRACHEAL, Murphy, with Cuff, 4 mm ID, 10s	2
6515-00-105-0707	TUBE, ENDOTRACHEAL, Radiopaque, Murphy, with Cuff, 5 mm ID, 10s	3
6515-00-105-0720	TUBE, ENDOTRACHEAL, Radiopaque, Murphy, with Cuff, 6 mm ID, 10s	4

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:02Z
Check the source to verify that this is the current version before use.

6515-00-105-0744	TUBE, ENDOTRACHEAL, Radiopaque, Murphy, with Cuff, 7 mm ID, 10s	5
6515-00-105-0759	TUBE, ENDOTRACHEAL, Radiopaque, Murphy, with Cuff, 8 mm ID, 10s	7
6515-00-105-0808	TUBE,ENDOTRACHEAL, Radiopaque, Murphy, with Cuff, 9 mm ID, 10s	9
6515-00-105-0815	TUBE, ENDOTRACHEAL, Radiopaque, Murphy, with Cuff, 10 mm ID, 10s	1(
6515-01-036-9034	TUBE, ENDOTRACHEAL, Radiopaque, Murphy, with Cuff, 7.5 mm ID, 10s	6
6515-01-036-9035	TUBE, ENDOTRACHEAL, Radiopaque, Murphy, with Cuff, 8.5 mm ID, 10s	8

6.4 Changes from the previous issue. The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Offerors and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the previous issue.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activity:
DOD-MB

Preparing activity:
Defense Logistics Agency-DM

Project No. 6515-2073

☆U.S. GOVERNMENT PRINTING OFFICE 1977-703-133/633

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:02Z
Check the source to verify that this is the current version before use.