A-A-53381
22 February 1988

# COMMERCIAL ITEM DESCRIPTION

## ARTICULATOR, DENTAL (Plain Line, Lingual Access Design)

The General Services Administration has authorized the use of this Commercial Item Description.

This commercial item description covers a plain line, lingual access design, dental articulator for use within the dental prosthetic laboratory.

Salient characteristics:

The articulator shall be suitable for occluding maxillary and mandibular anatomical casts, made from mouth impressions, so that accurate prothesis may be fabricated. The articulator shall have as part of the device, condylar and auditory pins for use with face bows, (see "Note" section), allowing full lingual visibility and accessibility to mounted casts. The articulator shall consist of a lower member and an upper member. The lower member shall be provided with right and left brass condylar elements. The lower member shall have two (2) small holes in the anterior portion for the purpose of accepting the mounting guide component, (see "NOTE" section). The upper member shall be provided with a tempered spring steel mechanism for holding the upper member upon the condylar element, permitting a protrusive movement of 30 $\pm$ 1 degrees, permitting right and left lateral excursions of 15 $\pm$ 1 degrees, permitting a Bennett movement of 15 $\pm$ 1 degrees and permitting removal of the upper member which is facilitated by quick release dogs. The articulator shall have an adjustable incisal pin with a lock thumbscrew which will provide a positive anterior stop. The articulator shall have a permanently fixed incisal guide, on the lower member, which shall provide a 10 $\pm$ 1 degrees protrusive incisal guidance. The articulator shall not require the use of a cast mounting ring, but shall have depressions into which stone shall be poured. After the stone has set, the casts shall be removed by pulling the tapered pins. The design and dimensions of the articulator shall be in accordance with Defense Personnel Support Center (DPSC) drawing number 24516.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Materials. All materials used shall be of high commercial quality and shall be entirely suitable for the purpose intended. The materials shall be as specified in Table I for the components and quantities listed.

Hardness. The steel spring shall have a hardness, as specified in Table I when tested as specified herein.

Weight. The weight of the articulator shall not exceed 22 ounces.

Finish. The finish of the components shall be as specified in Table I. Components shall be free of burrs and sharp edges. Plating shall be adherent, uniform in appearance, free from visible porosity, blisters, or other defects that may cause peeling.

Identification marking. Articulator shall be marked with the name or registered trademark of the manufacturer. Articulator of foreign (other than U.S.) manufacture shall be marked with the name of the country of origin, together with the name or registered trademark of the contractor. All marking shall be permanent and legible.

Descriptive literature. Each articulator shall be furnished with literature (leaflet form) containing articulator descriptive information together with manufacturer's part numbers for all components.

Workmanship. The articulator shall be free from defects which detract from its appearance or impair its serviceability.

TABLE I. Components, materials and finishes.

| Quantity | Component | Material | Finish |
|---|---|---|---|
| 1 | Upper member | Cast Aluminum Alloy No. 380 in accordance with QQ-A-591 for composition only. | Baked satin, blue, epoxy |
| 1 | Lower member | Cast Aluminum Alloy No. 356 in accordance with QQ-A-601 for composition only. | Baked satin, blue, epoxy |
| 1 | Thumbscrew | Thread size 10-32. | Knurled grasping edges |



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

TABLE I. Components, materials and finishes (continued).

| Quantity | Component | Material | Plating or Coating | Finish |
|---|---|---|---|---|
| 2 | Ball end | | | |
| 1 | Incisal Pin | Aluminum Alloy | | Machined |
| 2 | Cap | | | Machined sand blasted |
| 2 | Tapered pin | Austenitic 18 & 8 wire Stainless steel | - | Machined |
| 2 | Condylar shaft | Austenitic | | |
| 2 | Rollpin | Carbon Steel | Nickel | |
| 1 | Rollpin | | | |
| 2 | Lever | Brass (Extrusion) | | Machined |
| 1 | Incisal | Zinc Alloy | | Smooth |
| 1 | Guide Spring | Spring Steel C1075 or C1095 hardness 30N 66-68.5 Rockwell | Nickel | Smooth |
| 2 | Condylar Element | Brass Alloy 270 | | Smooth |

Unit. Each (EA). One articulator, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one articulator. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Tests.

Chemical analysis of the metals. In the absence of certificates of quality for the raw material, chemical analysis shall be conducted in accordance with Method 111 or 112 of FED-STD-151.

Plating thickness. Thickness of plating measurements shall be made by any one of the test methods described in QQ-N-290.

Hardness. Hardness shall be determined, in accordance with American Society for Testing and Materials (ASTM) Designation E 18 (Rockwell) or E 92 (Vickers or Diamond Pyramid).

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.



Source: https://assist.dla.mil -- Downloaded: 202640-26T09:18Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

| NSN | ITEM IDENTIFICATION |
|---|---|
| 6520-01-114-0485 | ARTICULATOR, DENTAL (Plain Line, Lingual Access Design). |
| 6520-00-935-5826 | FACE BOW, ARTICULATOR, DENTAL. |
| 6520-01-005-2557 | FACE BOW, ARTICULATOR, DENTAL. |
| 6520-01-157-9124 | FACE BOW, ARTICULATOR, DENTAL. |



Source: https://assist.dla.mil -- Downloaded: 202610-26T09:18Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

CIVIL AGENCY COORDINATING ACTIVITIES:
VA-OSS
PHS
FDA-MPQAS

49113

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

Project No. 6520-2009

Location: NCR CID DISK 3



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.