A-A-53385
12 February 1988

COMMERCIAL ITEM DESCRIPTION

ADHESIVE, IMPRESSION MATERIAL, DENTAL (Rubber Base)

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers rubber base dental impression material adhesive suitable for bonding dental rubber base impression material to impression trays.

Salient Characteristics:

The adhesive material shall be suitable for use as a bonding agent between polysulfide rubber base impression material and a metal or plastic tray during restorative dental procedures, when used in accordance with manufacturer's instructions.

The impression material adhesive shall be a fast-drying solution of a synthetic elastomer, suitable for bonding dental impression material to impression trays. The adhesive solvents shall be methyl ethyl ketone, toluol, and acetone. The solids content of the synthetic elastomer shall be 24 plus or minus 3 percent when tested as specified in test section under solids content.

The flash point of the solvents shall be 0-18 degrees Fahrenheit (^{o}F) when tested in accordance with ASTM D1310-86.

Container. Immediate container shall be a light-resistant glass bottle with a metal screw cap with liner. As an alternate, the screw cap shall be a phenolic cap with a low density polyethylene liner. An applicating brush shall be affixed to the underside of the screwcap. The brush shall be fabricated of animal-type bristles, or nylon filaments. The bristles shall be of sufficient quantity to hold and apply the adhesive in the function intended.

The applicating brush shall be strong and durable to withstand repeated applications and remain functional for the life of the adhesive. The liner and brush shall be compatible with the adhesive, with no deterioration of the liner or contamination of the adhesive.

AMSC/NA FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Irritation. Unless otherwise specified, prior to delivery the manufacturer shall furnish a statement directly to the Directorate of Medical Materiel, Technical Services Branch, DPSC-RST, Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pennsylvania 19101-8419 stating that the material does not normally irritate oral tissues and does not contain poisonous ingredients in sufficient concentration to be harmful to human beings when used as directed.

Identification marking. A permanent label shall be applied to each screw-capped bottle. The legend on the label shall include the manufacturer's name, item name, lot or control numbers, and the term "Flammable." The legend shall be permanent and easily legible.

Instructions for use. Adequate and accurate instructions for using the adhesive shall accompany each unit package.

Delivery schedule. Not more than 3 months shall have elapsed from the date of manufacture of the impression adhesive to the date of delivery to the Government.

Workmanship. The adhesive shall be free of lumps and foreign matter and any defect which detracts from its appearance or impairs its serviceability.

Unit. Package (PG). One package, containing one bottle (2 oz) and one applicator, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one bottle of adhesive. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Tests.

Solids content. The solids content shall be tested for "Loss on Drying" in accordance with the Pharmacopeia of the United States of America (USP). The sample shall be dried at 120° plus or minus 2°F for 24 plus or minus 1 hours. Care should be exercised when volatile materials are of such a nature as to constitute toxic, fire, or explosion hazards. The percentage of total solids shall be calculated as follows:

$$\text{Total solids, percent} = \frac{\text{Weight of residue}}{\text{Weight of sample}} \times 100$$

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

National Stock Number	Item Identification
6520-00-765-0986	ADHESIVE, IMPRESSION MATERIAL, DENTAL, 2 oz.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

48359

Project No. 6520-2027

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C
POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-ATTS
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2 DOCUMENT TITLE

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5 PROBLEM AREAS

a. Paragraph Number and Wording

b Recommended Wording

c Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.