A-A-53393
1 March 1988

COMMERCIAL ITEM DESCRIPTION

OSTEOTOME SET, (Lambotte)

The General Services Administration has authorized the use of this Commercial Item Description.

This commercial item description covers a Lambotte style osteotome set suitable for use in bone cutting surgical procedures.

Salient characteristics:

General. The osteotome set shall be a Lambotte style set of twelve osteotomes, with a rack. The osteotome set shall contain six straight and six curved osteotomes.

Components. Blade widths of the osteotomes shall be as follows:

Style	Blade width (inches)
Straight	1/4, 1/2, 3/4, 1-1/4, 1-1/2
Curved	1/4, 1/2, 3/4, 1-1/4, 1-1/2

The rack shall be designed to securely hold, organize and protect the osteotomes during sterilization procedures and storage. The rack shall have a handle which enables the osteotome set to be carried in an upright position.

Material. The osteotomes shall be fabricated of Class 4 martensitic corrosion-resistant steel in accordance with ASTM Designation F899. The rack shall be fabricated of corrosion-resistant metal alloy.

Hardness. The osteotome set shall have a Rockwell Hardness C Scale (HRC) 50 to 60. The hardness shall be tested in accordance with ASTM Designation E18 or ASTM Designations E92 and E140.

Style, design and dimensions. Style, design and dimensions of the osteotome shall be in accordance with Figure 1.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

Passivation. The osteotome set shall be properly passivated.

Finish The osteotomes shall have a mirror or satin finish and shall show no sharp exterior edges except where a sharp edge serves a functional purpose.

Sterilization. The osteotome set shall be capable of withstanding repeated sterilizations and show no evidence of corrosion when autoclaved at 270 ± 5 degrees F and 27 ± 1 psi for a minimum of 20 minutes.

Identification markings Osteotomes shall be marked "stainless " Manufacturers name (or registered trademark), manufacturers catalog number, and blade width shall be marked on the instrument.

Workmanship. The osteotome set shall be free from burrs, cracks, scratches, slivers and other defects which detract from its appearance or impair its serviceability.

Set (SE . One osteotome set, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Responsibility for compliance. All items must meet all requirements specified herein. The inspection set forth in this specification shall become a part of the contractor's overall inspection system or quality program. The absence of any inspection requirements in the specification shall not relieve the contractor of the responsibility of assuring that all products or supplies submitted to the Government for acceptance comply with all the requirements of the contract. Sampling in quality conformance does not authorize submission of known defective material, either indicated or actual, nor does it commit the Government to acceptance of defective material.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The acceptable quality level (AQL) shall be 0.65 for an inspection level of S-2 The unit of product for test purposes shall be one osteotome set.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound-units a request should be made to the contracting officer to determine if the product is acceptable

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE The following National Stock Number is covered by this document

NSN	ITEM IDENTIFICATION
6515-01-124-9778	OSTEOTOME SET, Lambotte, Set of 12

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS

Custodians

Army - MD
Navy - MS
Air Force - 03

Preparing Activity

DoD-MB

Agent

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:


VA-OSS
PHS
FDA-MPQAS

PGC 59308


Project No. 6515-3896

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

A-A 53393



TYPICAL PLAN VIEW FOR BOTH STRAIGHT AND CURVED OSTEOTOMES



W
.25
50
.75
1.00
1.25
1.50

NOTE.
ALL DIMENSIONS ARE IN INCHES.

FIGURE 1 LAMBOTTE OSTEOTOMES

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 Alexandria, VA

POSTAGE WILL BE PAID BY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53393	OSTEOTOME SET, (Lambotte)

3a. NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER (Last, First, MI) – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS (Street, City, State, ZIP Code) – Optional	8 DATE OF SUBMISSION (YYMMDD)

DD FORM 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.