COMMERCIAL ITEM DESCRIPTION

WAX, DENTAL, Occlusal Indicator

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a dental wax suitable for use in occlusal registration purposes.

Salient characteristics:

The dental wax shall be soft, pliable and easy to manipulate. The material shall be furnished in stamped, segmented sheets mounted on heavy paper. The paper shall hold the wax and keep the sheets separated while providing for easy removal. The wax sheets shall be coated on one side with a water soluble adhesive. The adhesive shall be suitable for use in the mouth in dental procedures.

The dental wax shall be suitable for use in the oral cavity for diagnosing occlusal prematurities to obtain balanced occlusal equilibrium of natural teeth, individual restorations, fixed bridges, partial and full dentures. The wax when used in accordance with the manufacturer's instructions shall indicate occlusal prematurities by the development of transparent areas in the wax on the contact teeth. The dental wax shall be a mixture of waxes, resins, and hydrocarbons of the paraffin series. The dental wax shall have a minimum flow of 60.0 percent at 38.0 $\pm$ 0.1 degrees Celsius when tested as specified herein.

The dental wax shall be dark green in color to contrast the tooth structure. Any color additives shall be approved and certified by the Food and Drug Administration under the Federal Food, Drug and Cosmetic Act.

The dental wax shall be 30 sheets minimum and weigh a minimum of 1-3/4 ounces. The wax sheet shall measure 3-3/4 $\pm$ 1/32 inches by 1-3/4 $\pm$ 1/32 and shall have a thickness of 0.015 $\pm$ 0.001 inch. Each sheet shall be stamped into six segments or strips for ease in separating sufficient wax for each treatment. Each segment or strip of wax shall measure 1-3/4 $\pm$ 1/32 inches by 5/8 $\pm$ 1/32 inch.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

Each unit package shall be provided with a soft, water-proof pencil. The pencil shall be suitable for use in the oral cavity for marking the areas indicated for occlusal adjustment through the transparent areas of the dental wax.

Adequate, accurate and complete instructions for use shall accompany each package.

Not more than three months shall have elapsed from the date of manufacture of the dental wax to the date of delivery to Government.

Workmanship. The wax shall be clean, uniform in quality and free from foreign material or any other defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package, containing 30 sheets of wax, with marking pencil, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:19Z
Check the source to verify that this is the currenversion before use.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more characteristics. The unit of product for inspection purposes shall be one package of wax with marking pencil. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Tests.

Flow. The flow shall be tested in accordance with ANSI/ADA Specification No. 4 Dental Inlay Casting Wax, with the following exception: The value for flow at any temperature shall be the average value for five specimens from each lot.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:19Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented for Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6520-00-784-4974	WAX, DENTAL, Occlusal Indicator, 30's, Green

Ordering data. Intermediate/exterior quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTEREST:

Custodians:

Army-MD
Navy-MS
Air Force - 03

Preparing Activity:

DoD-MB

AGENT:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PROJECT NO. 6520-2047

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-ATTS
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE

3a. NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5 PROBLEM AREAS

a. Paragraph Number and Wording

b Recommended Wording

c Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last, First, MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.