INCH-POUND

A-A-53400/19
20 March 1992

COMMERCIAL ITEM DESCRIPTION SHEET

FORCEPS, HEMOSTATIC, KOCHER, 5-1/2 INCH

The General Services Administration has authorized the use of this Commercial Item Description Sheet as a replacement for Military Specification MIL-F-42045 which is cancelled.

The complete requirements for procuring Forceps, Hemostatic, Kocher, 5-1/2 inch shall consist of this document and the issue in effect of Commercial Item Description A-A-53400.

Salient characteristics:

Style, design and dimensions. This Commercial Item Description Sheet covers the following types of Kocher forceps:

Type I - Straight

Type II - Curved

Shall be in accordance with figures 1 or 2 for applicable type, design, and dimensions.

NOTES:

a. Intended use: The Hemostatic Forceps of the Kocher design covered in this Commercial Item Description Sheet are intended for accomplishing hemostasis in general surgical procedures.

b. Copies of Commercial Item Description A-A-53400 may be obtained by writing Navy Publishing and Printing Service Office, Bldg 4D NPM-DoDSSP, 700 Robbins Avenue, Philadelphia, PA 19111-5094.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

c. National Stock Number and Item Identification.

NSN	Item Identification	Type
6515-01-138-5014	FORCEPS, HEMOSTATIC, Straight, Kocher, 5-1/2 Inch.	I
6515-01-138-5015	FORCEPS, HEMOSTATIC, Curved, Kocher, 5-1/2 Inch.	II

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:


DLA-DM


CIVIL AGENCY COORDINATING ACTIVITIES:


VA-OSS
USPHS
FDA-MPQAS


Project No. 6515-5242

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.










NOTE:
DIMENSIONS IN FULLY CLOSED POSITION.


UNLESS OTHERWISE SPECIFIED DIMENSIONS ARE IN INCHES		
TOLERANCES ON		
FRACTIONS	DECIMALS	ANGLES
± 1/64	±	±


NSN 6515-01-138-5014


FIGURE 1 FORCEPS, HEMOSTATIC, 5-1/2 INCH, STRAIGHT, KOCHER


Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.











NOTE:
DIMENSIONS IN FULLY CLOSED POSITION.

UNLESS OTHERWISE SPECIFIED DIMENSIONS ARE IN INCHES		
TOLERANCES ON		
FRACTIONS	DECIMALS	ANGLES
±1/64	±	±

NSN 6515-01-138-5015

FIGURE 2 FORCEPS, HEMOSTATIC, 5-1/2 INCH, CURVED, KOCHER

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-53400/19	20 March 1992

3. DOCUMENT TITLE: FORCEPS, HEMOSTATIC, KOCHER, 5-1/2 INCH

4. NATURE OF CHANGE (*Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.*)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (*Last, First, Middle Initial*)	b. ORGANIZATION	
c. ADDRESS (*Include Zip Code*)	d. TELEPHONE (*Include Area Code*) (1) Commercial (2) AUTOVON (*If applicable*)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (*Include Area Code*)	
Defense Personnel Support Center ATTN: MSE	(1) Commercial (215)737-2870	(2) AUTOVON 444-2870
c. ADDRESS (*Include Zip Code*) 2800 S. 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

Source: https://assist.dla.mil -- Downloaded: 2020-10-20T09:19Z
Check the source to verify that this is the current version before use.

DD Form 1426, OCT 89 *Previous editions are obsolete.* 198-290