COMMERCIAL ITEM DESCRIPTION

FORCEPS, HEMOSTATIC

The General Services Administration has authorized the use of this Commercial Item Description in preference to Federal Specification GG-I-526B.

This commercial item description covers metallic, reusable, hemostatic forceps used for accomplishing hemostasis during surgery.

General.

Commercial Item Description Sheet. The forceps shall be in accordance with all requirements specified herein and the applicable Commercial Item Description Sheet (slash sheet-see Notes).

Salient Characteristics:

The definitions of the terms and components used in this specification are delineated in ASTM F921. Performance shall be in accordance with the requirements specified in ASTM F1026. Corrosion requirements shall be in accordance with ASTM F1089.

Material. The material shall be class 4, corrosion resistant steel, in accordance with ASTM F899.

Hardness. The Rockwell hardness of the instrument shall be C40 to C48.

Style, design and dimensions. Shall be in accordance with the applicable Commercial Item Description Specification Sheet and as specified herein (See Notes).

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

Finish. The finish of the hemostatic forceps shall be mirror, satin, or matte. The matte finish shall exhibit a uniform nondirectional surface texture equivalent to that produced by lightly blasting the prepared surface with fine abrasives under pressure.

Identification markings. In addition to the requirements specified in ASTM F1026 the forceps shall also be marked "stainless" and shall bear the name or registered trademark of the contractor.

Tests. Corrosion resistance, hardness, performance, and dimensions shall be tested as specified under Quality Assurance Provisions.

Workmanship. The forceps shall be free from defects which detract from its appearance or impair its serviceability and in accordance with ASTM F1026.

Unit. Each (EA). One forceps, as specified, constitutes one unit.

Product Quality Level.

Visual and Dimensional Examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

QUALITY ASSURANCE PROVISIONS

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and Inspection Level II. Defects shall be those listed below. The unit of product for examination purposes shall be one forceps.

DEFECTS

Instrument not free from cracks, pits, burrs, sharp edges, scale, rough areas and corrosion.

Finish of the instrument not as specified.

Style and design not as specified.

Identification markings not present, complete, correct, or legible.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified below. The unit of product for test purposes shall be one forceps.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

Character-istic	Requirement	Test Procedures	Inspec. level	AQL
Composition Analysis	As specified herein	As specified herein and ASTM F899	*	----
Hardness	As specified herein	ASTM F1026	S-2	1.0
Dimensions	As specified herein		S-1	2.5
Copper Sulfate test	ASTM F1089	ASTM F1089	S-2	1.0
Boil test	ASTM F1089	ASTM F1089	S-2	1.0
Performance	ASTM F1026	ASTM F1026	S-2	1.0

* Sufficient amount from each lot of each metal used in the construction of the instrument shall be selected for composition analysis.

Tests.

Corrosion resistance. Shall be tested as specified in ASTM F1089.

Composition analysis. Shall be subjected to analysis in accordance with methods 111 or 112 Federal Test Method Standard No. 151.

Hardness. Hardness shall be tested as specified in accordance with ASTM F1026.

Performance. Performance shall be tested in accordance with applicable requirements of ASTM F1026.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting of rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, th offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances and regulations.

Recovery material. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

Notes.

Ordering Data. Purchasers should select the preferred options permitted herein, and include the following information in procurement documents:

(a) Title, number, and date of this specification.
(b) Title, number, and date of the applicable specification sheet which shall include this specification letter and number, and the slash number. Examples:

A-A-53400/1	Forceps, Hemostatic, Halsted.
A-A-53400/2	Forceps, Hemostatic, Rochester-Pean.
A-A-53400/3	Forceps, Hemostatic, Kelly.
A-A-53400/4	Forceps, Hemostatic, Crile.
A-A-53400/5	Forceps, Hemostatic, Hartman.
A-A-53400/6	Forceps, Hemostatic, Rochester-Ochsner.
A-A-53400/7	Forceps, Hemostatic, Mayo-Carmalt.
A-A-53400/8	Forceps, Hemostatic, Mixter.
A-A-53400/9	Forceps, Hemostatic, Pean.
A-A-53400/10	Forceps, Hemostatic, Rankin.
A-A-53400/11	Forceps, Hemostatic, Pratt.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4246

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge (*DO NOT STAPLE*), and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53400	FORCEPS, HEMOSTATIC

3a. NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street City State ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a. NAME OF SUBMITTER *(Last First MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street City State ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:19Z
Check the source to verify that this is the current version before use.