
FED. SUP CLASS
6640



MS36451-1 MS36451-2 MS36451-3 MS36451-7
MS36451-8

CANCELLED
CANCELLED AFTER 10 JUN 1991
NO SUPERSEDING STANDARD

	150	16
		16
		16

MATERIAL: Tube ... be of clear, colorless plastic. Friction type caps shall be of rigid plastic. Applicators may be of wood or plastic, tipped with surgical cotton or polyester-fiber.

DESIGN: Applicator shall consist of a wood or plastic stick tipped with either cotton or plastic fibers and mounted in a cylindrical closed bottom tube with a flexible friction type or screw type cap. All screw capped swab applicators shall be coded as follows:

Red Cap - denotes cotton tip
Green Cap - denotes polyester fiber tip

MS36451-1 shall have a friction cap that has two sockets to accommodate two cotton-tipped wood applicators.

MS36451-2 shall have a friction cap that supports one cotton-tipped applicator.

MS36451-3 shall have a friction cap that supports a cotton-tipped wood applicator shielded to assist insertion and reduce contamination of swab during insertion. Shield shall be molded of plastic.

MS36451-4 shall have a screw cap that supports a single wood cotton tipped applicator.

Ⓑ ENTIRE STANDARD REVISED Ⓒ CANCELLATION

P.A. FA Other Cust MS 99	INTERNATIONAL INTEREST	TITLE TUBE, BIOLOGICAL CULTURE SAMPLING	MILITARY STANDARD MS36451
Procurement Specifications NONE	SUPERSEDES:		SHEET 1 OF 2

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE Project Number 6640-1628-03

CANCELLATION DATE

APPROVED 14 December 1967 REVISED Ⓐ 17 August 1970 Ⓑ 19 Feb 1982 Ⓒ 10 JUN 1991

USER ACTIVITIES: ARMY - ME, DM
NAVY - MS

REVIEWER ACTIVITIES: ARMY - ME, MD
NAVY - MS
AIR FORCE - 03

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:41Z
Check the source to verify that this is the current version before use.

MS36451-5 shall have a screw cap that supports a single wood or plastic polyester fiber-tipped applicator.

MS36451-6 shall have a screw cap that supports a cotton-tipped flexible plastic applicator.

MS36451-7 shall have a screw cap that supports two cotton-tipped wood applicators.

MS36451-8 shall have a screw cap that supports two cotton-tipped plastic applicators.

NOTES:

1. Applicator assemblies shall be sterile and wrapped in plastic film strip packs with individually sealed compartments.
2. Applicator assembly shall be discarded after a single use.
3. Items differing from this Military Stand- iations in design or dimensions are considered document, provided such variations do e not in conflict with the ma-

CANCELLED
CANCELLED AFTER 10 JUN 1991
NO SUPERSEDING STANDARD

USER ACTIVITIES:

REVIEWER ACTIVITIES:

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

APPROVED REVISED

P.A. Other Cust	INTERNATIONAL INTEREST	TITLE TUBE, BIOLOGICAL CULTURE SAMPLING	MILITARY STANDARD MS36451
Procurement Specifications	SUPERSEDES:		SHEET 2 OF 2

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

DATE

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:41Z
Check the source to verify that this is the current version before use.