A-A-54152
15 May 1990

COMMERCIAL ITEM DESCRIPTION

CLAMP, UMBILICAL CORD, Plastic

The General Services Administration has authorized the use of this commercial item description as a replacement for Military Specification MIL-C-36136B, which is cancelled.

This commercial item description covers a sterile, disposable, self-locking plastic clamp used on the umbilical cord of a new born infant during obstetric or pediatric procedures.

Salient characteristics:

General. The clamp have a 'T' shaped lock mechanism with 3 contact points and once closed, shall open only when cut at the hinge. When closed, the jaw teeth shall intermesh. The lock design should be such that outward bowing of the clamp jaws causes the lock mechanism to strengthen in use. The jaw faces shall have matching relief grooves, centrally located to provide dual clamping action. Both outer jaw surfaces shall have serrations at the blind lock ends for gripping. The hinge shall have an umbilical cord guard. The clamp shall be 2-3/16 inches long and weigh 2-1/2 +/- 1/4 grams each. Every unit package of 125 clamps shall be provided with one clamp clipper. The clipper shall be a hand-held device with shielded cutting edge designed to cut through the clamp hinge with hand held pressure

Material. The clamp shall be fabricated from autoclavable, non-toxic, non-irritating, rigid plastic. The clipper shall be fabricated from rigid plastic with a nickel-plated brass eyelet and a stainless steel blade. As an alternate, the clipper may be fabricated from chromium over nickel-plated steel.

Style, design and dimensions. The clamp shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No. 20597 for style, design and dimensions

Sterilization. When tested as specified herein, the clamps shall not distort, crack or be adversely affected such as to impair their use.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

Sterility The clamps shall be sterile in accordance with the USP or AAMI requirements Sterility assurance level shall be 0.000001.

Instructions Each unit package shall be provided with instructions which include illustrations, describing proper application and removal of the clamp

Identification markings Each clamp shall be marked with the contractor's name or registered trademark. Each immediate container shall be marked "sterile" along with instructions for removal of the clamp.

Workmanship The clamp and clipper shall be free from defects which detract from their appearance or impair their serviceability

Unit Package (PG). One package containing 125 clamps and one clipper, as specified, constitutes one unit. Each clamp shall be individually packaged sterile in a suitable package capable of maintaining sterility of the contents unless package is damaged or opened.

Quality Assurance Provisions

Responsibility for inspection Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate

Inspection Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:29Z
Check the source to verify that this is the currentversion before use.

Sampling for examination Sampling for examination shall be conducted in accordance with MIL-STD-105, and as specified herein Unit of product for examination purposes shall be one clamp Sampling shall be level II and an AQL of 1.0.

Examination. Clamps shall be examined to determine compliance with specification requirements

Sampling for tests Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one clamp. AQL shall be 1.0 (percent defective) with an Inspection Level of S-2. (See Note 2).

Tests Tests shall be conducted to determine compliance with specification requirements. where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sterilization test Clamp shall be steam autoclaved, for 20 minutes at 250 degrees Fahrenheit (°F) and 15 pounds per square inch gage (psig), then allowed to cool room temperature (68-72 °F)

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as many be otherwise provided for under the provisions of the contract

Metric products Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and the regulations promulgated thereunder In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the Supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities, labels and markings shall be as specified in the contract and/or order

NOTE 1· The following National Stock Number is covered by this document

NSN	ITEM IDENTIFICATION
6515-00-890-1541	Clamp, Umbilical Cord, Plastic, Disposable, 2-1/4 Inch, 125's

NOTE 2: Examination and testing other than sterilization testing may be waived provided certificates of compliance can be supplied confirming compliance with specification requirements as verified through process control monitoring, in-house physical testing, measuring, and visual examination.

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4445

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 Alexandria, VA

POSTAGE WILL BE PAID BY

Commander
Defense Personnel Support Center
ATTN: DPSC-RST
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

(TO DETACH THIS FORM CUT ALONG THIS LINE.)

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-54152	CLAMP, UMBILICAL CORD, Plastic

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b ADDRESS *(Street, City, State ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5 PROBLEM AREAS

a Paragraph Number and Wording

b Recommended Wording

c Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last, First, MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street, City, State ZIP Code)* Optional	8 DATE OF SUBMISSION *(YYMMDD)*

DD FORM 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil - Downloaded: 2026-04-26T09:29Z
Check the source to verify that this is the current version before use.