A-A-54154
17 May 1990

COMMERCIAL ITEM DESCRIPTION

CATHETER AND NEEDLE UNIT, INTRAVENOUS, RADIOPAQUE, Disposable, 50s (Catheter into Needle)

The General Services Administration has authorized the use of this commercial item description as a replacement for Military Specification MIL-C-37994(DLA-DM), which is cancelled.

This commercial item description covers sterile, non-toxic, non-pyrogenic, disposable, catheter and needle units suitable for prolonged administration of intravenous medication from a single venipuncture. The unit shall be ready for use and shall permit aseptic insertion by the surgeon without the necessity of hand scrubbing or wearing of sterile gloves.

Salient characteristics:

General. Catheter and needle units shall be furnished as a combination of the following designations:

Catheter Size (Gage)	Needle Size (Gage)	Usable Catheter length (Inch)*	Material Type
A. 16 B. 18.5 or 19	a. 14 b. 17 c. 20	1. 8 2. 12 3. 24	I Polyvinylchloride (PVC)
C. 22		4. 36	II Polytetrafluroethylene (PTFE) III polyurethane

* Tolerance: +\- 1 inch

Material.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

Catheter. The catheter shall be fabricated from clear polyvinylchloride (PVC), polytetrafluroethylene (PTFE) or polyurethane which meets the requirements of the United States Pharmacopeia (USP) for Class IV plastics (See Note 2). The catheter hub shall be fabricated from rigid or semi-rigid plastic which meets the requirements of USP for class II plastics.

Needle and needle collar. The needle shall be fabricated from Class 3 austenitic stainless steel in accordance with ASTM Designation F899. The needle collar shall be fabricated from rigid plastic.

Wire stylet. The wire stylet shall be fabricated from Class 3 austenitic stainless steel in accordance with ASTM Designation F899.

Protective sleeve. The protective sleeve shall be fabricated from transparent or translucent flexible polyethylene.

Needle cover, needle guard, and flow control plug. The needle cover, needle guard, and flow control plug shall be fabricated from rigid plastic.

Style and design. The catheter shall consist of a needle, the lumen through which an indwelling catheter shall pass. The needle shall have a lumen 2 + 1/8 inches long and be of short bevel design. The needle point tip shall be ground to maximum keenness and the needle point heel shall be dull to prevent coring. The needle hub shall make a secure fit when pushed through the needle collar, which is permanently secured to the protective sleeve. The unit shall be supplied with a needle guard for safety prior to use and a folding needle cover for covering the needle after being withdrawn, following venipuncture.

Following venipuncture and needle withdrawal, the needle cover shall enclose and lock the entire needle cannula, preventing the needle bevel from shearing the catheter.

The catheter shall be permanently secured to the Luer-lock catheter hub if of two-piece construction. If of one piece construction, the proximal end shall incorporate a Luer-lock hub. The distal edges of the catheter shall be free from burrs and cuts. The wire stylet shall move freely within the catheter and have a flow control plug permanently secured to its proximal end.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

The flow control plug shall have a Luer taper to provide leakproof seal with the catheter hub. A protective sleeve, covering from the needle collar to over the flow control plug, shall maintain the sterility of the catheter unit opened.

Catheter hub and flow control plug. The female Luer-lock fitting on the catheter hub and the male Luer slip fitting on the flow control plug shall comply with the dimensional and performance requirements of ANSI/HIMA Standard MD70.1.

Catheter flexibility. With the wire stylet removed, the catheter shall be wrapped 360 degrees around a 3/8 inch diameter rod without kinking or cracking. Water shall then be admitted through the catheter hub within 5 seconds of being removed from the rod and observed for restricted flow from the distal tip.

Security of assembly. When the following joints are subjected to the indicated tensile forces, they shall not separate or pull apart: 16 gage catheter to hub-4 pounds, 18.5 or 19 gage catheter to hub 2- pounds, 22 gage catheter to hub-1 pound, flow control plug to wire stylet- 1.5 pounds minimum, protective sleeve to needle collar- 3 pounds, and needle (lumen) to hub- 10 pounds.

Catheter/hub leakage. The catheter/hub joint shall not leak when subjected to the water leakage test specified in ANSI/HIMA MD70.1. As an alternate, the joint shall not show leakage of air bubbles when checked underwater for 6 seconds with 35 psig air for Type I or II catheters, and 1 minute with psig air if Type III catheter.

Radiopacity. Catheters shall be radiopaque in accordance with ASTM Designation F640.

Heavy metals extractability. Plastic components coming in contact with intravenous solution or patient's blood shall be tested. The quantity of lead and cadmimum shall not exceed 5 parts per million (ppm) and the amount of barium shall not exceed 10 ppm when tested in accordance with the USP (See Note 3).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

Toxicity, pyrogenicity, sterility. The finished catheter and needle unit shall be nontoxic and nonpyrogenic in accordance with the USP for Transfusion and Infusion assemblies (the rabbit pyrogen test) or the Bacterial Endotoxins test (LAL) with the rabbit pyrogen test used for LAL test failure. The catheter and needle unit shall be sterile with a Sterility Assurance Level of 1 x 10-6 in accordance with AAMI guidelines.

Sterility lot and traceability. A sterility lot is that quantity of product sterilized in the same sterilizing chamber at the same time. Each quantity of product shall be traceable to a given manufacturing operation or lot. Each unit package shall be traceable to a given sterility or manufacturing lot (See Note 4).

Instructions. Each catheter and needle unit package shall contain instructions for use. These instructions shall include a caution note stating in effect "Do not attempt to retract catheter while needle is the vein".

Workmanship. The catheter and needle unit shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing 50 individually packaged catheter and needle units of one size only with "Instructions for Use" as specified, constitutes one unit. The individual package shall be a sealed, peel-open type container or other suitable sealed container capable of maintaining sterility of contents unless package is damaged or opened.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, Inspection Level S-1 and AQL 1.0. Test samples shall be selected from each manufacturing lot such as that each material lot is represented in the samples. The unit of product for test purposes shall be one catheter and needle unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE 1: The following National Stock Numbers, including item identification and designation code, are covered by this document.

NSN	ITEM IDENTIFICATION	DESIGNATION
6515-00-403-2971	Catheter and Needle Unit, Intravenous, Catheter, Into Needle, 16 gage catheter, 36 inches Long, 14 gage Needle, 2 Inch Long, Radiopaque, Disposable, 50's	Catheter size A, needle size a, usable catheter length 4, material Type III

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

6515-01-208-7328	Catheter and Needle Unit, Intravenous, Catheter Into Needle, 16 Gage catheter, 24 Inches Long, 14 Gage Needle, 2 Inches Long, Radiopaque, Disposable, 50's	Catheter size A, needle size, a usable catheter length 4, material Type I or II
6515-01-208-7329	Catheter and Needle Unit, Intravenous, Catheter Into Needle, 18.5 or 19 Gage catheter, 12 Inches Long, 17 Gage Needle, 2 Inches Long, Radiopaque, Disposable, 50's	Catheter size B, needle size b, usable catheter length 2, material Type I or II
6515-01-209-1174	Catheter and Needle Unit, Intravenous, Catheter Into Needle, 16 Gage catheter, 12 Inches Long, 14 Gage Needle, 2 Inches Long, Radiopaque, Disposable, 50's	Catheter size A, needle size b, usable catheter length 2, material Type I or II
6515-01-209-1175	Catheter and Needle Unit, Intravenous, Catheter Into Needle, 18.5 or 19 Gage Catheter, 24 Inches Long, Radiopaque, Disposable, 50's	Catheter size B, needle size b, usable catheter length 3, material Type I or II

NOTE 2:

The following is acceptable as an alternate to USP Class IV plastics lot testing: The catheter shall be initially qualified for Class IV plastics in accordance with the USP, and simultaneously, tested for tissue culture cytotoxicity in accordance with HIMA "Guideline For Evaluating The Safety of Materials Used in Medical Devices". Subsequent production lots which show greater toxicity than the initial qualification lot shall be subjected to USP Class IV biological testing.

In the event of a change in material formulation, USP Class IV initial qualification testing shall again be required.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

NOTE 3:

Certificates of heavy metals testing. A certificate of compliance, stipulating that offered items meet the heavy metals extractable testing, shall be acceptable, provided that the contractor has supplied acceptable items to the Government within the past three years and that there have been no changes to material, formulation and method of processing since actual testing was performed.

NOTE 4:

In lieu of having each material batch represented in the finished goods chemical, safety and biological test samples, The following shall be acceptable as an alternate: New raw materials shall be initially qualified through chemical, safety, and biological testing and again testing after assembly, sterilization and aging. Subsequent incoming materials will be tested physically for material consistency. Any changes to material or processing shall be requalification tested.

Ordering data (Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4706

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 Alexandria, VA

POSTAGE WILL BE PAID BY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

NO POSTAGE
NECESSARY
IF MAILED
IN THE
UNITED STATES

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-54154	CATHETER AND NEEDLE UNIT, INTRAVENOUS, RADIOPAQUE, Disposable, 50s (Catheter into Needle)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 1426 *(EG) (PE) USE THIS VERSION UNTIL EXHAUSTED.*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.