COMMERCIAL ITEM DESCRIPTION

RETRACTOR, ORTHOPEDIC (5-1/2 Inch, Smillie)

The General Services Administration has authorized the use of this commercial item description

This commercial item description covers curved or angled Smillie retractors suitable for use in orthopedic surgery

Salient characteristics

General The retractor shall be 5-1/2 inches long with either an angled or curved blade

Material Material of the retractor shall be Class 3 austenitic or Class 4 martensitic stainless steel in accordance with ASTM Designation F899

Hardness The retractor shall have a Rockwell Hardness B scale (HRB) of 80 to 95 when fabricated from Class 3 steel and Rockwell Hardness C scale (HRC) 40 to 47 when fabricated from Class 4 and when tested in accordance with ASTM Designation E18 or E92 whichever is applicable Conversion of hardness values shall be in accordance with ASTM Designation E140

Style, design and dimensions The retractor shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No 24455 for style, design and dimensions

Finish The retractor shall have a mirror, satin or matte finish The matte finish shall exhibit a uniform nondirectional surface texture equivalent to that produced by lightly blasting the prepared surface with fine abrasives [illegible]

Passivation If fabricated from Class 4 steel the retractor shall be properly passivated

Corrosion resistance Class 4 retractors shall show no signs of corrosion or plating of copper when copper sulfate and [illegible] tested in accordance with ASTM Designation F1089 Class 3 retractors shall show no signs of corrosion when salt tested in accordance with ASTM Designation F.089

DISTRIBUTION STATEMENT A Approved for public release distribution is unlimited

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

Identification marking and labeling Retractors shall be legibly marked with the contractor's name or trademark and the word 'stainless' Foreign manufactured retractors shall bear the name of the country of origin

Workmanship The retractor shall be free from defects which detract from its appearance or impair its serviceability

Visual and dimensional examination Retractors shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements

Materials The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements

Unit Each (EA) One retractor, as specified, constitutes one unit

[illegible]

Responsibility for Inspection Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein [illegible] The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements

Records Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate

Inspection Inspection, as used herein is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

Sampling for examination Sampling of examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one retractor. Sampling shall be Inspection level II with an acceptable quality level (AQL) of 1 0

Examination Examination shall be conducted to determine compliance with specification requirements.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein Unit of product for tests purposes shall be one retractor Sampling shall be Inspection level [illegible] with an AQL of 1 0

Tests Tests shall be conducted to determine compliance with specification requirements Where feasible, the same sample shall be used for the determination of two or more test

Contractor certification The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract characteristics.

Metric products Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No 376, and all other requirements of this document are met

If a product is manufactured to metric dimensions and these dimensions exceed the tolerances specified in the inch/pound units a request should be made to the contracting officer to determine if the product is acceptable

The contracting officer has the option of accepting or rejecting the product

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

Regulatory requirements

Federal Food, Drug and Cosmetic Act If the product covered by this document has been determined by the U S Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations

Recovered materials The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23 4 to the maximum extent practical

Preservation, packaging, packing, labeling, and marking Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling etc , involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity and shall conform to the applicable carrier's rules and regulations Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order

NOTE The following National Stock Number are covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-139-7972	Retractor, Orthopedic, Smillie, Reverse Curve 5 1/2 Inch 2-1 8 by 1/2 Inch Blade
6515-01-139-7973	Retractor, Orthopedic Smillie Small Angled, 5-1/2 Inch Long 1 1 4 by 1-3/8 Inch Blade
6515 01 139 7974	Retractor, Orthopedic, Smillie, Medium Angled 5 1/2 Inch Long, 1 4 by 1 Inch Blade

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

NSN	ITEM IDENTIFICATION
6515-01-139-7975	Retractor, Orthopedic, Smillie, Large Angled, 5-1/2 Inch Long, 3/4 by 2-3/8 Inch Blade
6515-01-139-7976	Retractor, Orthopedic, Smillie, Curved, 5-1/2 Inch Long, 1/2 by 5/8 Inch Blade

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order)

MILITARY INTERESTS

Custodians:

Army - MI
Navy - MS
Air Force - 03

Preparing Activity

DoD - MB

Agent

DLA - DM

CIVIL AGENCY COORDINATING ACTIVITIES

VA-OSS
PHS
FDA MEDAS

Project No 6515-4731

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements All users of military standardization documents are invited to provide suggestions This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements

(Fold along this line)

(Fold along this line)

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 Alexandria, VA

POSTAGE WILL BE PAID BY

Commander
Defense Personnel Support Center
ATTN: DPSC-RST
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

(T) DETACH THIS FORM, CUT ALONG THIS LINE.)

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-54155	RONGEUR, ORTHOPEDIC (5-1/2 Inch, Smillie)

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)* _____
b ADDRESS *(Street, City, State, ZIP Code)*	

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last, First, MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

DD FORM 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.