INCH-POUND

A-A-55708
9 August 1996

COMMERCIAL ITEM DESCRIPTION

ROTATOR, LABORATORY,
Variable Speed

The General Services Administration has authorized the use of this Commercial Item Description for all Federal Agencies.

1. SCOPE. This Commercial Item Description covers a portable variable speed platform rotator for use in a variety of laboratory procedures.

2. SALIENT CHARACTERISTICS:

2.1 The unit shall be a portable, variable speed, platform rotator for use in a variety of laboratory procedures including standard serological procedures.

2.2 The unit shall consist of a base and a top mounted platform.

2.3 The unit shall have the nominal dimensions of 12 inches Wide by 13 inches Deep by 6 inches High.

2.4 The unit shall be equipped with:

2.4.1 A solid state speed control for selecting speeds in a nominal range of 80-210 rpm’s.

2.4.2 A 0-30 minute timer and a selector switch for timed or continuous operation.

Beneficial comments, recommendations, additions, deletions, clarifications, etc. and any other data which may be used to improve this document should be sent to: Defense Personnel Support Center, Directorate of Medical Materiel, ATTN: DPSC-MBP, 2800 South 20th Street, Philadelphia, PA 19145-0530, by letter or using the self-addressed Standardization Document Improvement Proposal, (DD Form 1426) appearing at the end of this document.

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:56Z
Check the source to verify that this is the current version before use.

2.4.3 A rimmed platform that is padded with a material that prevents labware from sliding during operation. The platform shall have nominal dimensions of 13 inches by 13 inches.

2.5 Workmanship . The laboratory rotator shall be free from defects which detract from its appearance or impair its serviceability.

2.6 Unit . Each (EA). One laboratory rotator as specified, constitutes one unit. Each unit shall be packaged in accordance with the requirements of ASTM D 3951.

3. REGULATORY REQUIREMENTS .

3.1 Federal Food, Drug, and Cosmetic Act . If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/ contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

3.2 Recovered materials . The offeror/contractor is encouraged to use recovered materials in accordance with the Federal Acquisition Regulation Subpart 23.403 to the maximum extent practical.

4. QUALITY ASSURANCE PROVISIONS .

4.1 Product Conformance . The products provided shall meet the salient characteristics of this commercial item description, conform to the producer’s own drawings, specifications, standards, and quality assurance practices, and be the same product offered for sale in the commercial market. The government reserves the right to require proof of such conformance.

4.2 Metric products . Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units. Dimensions and tolerances specified shall be converted using conversion tables contained in the latest revision of ASTM E 380.

4.3 Market acceptability . The following market acceptability criteria are necessary to document the quality of the product to be provided under this Commercial Item Description.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:56Z
Check the source to verify that this is the current version before use.

4.3.1 The company producing the item must have been producing a product meeting the requirements of this Commercial Item Description for at least two (2) years.

4.3.2 The company must have sold 1,000 units meeting this Commercial Item Description in the commercial market place over the past two (2) years.

4.4 Other quality assurance provision . Shall be as indicated in Section E, Inspection, of the contract and/or order.

5. PACKAGING .

5.1 Preservation, packaging packing, labeling and marking. Shall be as specified in the contract and/or order. Specific requirements for this item are indicated in Section C, Description/Specification; general packaging and marking requirements are indicated in Section D, Packaging and Marking.

6. NOTES .

6.1 National Stock Number Data . The following National Stock Number and its respective Item Identification is covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6640-01-172-1132	ROTATOR, LABORATORY, Variable Speed, 120/230 volt, 50/60 Hz AC 6

6.2 Source of Documents :

6.2.1 Government Documents

Defense Printing Service Detachment Office
Bldg. 4D (NPM-DODSSP)
700 Robbins Avenue
Philadelphia, PA 19111-5094

6.2.2 Non-Government Documents .

ASTM documents.

(Applications for copies should be addressed to the American Society for Testing and Materials (ASTM), 100 Barr Harbor Drive, West Conshohocken, PA 19428-2959.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:56Z
Check the source to verify that this is the current version before use.

ANSI documents.

(Applicat ions for copies should be addressed to the American National Standards Institute, 11 West 42nd Street, New York, NY 10036-8002.)

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS
GSA-FSS-FCGC

Project Number: 6640-1879
Location: ROTATC14.DOC (WORD 6)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:56Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-55708	960809

3. DOCUMENT TITLE
ROTATOR, LABORATORY, Variable Speed

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MBP	(215)737-8431	444-8431
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 *Previous editions are obsolete.* 298/290

Source: https://assist.dla.mil - Downloaded: 2026-10-26T09:56Z
Check the source to verify that this is the current version before use.