INCH-POUND

A-A-55711
16 August 1996

COMMERCIAL ITEM DESCRIPTION

FREEZER, UNDER COUNTER, BLOOD PLASMA

The General Services Administration has authorized the use of this Commercial Item Description for all Federal Agencies.

1. SCOPE. This Commercial Item Description covers an under-counter, blood plasma freezer.

2. SALIENT CHARACTERISTICS:

2.1 The item shall be an under the counter freezer capable of achieving and maintaining a temperature as low as –30°F suitable for freezing and maintaining plasma.

2.2 The freezer shall have a storage capacity to accommodate at least 75 units of frozen plasma packaged in standard approved plastic bags and corrugated boxes measuring 7-1/8 inches by 5-5/8 inches by 1-7/8 inches (filled and bulging dimensions) .

2.3 The unit interior and exterior shall be constructed of stainless steel with a solid door and shall have three stainless steel, roll–out drawers.

2.4 The unit shall have a total exterior cube not to exceed 12.5 cubic feet.

2.5 The unit shall comply with all applicable Federal laws to include the requirement to use CFC free refrigerant, insulation and packaging.

Beneficial comments, recommendations, additions, deletions, clarifications, etc. and any other data which may be used to improve this document should be sent to: Defense Personnel Support Center, Directorate of Medical Materiel, ATTN: DPSC-MBP, 2800 South 20th Street, Philadelphia, PA 19145-0530, by letter or using the self–addressed Standardization Document Improvement Proposal, (DD Form 1426) appearing at the end of this document.

AMSC N/A FSC 4110

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:56Z
Check the source to verify that this is the current version before use.

2.6 The unit shall meet all American Association of Blood Bank requirements for plasma freezers.

2.7 The freezer shall be equipped with:

2.7.1 A continuously recording thermometer capable of recording temperatures for seven days on one chart.

2.7.2 An alarm system with an audible and visual signal for violation of temperature requirements and loss of electrical power. Alarm system shall have an automatically recharged battery backup.

2.7.3 A one year supply of temperature recording charts.

2.8 The unit shall operate satisfactorily from 115/230 volts 50/60 Hz. consistent with the requirements contained in MIL–D-42048.

2.9 Dielectric strength. The unit shall be tested per UL-471 and show no sign of dielectric failure.

2.10 Identification marking. The unit shall have a permanently affixed nameplate which is legibly marked in accordance with MIL-D-42048.

2.11 Wiring diagram. A wiring diagram shall be affixed by means of an adhesive backed mylar overlay to the back of the cabinet or to the machine compartment cover. The diagram shall portray each electrical component by sketch, outline, or symbol, and shall include the name of each component. The wiring diagram shall show the wiring to all components and any coding used for identification of wires.

2.12 Service data. The contractor shall furnish with each unit two copies each of a technical manual which contains complete instructions for installation, operation, maintenance and lists of component parts. All manuals, handbooks, and/or brochures will be written in the English language, and all schematics and wiring diagrams will use American electrical and electronics symbols. The manuals will include electrical data and connection diagrams for all applicable utilities. The instruction shall also contain a complete list of all replaceable parts showing part number, name, and quantity required. The manual shall comply with the requirements of MIL-M–7298 Commercial Off the Shelf Manuals. As soon as practical after award of the contract or purchase order, the contractor shall furnish to the

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:56Z
Check the source to verify that this is the current version before use.

Contracting Officer two preliminary copies of his proposed technical manual for review and comment by the Directorate of Medical Materiel at Defense Personnel Support Center (based on a maximum period of 30 days for such review) and to allow incorporation by the contractor of any required revisions and corrections in the final manual without delaying delivery under the contract or purchase order. In addition, the Contractor shall furnish 15 copies of the approved technical manual to the Contracting Officer for distribution. Should the contractor have furnished manuals identical with those previously accepted within two years previous to the date of contract, then the requirements for review of drafts and submission of 15 additional may be waived upon request.

2.13 Finish. The unit shall be fabricated from materials which are inherently resistant to corrosion or which are treated to render them corrosion-resistant. Where paining or plating is used, appropriate cleaning and priming shall be performed prior to finishing. Unless otherwise specified finish shall be equivalent or superior to that of comparable commercial equipment, and shall be free of gaps, scratches, chips, sharpness, and other imperfections.

2.14 Workmanship. The freezer shall be free from defects which detract from its appearance or impair its serviceability.

2.15 Unit. Each (EA). One freezer, as specified, constitutes one unit. Each freezer shall be packaged in accordance with the requirements of ASTM D 3951.

3. REGULATORY REQUIREMENTS.

3.1 Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/ contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

3.2 Recovered materials. The offeror/contractor is encouraged to use recovered materials in accordance with the Federal Acquisition Regulation Subpart 23.403 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:56Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS.

4.1 Product Conformance. The products provided shall meet the salient characteristics of this commercial item description, conform to the producer's own drawings, specifications standards, and quality assurance practices, and be the same product offered for sale in the commercial market. The government reserves the right to require proof of such conformance.

4.2 Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch–pound units. Dimensions and tolerances specified shall be converted using conversion tables contained in the latest revision of ASTM E 380.

4.3 Market acceptability. The following market acceptability criteria are necessary to document the quality of the product to be provided under this Commercial Item Description.

4.3.1 The company producing the item must have been producing a product meeting the requirements of this Commercial Item Description for at least two (2) years.

4.3.2 The company must have sold 1,000 units meeting this Commercial Item Description in the commercial market place over the past two (2) years.

4.4 Other quality assurance provision. Shall be as indicated in Section E, Inspection, of the contract and/or order.

5. PACKAGING.

5.1 Perservation, packaging packing, labeling and marking. Shall be as specified in the contract and/or order. Specific requirements for this item are indicated in Section C, Description/Specification; general packaging and marking requirements are indicated in Section D, Packaging and Marking.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:56Z
Check the source to verify that this is the current version before use.

6. NOTES.

6.1 National Stock Number Data. The following National Stock Number and its respective Item Identification is covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
4110-01-424-9408	FREEZER, UNDER-COUNTER, BLOOD PLASMA

6.2 Source of Documents:

6.2.1 Government Documents

Defense Printing Service Detachment Office
Bldg. 4D (NPM-DODSSP)
700 Robbins Avenue
Philadelphia, PA 19111-5094

6.2.2 Non-Government Documents.

ASTM documents.

(Applications for copies should be addressed to the American Society for Testing and Materials (ASTM), 100 Barr Harbor Drive, West Conshohocken, PA 19428-2959.)

ANSI documents.

(Applications for copies should be addressed to the American National Standards Institute, 11 West 42nd Street, New York, NY 10036-8002.)

Underwriters Laboratories documents.

(Applications for copies should be addressed to the: Publications Stock, 333 Pfingsten Road, Northbrook, IL 60062.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:56Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force – 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS
GSA-FSS-FCGC

Project Number: 4110-0504
Location: FRZUCC08.DOC (WORD 6)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:56Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-55711	960816

3. DOCUMENT TITLE
FREEZER, UNDER COUNTER, BLOOD PLASMA

4. NATURE OF CHANGE (*Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.*)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MBP	(215)737-8431	444-8431
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

Source: https://assist.dla.mil -- Downloaded: 2026-01-26T09:56Z
Check the source to verify that this is the current version before use.