INCH-POUND

A-A-55712
16 August 1996

COMMERCIAL ITEM DESCRIPTION

METER, HYDROGEN ION TEST,
115/230 volts, 50/60 Hz, AC

The General Services Administration has authorized the use of this Commercial Item Description for all Federal Agencies.

1. SCOPE. This Commercial Item Description covers a Portable Solid State pH Meter.

2. SALIENT CHARACTERISTICS:

2.1 The unit shall be a portable, solid state pH meter suitable to be used in clinical laboratories for measuring the hydrogen ion content in aqueous based buffers and bacteriological media preparations.

2.2 The unit shall be a general purpose analog pH meter with range capabilities that include 0 to 14 pH units and from +400 millivolts (mv). The unit shall have an accuracy of ±0.1 pH units. The pH scale shall be graduated in 0.2 pH increments. The millivolt (mv) scale shall be graduated in 20 millivolt (mv) increments.

2.3 The unit shall have a mirror-backed scale and a scale length of 120 millimeters (mm) with an isopotential point of pH 7.0. The unit shall have a front mount temperature compensation scale that covers 0 to 100 ^{o}C with a manually operated selector switch.

Beneficial comments, recommendations, additions, deletions, clarifications, etc. and any other data which may be used to improve this document should be sent to: Defense Personnel Support Center, Directorate of Medical Materiel, ATTN: DPSC-MBP, 2800 South 20th Street, Philadelphia, PA 19145-0530, by letter or using the self-addressed Standardization Document Improvement Proposal, (DD Form 1426) appearing at the end of this document.

AMSC N/A FSC 6630

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:56Z
Check the source to verify that this is the current version before use.

2.4 The unit shall be supplied with a laboratory grade, potassium chloride gel-filled combination pH electrode.

2.5 The unit shall be supplied with a support rod which has an adjustable electrode holder. The electrode holder shall be capable of being vertically adjusted on the support rod to facilitate the insertion of the pH electrode into a test solution.

2.6 The unit shall be comparable with the use of pH buffer's standards (pH 4 buffer, pH 7 buffer, pH 10 buffer).

2.7 The unit shall operate satisfactorily from 115/230 volts 50/60 Hz. consistent with the requirements contained in MIL-D-42048.

2.8 Reusable container . Section 3.6 that is specified in MIL-D-42048 is not required.

2.9 Dielectric strength . The unit shall be tested per UL and show no sign of dielectric failure.

2.10 Identification marking . The unit shall have a permanently affixed nameplate which is legibly marked in accordance with MIL-D-42048.

2.11 Service data . The contractor shall furnish with each unit two copies of a technical manual which contains complete instructions for installation, operation, maintenance and lists of component parts. All manuals, handbooks, and/or brochures will be written in the English language, and all schematics and wiring diagrams will use American electrical and electronics symbols. The manuals will include electrical data and connection diagrams for all applicable utilities. The instruction shall also contain a complete list of all replaceable parts showing part number, name, and quantity required. The manual shall comply with the requirements of MIL-M-7298, Commercial Off the Shelf Manuals. As soon as practical after award of the contract or purchase order, the contractor shall furnish to the Contracting Officer two preliminary copies of his proposed technical manual for review and comment by the Directorate of Medical Materiel at Defense Personnel Support Center (based on a maximum period of 30 days for such review) and to allow incorporation by the contractor of any required revisions and corrections in the final manual without delaying delivery under the contract or purchase order. In addition, the Contractor shall furnish 15 copies of the approved technical manual to the

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:56Z
Check the source to verify that this is the current version before use.

Contracting Officer for distribution. Should the contractor have furnished manuals identical with those previously accepted within two years previous to the date of contract, then the requirements for review of drafts and submission of 15 additional may be waived upon request.

2.12 Finish. The unit shall be fabricated from materials which are inherently resistant to corrosion or which are treated to render them corrosion-resistant. Where painting or plating is used appropriate cleaning and printing shall be performed prior to finishing. Unless otherwise specified, finish shall be equivalent or superior to that of comparable commercial equipment, and shall be free of gaps, scratches, chips, and other imperfections.

2.13 Workmanship. The pH meter shall be free from defects which detract from its appearance or impair its serviceability.

2.14 Unit. Each (EA). One pH meter, as specified, constitutes one unit. Each unit shall be packaged in accordance with the requirements of ASTM D 3951.

3. REGULATORY REQUIREMENTS.

3.1 Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/ contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

3.2 Recovered materials. The offeror/contractor is encouraged to use recovered materials in accordance with paragraph 23.403 of the Federal Acquisition Regulation (FAR) to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:56Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS .

4.1 Product Conformance . The products provided shall meet the salient characteristics of this commercial item description, conform to the producer's own drawings, specifications, standards, and quality assurance practices, and be the same product offered for sale in the commercial market. The government reserves the right to require proof of such conformance.

4.2 Metric products . Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units. Dimensions and tolerances specified shall be converted using conversion tables contained in the latest revision of ASTM E 380.

4.3 Market acceptability . The following market acceptability criteria are necessary to document the quality of the product to be provided under this Commercial Item Description.

4.3.1 The company producing the item must have been producing a product meeting the requirements of this Commercial Item Description for at least two (2) years.

4.3.2 The company must have sold 1,000 units meeting this Commercial Item Description in the commercial market place over the past two (2) years.

4.4 Other quality assurance provision . Shall be as indicated in Section E, Inspection, of the contract and/or order.

5. PACKAGING .

5.1 Preservation, packaging packing, labeling and marking. Shall be as specified in the contract and/or order. Specific requirements for this item are indicated in Section C, Description/Specification; general packaging and marking requirements are indicated in Section D, Packaging and Marking.

6. NOTES .

6.1 National Stock Number Data . The following National Stock Number and its respective Item Identification is covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6630-01-188-5382 (W)	METER, HYDROGEN ION TEST, 115/230 volts, 50/60 Hz

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:56Z
Check the source to verify that this is the current version before use.

6.2 Government Documents

Standardization Documents Order Desk
Bldg. 4D (NPM-DODSSP)
700 Robbins Avenue
Philadelphia, PA 19111-5094

6.3 Non-Government Documents .

ASTM documents.

(Applications for copies should be addressed to the American Society for Testing and Materials (ASTM), 100 Barr Harbor Drive, West Conshohocken, PA 19428-2959.)

ANSI/ASQC Z1.4 - Sampling Procedures and Tables for Inspection of Attributes.

(Applications for copies should be addressed to the American National Standards Institute, 11 West 42nd Street, New York, NY 10036-8002.)

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS
GSA-FSS-FCGC

Project Number: 6630-0564
Location: METERC08.DOC (WORD 6)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:56Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-55712	960816

3. DOCUMENT TITLE

METER, HYDROGEN ION TEST, 115/230 VOLTS, 50/60 HZ, AC

4. NATURE OF CHANGE *(Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)*

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code)	
DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MBP	(1) Commercial (215)737-8431	(2) AUTOVON 444-8431
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:56Z
Check the source to verify that this is the current version before use.