A-A-50890A
24 March 1989
SUPERSEDING
A-A-50890
13 July 1983

COMMERCIAL ITEM DESCRIPTION

SCREEN, X-RAY PROTECTIVE

Mobile

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-S-15253B which is cancelled.

This Commercial Item Description covers a mobile, x-ray protective screen for protecting operators from primary and secondary radiation when operating x-ray and photofluorographic equipment.

Salient characteristics:

The x-ray protective screen shall be a mobile, winged screen mounted on casters. The screen height shall be approximately 84 inches high and approximately 44 inches overall width, which includes the wings. The screen shall be suitable for protecting operators from primary and secondary radiation when operating x-ray and photofluorographic equipment.

The screen shall have a lead lining of at least 1/16 inch. Sheet steel used for encasing the lead shall be a minimum of 0.031 inch (22 gage) thick. All exposed metal parts shall be finished with baked on enamel of the manufacturer's standard color. The finish shall be free of bare, chipped out, or excessively scratched areas.

The center portion of the screen shall contain an 8 inch by 10 inch viewing window with x-ray attenuation equivalent to at least 0.16 cm of lead equivalent at 100 kVp. The window shall be transparent, clear, and free of cracks and flaws.

AMSC N/A FSC 6525

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

The screen shall be mounted on three or four smooth-rolling, quiet casters which are at least 3 inches in diameter.

The screen metal components shall be free of cracks, tears, ruptures, ripples, or noticeably deformed areas. The edges and corners of the screen shall be free of burrs and sharpness. The screen shall be free of oil, grease, dirt, or other contamination.

Workmanship. The x-ray protective screen shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One screen, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one x-ray protective screen as specified in this description.

Defects.

Screen is not complete with all specified component parts.

Screen casters are not at least 3 inches in diameter.

Component parts are not properly fitted and assembled.

Screen window is not transparent, clear, and free of cracks and flaws.

Metal parts are not free of cracks, tears, ruptures, ripples, or noticeably deformed areas.

Edges and corners of screen are not free of burrs and sharpness.

Finish is not enamel.

Finish is not free of bare, chipped out, or excessively scratched areas.

Screen is not free of oil, grease, dirt, or other contamination.

Sampling for tests. Sampling for tests, including dimensional test, but excluding lead equivalency test, shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-3. The unit of product for test purposes shall be one x-ray protective screen.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

a. First article. Where a first article sample is required, all tests shall be performed on that unit.

b. Acceptance testing. When a first article sample has not been required, one unit shall be subjected to the performance test specified herein. If a test article sample has successfully passed all tests, performance test shall not be required at acceptance.

Tests

Lead Equivalency. With the x-ray screen window removed, a suitable penetrometer shall be placed such that there is an unobstructed path for x-ray through the window opening and that part of the screen near the window, both which can be x-rayed at the same time on one 14 inch by 17 inch x-ray radiograph. An x-ray radiograph shall be taken in any random area around the window at 100 kVp. The x-ray exposure shall be examined for comparison purposes. A lead equivalency of less than 0.16 cm of lead shall constitute a defect.

Density Uniformity. The x-ray screen shall be tested by x-ray exposure at 100 kVp in at least two random places on each panel of the screen. Any essentially different density shall constitute a defect.

Rigidity Test. With the x-ray screen in the normal position, and the base held firmly, a force of 50 pounds shall be applied at any point and in any direction at the top of the screen. Play in any direction, exceeding 1/2 inch, shall constitute a defect.

Rolling Test. The x-ray screen shall be placed on a smooth horizontal concrete floor. The casters shall be clean. A horizontal force shall be applied gently to the top of the screen at any point in any direction. The force shall be gradually increased until the screen rolls or upsets. If the screen upsets, it shall constitute a defect.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6525-01-086-7025

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6525-1245

Location: CID/SCREEN/H13-31

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-50890A	SCREEN, X-RAY PROTECTIVE, Mobile

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b ADDRESS *(Street City State ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last First MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street City State ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge (*DO NOT STAPLE*), and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4968 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC- RSIE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.