INCH-POUND
A-A-50891A
17 September 1993
SUPERSEDING
A-A-50891
22 August 1985

COMMERCIAL ITEM DESCRIPTION

VIBRATOR, DENTAL MOLDING

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers an electrically operated vibrator suitable for use in dental prosthetic laboratories for packing gypsum materials.

Salient characteristics:

1. The vibrator shall consist of a top vibrating deck activated by an magnetic vibrating system contained in a base assembly, mounted on suitably cushioned feet.

2. The vibrating system shall include a line cord, an on-off, double-pole, single throw switch, a half-wave silicon-power rectifier, a coil with core, an air gap, a spring-mounted armature attached to the vibrating deck, and a wire-wound rheostat.

3. The rectifier shall be capable of withstanding peak inverse voltages up to 400 volts, and shall produce an input of pulsating direct current to the coil, and drive the armature (and deck) at a constant rate of 60 vibrations per second on 115V, 60Hz, and 50 vibrations/sec on 115V, 50Hz.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

4. The rheostat shall be capable of varying the displacement of vibration, without load, from a nominal 0.010 inch to a nominal 0.060 inch.

5. The air gap shall be an adjustable type, readily accessible from an exterior surface.

6. The unit shall be 115 volts and shall operate from 50 to 60 Hertz. The service instructions shall include instructions for adjusting the unit from 50 to 60Hz. The unit shall initially be delivered with the 115V, 60Hz configuration.

7. Metal components shall be suitably protected against corrosion. Electrical components shall be suitably coated or protected against moisture.

8. The vibrating system shall have a nominal weight of 21 pounds and shall be capable of vibrating a load of 10 pounds at a nominal displacement of 0.035 inch.

9. The vibrator shall be UL 544 certified for the 115V, 60Hz version of the unit and shall meet the requirements of UL 544 for the 115V, 50Hz version.

10. Two copies of a technical manual containing complete instructions for installation, operation, and maintenance, and a list of component parts, shall be furnished with each vibrator:

Workmanship. The dental molding vibrator shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One dental molding vibrator, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one dental molding vibrator.

Sampling for tests. Sampling for tests, including dimensional tests, shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for test purposes shall be one dental molding vibrator.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics. Tests shall include, but shall not be limited to, the following:

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Operational test. A 5-pound load shall be centrally positioned on the deck within the deck ribs. With the rheostat set to 3/4 of the maximum setting the vibrator shall operate for a period of 72 hours on 115V, 60Hz power. The operational test shall also be performed as described on 220V, 50Hz power.

Vibration-displacement test. The vibrator shall vibrate at 3600 vibrations per minute on 115V, 60 Hz power. The table displacement shall be measured from 0.010 to 0.060 plus or minus 0.010 inch with a centrally located 5-pound load.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug, and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6520-01-205-2427

MILITARY INTERESTS:

Custodians:

Air Force-03
Army -MD
Navy-MS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
USPHS
VA-OSS

Project Number: 6520-2565
Location: VIBRATOR16-CID

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I. RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-50891A	17 September 1993

3. DOCUMENT TITLE: VIBRATOR, DENTAL MOLDING

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MSE	b. TELEPHONE (Include Area Code) (1) Commercial: AC 215-737-2870 (2) AUTOVON: AV 444-2870
c. ADDRESS (Include Zip Code) 2800 SO. 20TH STREET PHILADELPHIA, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete — 9879C

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.