COMMERCIAL ITEM DESCRIPTION

LIGHT, DIAGNOSTIC-EXAMINATION

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a diagnostic-examination light, 115 volt, 60 Hz, AC.

Salient characteristics:

Shall be a hand-held light which provides long-wave ultraviolet light and daylight fluorescent illumination, with operator observation capability through an integral magnifier lens, for use in medical examination and diagnosis.

Main body of light shall consist of a rectangular housing, approximately 7 inches long by 5-1/2 inches wide by 2-1/2 inches thick, containing the following:

a. Sockets, ballast and wiring for four lamps.
b. Two ultraviolet lamps, 4 watts, 6 inch miniature bipin, NSN 6240-00-319-4549.
c. Two cold white fluorescent lamps, 4 watts, 6 inch miniature bipin, NSN 6240-00-796-6240.
d. Switch or switches to provide ON-OFF control and selection of either ultraviolet or daylight fluorescent light.
e. Magnifier lens.

A handle, approximately 6 inches long, shall be attached midway on the long side of the main body housing.

Light unit shall operate on a nominal 115 volts, 50/60 Hz. Shall be provided with a 3 conductor power cord (with ground lead) approximately 6 feet long and terminating in a 3 prong male plug that complies with American National Standards Institute (ANSI) Standard C73.11.

Lamps shall be mounted parallel to the long plane of the magnifier lens and arranged to provide one lamp of each type on each side of that lens.

Ultraviolet light shall be long wave, above 3200 Angstrom units in the ultraviolet spectrum, with peak average of 3500 to 3700 Angstrom units.

Magnifier lens shall provide distortion-free viewing of at least 3 power magnification, be optically ground and polished glass, measure approximately 4 inches long, 2 inches wide, and have a nominal 8 inch focal length. Lens shall be centrally mounted in the main body housing with its long plane parallel to long plane of housing and at a right angle to the handle bracket.

Weight of complete light unit shall not exceed 3-1/2 pounds.

Shall meet requirements of Underwriters' Laboratories, Inc.

Technical manual. Each light shall be furnished with two copies of a technical manual containing, as a minimum, complete instructions for operation and maintenance, a

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:48Z
Check the source to verify that this is the current version before use.

wiring diagram, as well as illustrations and a listing identifying all component parts. The manufacturer's part numbers shall be given for each component other than common fasteners.

Workmanship. Lights shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Each (EA). One diagnostic-examination light, as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices, and is the same as the product sold in the commercial marketplace. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

The following National Stock Number is covered by this document: 6530-00-480-8286

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians

Army-MD
Navy-MS
Air Force-03

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

PREPARING ACTIVITY:

DoD-MB

AGENT:

DLA-DM

50215

Project No. 6530-1589

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:48Z
Check the source to verify that this is the currentversion before use.